

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102622\
 Data File : BN022330.D
 Acq On : 26 Oct 2022 10:10
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 27 05:57:16 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 27 05:55:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	99	0.00
2	1,4-Dioxane	0.400	0.377	5.8	98	0.00
3	n-Nitrosodimethylamine	0.400	0.373	6.8	101	0.00
4 S	2-Fluorophenol	0.400	0.432	-8.0	117	0.00
5 S	Phenol-d6	0.400	0.451	-12.7	125	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	109	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	117	0.00
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	128	0.00
9	Naphthalene	0.400	0.386	3.5	121	0.00
10	Hexachlorobutadiene	0.400	0.372	7.0	116	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.349	12.8	111	0.00
12	2-Methylnaphthalene	0.400	0.495	-23.7	173	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	130	0.00
14 S	2,4,6-Tribromophenol	0.400	0.358	10.5	136	0.00
15 S	2-Fluorobiphenyl	0.400	0.387	3.3	131	0.00
16	Acenaphthylene	0.400	0.389	2.8	143	0.00
17	Acenaphthene	0.400	0.387	3.3	135	0.00
18	Fluorene	0.400	0.431	-7.7	154	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	119	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.322	19.5	70	0.00
21	4-Bromophenyl-phenylether	0.400	0.409	-2.2	132	0.00
22	Hexachlorobenzene	0.400	0.395	1.3	127	0.00
23	Atrazine	0.400	0.382	4.5	133	0.00
24	Pentachlorophenol	0.400	0.325	18.8	121	0.00
25	Phenanthrene	0.400	0.384	4.0	124	0.00
26	Anthracene	0.400	0.375	6.3	127	0.00
27 SURR	Fluoranthene-d10	0.400	0.328	18.0	109	0.00
28	Fluoranthene	0.400	0.319	20.3	105	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	81	0.00
30	Pyrene	0.400	0.480	-20.0	103	0.00
31 S	Terphenyl-d14	0.400	0.511	-27.7#	109	0.00
32	Benzo(a)anthracene	0.400	0.393	1.8	88	0.00
33	Chrysene	0.400	0.383	4.3	81	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.469	-17.2	108	0.00
35 I	Perylene-d12	0.400	0.400	0.0	83	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.372	7.0	86	0.00
37	Benzo(b)fluoranthene	0.400	0.353	11.8	81	0.00
38	Benzo(k)fluoranthene	0.400	0.369	7.8	83	0.00
39 C	Benzo(a)pyrene	0.400	0.372	7.0	86	0.00
40	Dibenzo(a,h)anthracene	0.400	0.372	7.0	85	0.00
41	Benzo(g,h,i)perylene	0.400	0.375	6.3	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4