

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102720\
 Data File : BN012403.D
 Acq On : 27 Oct 2020 09:39
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 27 11:35:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 27 11:09:04 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.651	0.591	9.2	121	0.00
3	n-Nitrosodimethylamine	1.581	1.427	9.7	113	0.00
4 S	2-Fluorophenol	1.398	1.260	9.9	113	0.00
5 S	Phenol-d6	1.655	1.524	7.9	117	0.00
6	bis(2-Chloroethyl)ether	1.127	0.994	11.8	110	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
8 S	Nitrobenzene-d5	0.441	0.397	10.0	124	0.00
9	Naphthalene	1.120	0.980	12.5	117	0.00
10	Hexachlorobutadiene	0.237	0.211	11.0	116	0.00
11 SURR	2-Methylnaphthalene-d10	0.643	0.586	8.9	123	0.00
12	2-Methylnaphthalene	0.727	0.654	10.0	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.284	0.249	12.3	127	0.00
15 S	2-Fluorobiphenyl	1.488	1.291	13.2	122	0.00
16	Acenaphthylene	1.975	1.670	15.4	122	0.00
17	Acenaphthene	1.262	1.086	13.9	121	0.00
18	Fluorene	1.608	1.379	14.2	122	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
20	4,6-Dinitro-2-methylphenol	0.093	0.075	19.4	114	0.00
21	4-Bromophenyl-phenylether	0.243	0.213	12.3	124	0.00
22	Hexachlorobenzene	0.294	0.259	11.9	121	0.00
23	Atrazine	0.232	0.204	12.1	124	0.00
24	Pentachlorophenol	0.137	0.125	8.8	133	0.00
25	Phenanthrene	1.167	0.999	14.4	123	0.00
26	Anthracene	1.081	0.923	14.6	124	0.00
27 SURR	Fluoranthene-d10	1.235	1.058	14.3	121	0.00
28	Fluoranthene	1.388	1.191	14.2	122	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
30	Pyrene	1.353	1.180	12.8	122	0.00
31 S	Terphenyl-d14	0.939	0.829	11.7	123	0.00
32	Benzo(a)anthracene	1.217	1.096	9.9	126	0.00
33	Chrysene	1.365	1.153	15.5	121	0.00
34	Bis(2-ethylhexyl)phthalate	0.797	0.683	14.3	120	0.00
35	Indeno(1,2,3-cd)pyrene	1.961	1.739	11.3	120	0.00
36 I	Perylene-d12	1.000	1.000	0.0	118	0.00
37	Benzo(b)fluoranthene	1.262	1.134	10.1	126	0.00
38	Benzo(k)fluoranthene	1.375	1.166	15.2	118	0.00
39 C	Benzo(a)pyrene	1.225	1.059	13.6	118	0.00
40	Dibenzo(a,h)anthracene	1.383	1.212	12.4	119	0.00
41	Benzo(g,h,i)perylene	1.453	1.246	14.2	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4