

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102720\
 Data File : BN012403.D
 Acq On : 27 Oct 2020 09:39
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 27 11:35:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 27 11:09:04 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	112	0.00
2	1,4-Dioxane	0.400	0.363	9.3	121	0.00
3	n-Nitrosodimethylamine	0.400	0.361	9.8	113	0.00
4 S	2-Fluorophenol	0.400	0.360	10.0	113	0.00
5 S	Phenol-d6	0.400	0.368	8.0	117	0.00
6	bis(2-Chloroethyl)ether	0.400	0.353	11.8	110	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	118	0.00
8 S	Nitrobenzene-d5	0.400	0.360	10.0	124	0.00
9	Naphthalene	0.400	0.350	12.5	117	0.00
10	Hexachlorobutadiene	0.400	0.357	10.8	116	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.364	9.0	123	0.00
12	2-Methylnaphthalene	0.400	0.360	10.0	119	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.400	0.351	12.3	127	0.00
15 S	2-Fluorobiphenyl	0.400	0.347	13.3	122	0.00
16	Acenaphthylene	0.400	0.338	15.5	122	0.00
17	Acenaphthene	0.400	0.344	14.0	121	0.00
18	Fluorene	0.400	0.343	14.2	122	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	121	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.322	19.5	114	0.00
21	4-Bromophenyl-phenylether	0.400	0.351	12.3	124	0.00
22	Hexachlorobenzene	0.400	0.353	11.8	121	0.00
23	Atrazine	0.400	0.351	12.3	124	0.00
24	Pentachlorophenol	0.400	0.365	8.8	133	0.00
25	Phenanthrene	0.400	0.342	14.5	123	0.00
26	Anthracene	0.400	0.342	14.5	124	0.00
27 SURR	Fluoranthene-d10	0.400	0.343	14.2	121	0.00
28	Fluoranthene	0.400	0.343	14.2	122	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	120	0.00
30	Pyrene	0.400	0.349	12.8	122	0.00
31 S	Terphenyl-d14	0.400	0.353	11.8	123	0.00
32	Benzo(a)anthracene	0.400	0.360	10.0	126	0.00
33	Chrysene	0.400	0.338	15.5	121	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.343	14.2	120	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.355	11.3	120	0.00
36 I	Perylene-d12	0.400	0.400	0.0	118	0.00
37	Benzo(b)fluoranthene	0.400	0.359	10.3	126	0.00
38	Benzo(k)fluoranthene	0.400	0.339	15.3	118	0.00
39 C	Benzo(a)pyrene	0.400	0.346	13.5	118	0.00
40	Dibenzo(a,h)anthracene	0.400	0.351	12.3	119	0.00
41	Benzo(g,h,i)perylene	0.400	0.343	14.2	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4