

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102720\
 Data File : BN012413.D
 Acq On : 27 Oct 2020 17:38
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 27 18:07:44 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 27 11:36:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	94	0.00
2	1,4-Dioxane	0.400	0.334	16.5	93	0.00
3	n-Nitrosodimethylamine	0.400	0.329	17.8	86	0.00
4 S	2-Fluorophenol	0.400	0.360	10.0	95	0.00
5 S	Phenol-d6	0.400	0.354	11.5	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.293	26.8#	77	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.331	17.3	92	0.01
9	Naphthalene	0.400	0.364	9.0	98	0.00
10	Hexachlorobutadiene	0.400	0.374	6.5	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.355	11.3	97	0.00
12	2-Methylnaphthalene	0.400	0.356	11.0	95	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.400	0.360	10.0	97	0.00
15 S	2-Fluorobiphenyl	0.400	0.365	8.8	96	0.00
16	Acenaphthylene	0.400	0.357	10.8	96	0.00
17	Acenaphthene	0.400	0.358	10.5	94	0.00
18	Fluorene	0.400	0.357	10.8	95	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.319	20.3	82	0.01
21	4-Bromophenyl-phenylether	0.400	0.347	13.3	90	0.00
22	Hexachlorobenzene	0.400	0.378	5.5	95	0.01
23	Atrazine	0.400	0.359	10.3	92	0.00
24	Pentachlorophenol	0.400	0.373	6.8	100	0.01
25	Phenanthrene	0.400	0.368	8.0	97	0.00
26	Anthracene	0.400	0.338	15.5	90	0.01
27 SURR	Fluoranthene-d10	0.400	0.362	9.5	94	0.00
28	Fluoranthene	0.400	0.366	8.5	95	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	86	0.01
30	Pyrene	0.400	0.384	4.0	96	0.00
31 S	Terphenyl-d14	0.400	0.379	5.3	94	0.00
32	Benzo(a)anthracene	0.400	0.373	6.8	93	0.00
33	Chrysene	0.400	0.370	7.5	94	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.383	4.3	96	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.361	9.8	87	0.00
36 I	Perylene-d12	0.400	0.400	0.0	83	0.00
37	Benzo(b)fluoranthene	0.400	0.367	8.3	90	0.00
38	Benzo(k)fluoranthene	0.400	0.369	7.8	89	0.00
39 C	Benzo(a)pyrene	0.400	0.374	6.5	90	0.00
40	Dibenzo(a,h)anthracene	0.400	0.351	12.3	84	0.00
41	Benzo(g,h,i)perylene	0.400	0.363	9.3	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4