

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102721\
 Data File : BN017131.D
 Acq On : 27 Oct 2021 17:42
 Operator : CG/JU
 Sample : SST0.129
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST0.129

Quant Time: Oct 28 01:06:16 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 28 01:03:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.531	152	5050	0.400	ng/ul	0.00
4) Naphthalene-d8	10.304	136	18592	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.183	164	12259	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.935	188	27186	0.400	ng/ul	0.00
17) Chrysene-d12	21.149	240	27880	0.400	ng/ul	0.00
23) Perylene-d12	23.350	264	29855	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/ul	
6) 2-Methylnaphthalene-d10	11.905	152	2778	0.105	ng/ul	0.00
18) Fluoranthene-d10	18.975	212	7525	0.109	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.348	128	5398	0.096	ng/ul	97
7) 2-Methylnaphthalene	11.982	142	3663	0.105	ng/ul	100
8) 1-Methylnaphthalene	12.202	142	3794	0.106	ng/ul	99
10) Acenaphthylene	13.896	152	4959	0.100	ng/ul	95
11) Acenaphthene	14.243	153	4398	0.103	ng/ul	99
12) Fluorene	15.238	166	5041	0.102	ng/ul	98
15) Phenanthrene	16.977	178	9097	0.105	ng/ul	99
16) Anthracene	17.070	178	7268	0.100	ng/ul	97
19) Fluoranthene	19.007	202	10329	0.108	ng/ul	99
20) Pyrene	19.370	202	10998	0.112	ng/ul	96
21) Benzo(a)anthracene	21.132	228	9492	0.104	ng/ul	98
22) Chrysene	21.184	228	10884	0.105	ng/ul	99
24) Benzo(b)fluoranthene	22.692	252	11143	0.099	ng/ul	91
25) Benzo(k)fluoranthene	22.736	252	12829	0.104	ng/ul#	81
26) Benzo(a)pyrene	23.253	252	10345	0.103	ng/ul#	87
27) Indeno(1,2,3-cd)pyrene	25.568	276	14549	0.102	ng/ul#	93
28) Dibenzo(a,h)anthracene	25.592	278	11406	0.101	ng/ul#	92
29) Benzo(g,h,i)perylene	26.246	276	12217	0.099	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102721\
Data File : BN017131.D
Acq On : 27 Oct 2021 17:42
Operator : CG/JU
Sample : SST0.129
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.1229

Quant Time: Oct 28 01:06:16 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 28 01:03:15 2021
Response via : Initial Calibration

