

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102721\
 Data File : BN017169.D
 Acq On : 28 Oct 2021 23:50
 Operator : CG/JU
 Sample : SP5734
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP5734

Quant Time: Oct 29 03:15:19 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 28 12:50:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.531	152	6172	0.400	ng/ul	0.00
4) Naphthalene-d8	10.299	136	25334	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.183	164	18435	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.935	188	36605	0.400	ng/ul	0.00
17) Chrysene-d12	21.146	240	30194	0.400	ng/ul	0.00
23) Perylene-d12	23.347	264	41439	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.055	96	98622	14.372	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.905	152	32488	0.878	ng/ul	0.00
18) Fluoranthene-d10	18.975	212	82718	1.045	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102721\
Data File : BN017169.D
Acq On : 28 Oct 2021 23:50
Operator : CG/JU
Sample : SP5734
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP5734

Quant Time: Oct 29 03:15:19 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 28 12:50:34 2021
Response via : Initial Calibration

