

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102721\
 Data File : BN017137.D
 Acq On : 27 Oct 2021 21:21
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV235

Quant Time: Oct 28 01:22:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 28 01:21:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.531	152	5522	0.400	ng/ul	0.00
4) Naphthalene-d8	10.304	136	20891	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.183	164	12814	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.939	188	27902	0.400	ng/ul	0.00
17) Chrysene-d12	21.149	240	27995	0.400	ng/ul	0.00
23) Perylene-d12	23.350	264	28168	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.059	96	2659	0.433	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.905	152	12495	0.410	ng/ul	0.00
18) Fluoranthene-d10	18.979	212	31559	0.430	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.092	88	2854	0.433	ng/ul#	87
5) Naphthalene	10.348	128	24120	0.417	ng/ul	100
7) 2-Methylnaphthalene	11.982	142	16390	0.409	ng/ul	100
8) 1-Methylnaphthalene	12.201	142	16849	0.411	ng/ul	100
10) Acenaphthylene	13.896	152	20895	0.412	ng/ul	100
11) Acenaphthene	14.243	153	18735	0.421	ng/ul	99
12) Fluorene	15.243	166	21716	0.418	ng/ul	100
14) Pentachlorophenol	16.593	266	2080	0.331	ng/ul	99
15) Phenanthrene	16.977	178	37730	0.426	ng/ul	99
16) Anthracene	17.070	178	30757	0.419	ng/ul	100
19) Fluoranthene	19.007	202	43267	0.432	ng/ul	100
20) Pyrene	19.374	202	43888	0.427	ng/ul	100
21) Benzo(a)anthracene	21.134	228	37764	0.412	ng/ul	100
22) Chrysene	21.184	228	44981	0.433	ng/ul	100
24) Benzo(b)fluoranthene	22.692	252	44214	0.421	ng/ul	99
25) Benzo(k)fluoranthene	22.736	252	51872	0.436	ng/ul	99
26) Benzo(a)pyrene	23.256	252	39297	0.417	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	25.565	276	56708	0.424	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.588	278	44673	0.423	ng/ul	99
29) Benzo(g,h,i)perylene	26.242	276	48126	0.426	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102721\
Data File : BN017137.D
Acq On : 27 Oct 2021 21:21
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SICV235

Quant Time: Oct 28 01:22:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 28 01:21:35 2021
Response via : Initial Calibration

