

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102721\
 Data File : BN017168.D
 Acq On : 28 Oct 2021 23:14
 Operator : CG/JU
 Sample : PB140134BL
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK134

Quant Time: Oct 29 03:14:18 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 28 12:50:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.531	152	5807	0.400	ng/ul	0.00
4) Naphthalene-d8	10.300	136	22791	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.180	164	14372	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.936	188	30980	0.400	ng/ul	0.00
17) Chrysene-d12	21.148	240	25185	0.400	ng/ul	0.00
23) Perylene-d12	23.346	264	36136	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.055	96	29994	4.646	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.906	152	9071	0.273	ng/ul	0.00
18) Fluoranthene-d10	18.976	212	25595	0.388	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102721\
Data File : BN017168.D
Acq On : 28 Oct 2021 23:14
Operator : CG/JU
Sample : PB140134BL
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK134

Quant Time: Oct 29 03:14:18 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 28 12:50:34 2021
Response via : Initial Calibration

