

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102722\
 Data File : BN022351.D
 Acq On : 27 Oct 2022 01:55
 Operator : CG/JU
 Sample : N5236-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 LSP-GW-03

Quant Time: Oct 27 08:31:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 27 07:14:22 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 10/27/2022
 Supervised By : Jagrut Upadhyay 10/28/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.942	152	3753	0.400	ng	0.00
7) Naphthalene-d8	10.763	136	12097	0.400	ng	#-0.01
13) Acenaphthene-d10	14.610	164	6887	0.400	ng	0.00
19) Phenanthrene-d10	17.362	188	14501	0.400	ng	0.00
29) Chrysene-d12	21.564	240	9847	0.400	ng	# 0.00
35) Perylene-d12	24.029	264	6998	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.472	112	1496	0.150	ng	0.00
5) Phenol-d6	7.104	99	1241	0.099	ng	0.00
8) Nitrobenzene-d5	9.118	82	3152m	0.301	ng	0.00
11) 2-Methylnaphthalene-d10	12.361	152	5625	0.296	ng	0.00
14) 2,4,6-Tribromophenol	16.109	330	1171	0.396	ng	0.00
15) 2-Fluorobiphenyl	13.231	172	9836	0.334	ng	-0.01
27) Fluoranthene-d10	19.403	212	14743	0.368	ng	0.00
31) Terphenyl-d14	20.001	244	14299	0.557	ng	0.00
Target Compounds						
17) Acenaphthene	14.675	154	4634	0.204	ng	99
28) Fluoranthene	19.432	202	12381	0.230	ng	100
30) Pyrene	19.799	202	19550	0.431	ng	100
34) Bis(2-ethylhexyl)phtha...	21.466	149	11536	0.421	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102722\
Data File : BN022351.D
Acq On : 27 Oct 2022 01:55
Operator : CG/JU
Sample : N5236-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
LSP-GW-03

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 10/27/2022
Supervised By : Jagrut Upadhyay 10/28/2022

Quant Time: Oct 27 08:31:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 27 07:14:22 2022
Response via : Initial Calibration

