




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102722\
Data File : BN022338.D
Acq On : 26 Oct 2022 16:59
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 28 09:00:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 28 09:00:04 2022
Response via : Initial Calibration



TIC: BN022338.D\data.ms

(70) Pentachlorophenol (C)




17.022min (0.000) 26904.37 ng

response 441280

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.93
264.00	61.60	63.39
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102722\
Data File : BN022338.D
Acq On : 26 Oct 2022 16:59
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 28 09:00:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 28 09:00:04 2022
Response via : Initial Calibration



TIC: BN022338.D\data.ms

(77) Benzidine

19.633min (0.000) 0.00 ng

response 2380061

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.21
183.00	13.20	11.69
0.00	0.00	0.00