

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102722\
 Data File : BN022346.D
 Acq On : 26 Oct 2022 22:08
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102722

Quant Time: Oct 27 07:21:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 27 07:14:22 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.560	0.559	0.2	109	0.00
3	n-Nitrosodimethylamine	0.596	0.570	4.4	109	0.00
4 S	2-Fluorophenol	1.066	1.012	5.1	107	0.00
5 S	Phenol-d6	1.336	1.249	6.5	108	0.00
6	bis(2-Chloroethyl)ether	1.283	1.249	2.7	110	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
8 S	Nitrobenzene-d5	0.346	0.321	7.2	109	0.00
9	Naphthalene	1.196	1.169	2.3	111	0.00
10	Hexachlorobutadiene	0.203	0.201	1.0	111	-0.01
11 SURR	2-Methylnaphthalene-d10	0.628	0.598	4.8	111	0.00
12	2-Methylnaphthalene	0.238	0.202	15.1	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.172	0.154	10.5	110	0.00
15 S	2-Fluorobiphenyl	1.711	1.668	2.5	109	0.00
16	Acenaphthylene	1.938	1.776	8.4	111	0.00
17	Acenaphthene	1.321	1.272	3.7	110	0.00
18	Fluorene	1.841	1.763	4.2	108	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
20	4,6-Dinitro-2-methylphenol	0.064	0.049	23.4	105	0.00
21	4-Bromophenyl-phenylether	0.241	0.229	5.0	111	0.00
22	Hexachlorobenzene	0.276	0.270	2.2	112	0.00
23	Atrazine	0.202	0.178	11.9	106	0.00
24	Pentachlorophenol	0.114	0.087	23.7	102	0.00
25	Phenanthrene	1.372	1.305	4.9	110	0.00
26	Anthracene	1.155	1.052	8.9	110	0.00
27 SURR	Fluoranthene-d10	1.105	1.060	4.1	111	0.00
28	Fluoranthene	1.487	1.432	3.7	110	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
30	Pyrene	1.841	1.848	-0.4	110	0.00
31 S	Terphenyl-d14	1.042	1.073	-3.0	114	0.00
32	Benzo(a)anthracene	1.575	1.477	6.2	103	0.00
33	Chrysene	1.646	1.619	1.6	108	0.00
34	Bis(2-ethylhexyl)phthalate	1.207	1.404	-16.3	120	0.00
35 I	Perylene-d12	1.000	1.000	0.0	97	0.00
36	Indeno(1,2,3-cd)pyrene	1.976	1.780	9.9	85	0.00
37	Benzo(b)fluoranthene	1.831	1.821	0.5	100	0.00
38	Benzo(k)fluoranthene	1.821	1.740	4.4	98	0.00
39 C	Benzo(a)pyrene	1.490	1.425	4.4	96	0.00
40	Dibenzo(a,h)anthracene	1.543	1.384	10.3	85	0.00
41	Benzo(g,h,i)perylene	1.628	1.458	10.4	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN102722