

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102722\
 Data File : BN022346.D
 Acq On : 26 Oct 2022 22:08
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102722

Quant Time: Oct 27 07:21:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 27 07:14:22 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	0.00
2	1,4-Dioxane	0.400	0.399	0.3	109	0.00
3	n-Nitrosodimethylamine	0.400	0.383	4.3	109	0.00
4 S	2-Fluorophenol	0.400	0.380	5.0	107	0.00
5 S	Phenol-d6	0.400	0.374	6.5	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.390	2.5	110	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	110	0.00
8 S	Nitrobenzene-d5	0.400	0.372	7.0	109	0.00
9	Naphthalene	0.400	0.391	2.3	111	0.00
10	Hexachlorobutadiene	0.400	0.396	1.0	111	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.381	4.8	111	0.00
12	2-Methylnaphthalene	0.400	0.339	15.3	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.400	0.359	10.3	110	0.00
15 S	2-Fluorobiphenyl	0.400	0.390	2.5	109	0.00
16	Acenaphthylene	0.400	0.367	8.3	111	0.00
17	Acenaphthene	0.400	0.385	3.8	110	0.00
18	Fluorene	0.400	0.383	4.3	108	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	111	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.306	23.5	105	0.00
21	4-Bromophenyl-phenylether	0.400	0.379	5.3	111	0.00
22	Hexachlorobenzene	0.400	0.390	2.5	112	0.00
23	Atrazine	0.400	0.353	11.8	106	0.00
24	Pentachlorophenol	0.400	0.304	24.0	102	0.00
25	Phenanthrene	0.400	0.381	4.8	110	0.00
26	Anthracene	0.400	0.364	9.0	110	0.00
27 SURR	Fluoranthene-d10	0.400	0.384	4.0	111	0.00
28	Fluoranthene	0.400	0.385	3.8	110	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.402	-0.5	110	0.00
31 S	Terphenyl-d14	0.400	0.412	-3.0	114	0.00
32	Benzo(a)anthracene	0.400	0.375	6.3	103	0.00
33	Chrysene	0.400	0.393	1.8	108	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.518	-29.5#	120	0.00
35 I	Perylene-d12	0.400	0.400	0.0	97	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.360	10.0	85	0.00
37	Benzo(b)fluoranthene	0.400	0.398	0.5	100	0.00
38	Benzo(k)fluoranthene	0.400	0.382	4.5	98	0.00
39 C	Benzo(a)pyrene	0.400	0.383	4.3	96	0.00
40	Dibenzo(a,h)anthracene	0.400	0.359	10.3	85	0.00
41	Benzo(g,h,i)perylene	0.400	0.358	10.5	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN102722