

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102820\
 Data File : BN012415.D
 Acq On : 28 Oct 2020 09:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 28 10:29:32 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 27 11:36:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.651	0.574	11.8	92	0.00
3	n-Nitrosodimethylamine	1.581	1.428	9.7	89	0.00
4 S	2-Fluorophenol	1.398	1.215	13.1	86	0.00
5 S	Phenol-d6	1.655	1.431	13.5	86	0.00
6	bis(2-Chloroethyl)ether	1.127	1.076	4.5	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.441	0.381	13.6	92	0.00
9	Naphthalene	1.120	1.014	9.5	94	0.00
10	Hexachlorobutadiene	0.237	0.221	6.8	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.643	0.593	7.8	97	0.00
12	2-Methylnaphthalene	0.727	0.657	9.6	93	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.284	0.246	13.4	96	0.00
15 S	2-Fluorobiphenyl	1.488	1.314	11.7	95	0.00
16	Acenaphthylene	1.975	1.675	15.2	93	0.00
17	Acenaphthene	1.262	1.090	13.6	93	0.00
18	Fluorene	1.608	1.407	12.5	95	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
20	4,6-Dinitro-2-methylphenol	0.093	0.081	12.9	95	0.00
21	4-Bromophenyl-phenylether	0.243	0.208	14.4	94	0.00
22	Hexachlorobenzene	0.294	0.272	7.5	98	0.00
23	Atrazine	0.232	0.200	13.8	94	0.00
24	Pentachlorophenol	0.137	0.117	14.6	97	0.00
25	Phenanthrene	1.167	1.008	13.6	96	0.00
26	Anthracene	1.081	0.882	18.4	92	0.00
27 SURR	Fluoranthene-d10	1.235	1.060	14.2	94	0.00
28	Fluoranthene	1.388	1.189	14.3	94	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
30	Pyrene	1.353	1.205	10.9	94	0.00
31 S	Terphenyl-d14	0.939	0.838	10.8	95	0.00
32	Benzo(a)anthracene	1.217	1.066	12.4	93	0.00
33	Chrysene	1.365	1.184	13.3	94	0.00
34	Bis(2-ethylhexyl)phthalate	0.797	0.702	11.9	94	0.00
35	Indeno(1,2,3-cd)pyrene	1.961	1.696	13.5	89	0.00
36 I	Perylene-d12	1.000	1.000	0.0	89	0.00
37	Benzo(b)fluoranthene	1.262	1.081	14.3	91	0.00
38	Benzo(k)fluoranthene	1.375	1.196	13.0	91	0.00
39 C	Benzo(a)pyrene	1.225	1.065	13.1	89	0.00
40	Dibenzo(a,h)anthracene	1.383	1.152	16.7	85	0.00
41	Benzo(g,h,i)perylene	1.453	1.226	15.6	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4