

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102820\
 Data File : BN012415.D
 Acq On : 28 Oct 2020 09:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 28 10:29:32 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 27 11:36:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	0.00
2	1,4-Dioxane	0.400	0.352	12.0	92	0.00
3	n-Nitrosodimethylamine	0.400	0.361	9.8	89	0.00
4 S	2-Fluorophenol	0.400	0.348	13.0	86	0.00
5 S	Phenol-d6	0.400	0.346	13.5	86	0.00
6	bis(2-Chloroethyl)ether	0.400	0.382	4.5	94	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	92	0.00
8 S	Nitrobenzene-d5	0.400	0.346	13.5	92	0.00
9	Naphthalene	0.400	0.362	9.5	94	0.00
10	Hexachlorobutadiene	0.400	0.373	6.8	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.369	7.8	97	0.00
12	2-Methylnaphthalene	0.400	0.362	9.5	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.400	0.346	13.5	96	0.00
15 S	2-Fluorobiphenyl	0.400	0.353	11.8	95	0.00
16	Acenaphthylene	0.400	0.339	15.3	93	0.00
17	Acenaphthene	0.400	0.346	13.5	93	0.00
18	Fluorene	0.400	0.350	12.5	95	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.347	13.3	95	0.00
21	4-Bromophenyl-phenylether	0.400	0.342	14.5	94	0.00
22	Hexachlorobenzene	0.400	0.370	7.5	98	0.00
23	Atrazine	0.400	0.345	13.8	94	0.00
24	Pentachlorophenol	0.400	0.343	14.2	97	0.00
25	Phenanthrene	0.400	0.346	13.5	96	0.00
26	Anthracene	0.400	0.326	18.5	92	0.00
27 SURR	Fluoranthene-d10	0.400	0.343	14.2	94	0.00
28	Fluoranthene	0.400	0.343	14.2	94	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	91	0.00
30	Pyrene	0.400	0.356	11.0	94	0.00
31 S	Terphenyl-d14	0.400	0.357	10.8	95	0.00
32	Benzo(a)anthracene	0.400	0.350	12.5	93	0.00
33	Chrysene	0.400	0.347	13.3	94	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.352	12.0	94	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.346	13.5	89	0.00
36 I	Perylene-d12	0.400	0.400	0.0	89	0.00
37	Benzo(b)fluoranthene	0.400	0.343	14.2	91	0.00
38	Benzo(k)fluoranthene	0.400	0.348	13.0	91	0.00
39 C	Benzo(a)pyrene	0.400	0.348	13.0	89	0.00
40	Dibenzo(a,h)anthracene	0.400	0.333	16.8	85	0.00
41	Benzo(g,h,i)perylene	0.400	0.338	15.5	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4