

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102820\
 Data File : BN012429.D
 Acq On : 28 Oct 2020 19:35
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 29 01:25:04 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 28 10:32:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	108	0.00
2	1,4-Dioxane	0.400	0.359	10.3	116	0.00
3	n-Nitrosodimethylamine	0.400	0.352	12.0	106	0.00
4 S	2-Fluorophenol	0.400	0.380	5.0	115	0.00
5 S	Phenol-d6	0.400	0.356	11.0	109	0.00
6	bis(2-Chloroethyl)ether	0.400	0.305	23.8	92	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	111	0.00
8 S	Nitrobenzene-d5	0.400	0.349	12.8	113	0.00
9	Naphthalene	0.400	0.346	13.5	109	-0.01
10	Hexachlorobutadiene	0.400	0.343	14.2	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.349	12.8	111	0.00
12	2-Methylnaphthalene	0.400	0.353	11.8	111	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.400	0.341	14.8	107	0.00
15 S	2-Fluorobiphenyl	0.400	0.355	11.3	109	-0.01
16	Acenaphthylene	0.400	0.341	14.8	106	0.00
17	Acenaphthene	0.400	0.347	13.3	106	0.00
18	Fluorene	0.400	0.339	15.3	105	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	105	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.320	20.0	98	-0.01
21	4-Bromophenyl-phenylether	0.400	0.339	15.3	104	-0.01
22	Hexachlorobenzene	0.400	0.337	15.8	100	0.00
23	Atrazine	0.400	0.332	17.0	101	-0.01
24	Pentachlorophenol	0.400	0.317	20.8	100	-0.01
25	Phenanthrene	0.400	0.339	15.3	105	-0.01
26	Anthracene	0.400	0.335	16.3	105	0.00
27 SURR	Fluoranthene-d10	0.400	0.343	14.2	105	0.00
28	Fluoranthene	0.400	0.345	13.8	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	106	0.00
30	Pyrene	0.400	0.346	13.5	106	0.00
31 S	Terphenyl-d14	0.400	0.337	15.8	103	0.00
32	Benzo(a)anthracene	0.400	0.349	12.8	107	0.00
33	Chrysene	0.400	0.333	16.8	105	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.365	8.8	112	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.358	10.5	106	0.00
36 I	Perylene-d12	0.400	0.400	0.0	106	0.00
37	Benzo(b)fluoranthene	0.400	0.334	16.5	104	0.00
38	Benzo(k)fluoranthene	0.400	0.336	16.0	104	0.00
39 C	Benzo(a)pyrene	0.400	0.349	12.8	106	0.00
40	Dibenzo(a,h)anthracene	0.400	0.337	15.8	102	0.00
41	Benzo(g,h,i)perylene	0.400	0.346	13.5	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4