

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102920\
 Data File : BN012438.D
 Acq On : 29 Oct 2020 15:08
 Operator : CG/JU
 Sample : L4511-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 OB-9

Quant Time: Oct 29 15:39:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 28 10:32:42 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.578	152	673	0.40	ng	# 0.00
7) Naphthalene-d8	10.352	136	2772	0.40	ng	0.00
13) Acenaphthene-d10	14.217	164	1638	0.40	ng	0.00
19) Phenanthrene-d10	16.968	188	3309	0.40	ng	# 0.00
29) Chrysene-d12	21.174	240	3728	0.40	ng	# 0.00
36) Perylene-d12	23.350	264	4359	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.194	112	212	0.09	ng	0.00
5) Phenol-d6	6.765	99	156	0.06	ng	0.00
8) Nitrobenzene-d5	8.729	82	902	0.30	ng	0.00
11) 2-Methylnaphthalene-d10	11.953	152	1109	0.25	ng	0.00
14) 2,4,6-Tribromophenol	15.727	330	374	0.32	ng	0.00
15) 2-Fluorobiphenyl	12.835	172	2040	0.33	ng	-0.01
27) Fluoranthene-d10	19.008	212	3437	0.34	ng	0.00
31) Terphenyl-d14	19.618	244	4301	0.49	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.246	88	33	0.03	ng	# 1
3) n-Nitrosodimethylamine	3.447	42	76	0.03	ng	# 5
9) Naphthalene	10.402	128	183	0.02	ng	# 4
12) 2-Methylnaphthalene	12.024	142	136	0.03	ng	# 80
16) Acenaphthylene	13.938	152	255	0.03	ng	95
17) Acenaphthene	14.281	154	183	0.04	ng	89
18) Fluorene	15.271	166	285	0.04	ng	# 98
20) 4,6-Dinitro-2-methylph...	15.397	198	32	0.04	ng	# 1
21) 4-Bromophenyl-phenylether	16.177	248	84	0.04	ng	# 86
22) Hexachlorobenzene	16.275	284	115	0.05	ng	# 83
23) Atrazine	16.445	200	41	0.02	ng	# 25
24) Pentachlorophenol	16.628	266	79	0.07	ng	# 85
25) Phenanthrene	17.017	178	525	0.05	ng	96
26) Anthracene	17.102	178	424	0.05	ng	97
28) Fluoranthene	19.043	202	618	0.05	ng	# 87
30) Pyrene	19.405	202	656	0.05	ng	# 94
32) Benzo(a)anthracene	21.163	228	570	0.05	ng	# 80
33) Chrysene	21.216	228	539	0.04	ng	# 79
34) Bis(2-ethylhexyl)phtha...	21.099	149	2543	0.34	ng	# 90
35) Indeno(1,2,3-cd)pyrene	25.517	276	757	0.04	ng	# 89
37) Benzo(b)fluoranthene	22.707	252	721	0.05	ng	# 1
38) Benzo(k)fluoranthene	22.748	252	644	0.04	ng	# 1
39) Benzo(a)pyrene	23.255	252	615	0.05	ng	# 1
40) Dibenzo(a,h)anthracene	25.527	278	616	0.04	ng	# 1
41) Benzo(g,h,i)perylene	26.174	276	747	0.05	ng	# 55





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.578 min Scan# 571
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion:152 Resp: 673
Ion Ratio Lower Upper
152 100
150 149.0 116.6 174.8
115 79.4 52.9 79.3#



#2
1,4-Dioxane
Concen: 0.03 ng
RT: 3.246 min Scan# 33
Delta R.T. 0.073 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion: 88 Resp: 33
Ion Ratio Lower Upper
88 100
58 0.0 59.4 89.2#
43 221.2 32.3 48.5#









#3
n-Nitrosodimethylamine
Concen: 0.03 ng
RT: 3.447 min Scan# 58
Delta R.T. -0.040 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion: 42 Resp: 76
Ion Ratio Lower Upper
42 100
74 0.0 64.4 96.6#
44 61.8 2.8 4.2#



#4
2-Fluorophenol
Concen: 0.09 ng
RT: 5.194 min Scan# 275
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion: 112 Resp: 212
Ion Ratio Lower Upper
112 100
64 74.1 49.5 74.3
63 62.3 32.0 48.0#









#5
Phenol-d6
Concen: 0.06 ng
RT: 6.765 min Scan# 470
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion:	99	Resp:	156
Ion	Ratio	Lower	Upper
99	100		
42	99.4	22.4	33.6#
71	74.4	37.8	56.6#



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.352 min Scan# 823
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:	136	Resp:	2772
Ion	Ratio	Lower	Upper
136	100		
137	14.1	10.6	15.8
54	12.4	8.6	12.8
68	7.3	5.7	8.5









#8
Nitrobenzene-d5
Concen: 0.30 ng
RT: 8.729 min Scan# 695
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion: 82 Resp: 902
Ion Ratio Lower Upper
82 100
128 42.7 30.6 46.0
54 80.3 48.3 72.5#



#9
Naphthalene
Concen: 0.02 ng
RT: 10.402 min Scan# 827
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion: 128 Resp: 183
Ion Ratio Lower Upper
128 100
129 54.2 9.8 14.8#
127 51.4 12.2 18.4#









#11
2-Methylnaphthalene-d10
Concen: 0.25 ng
RT: 11.953 min Scan# 1013
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion:152 Resp: 1109
Ion Ratio Lower Upper
152 100
151 22.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.03 ng
RT: 12.024 min Scan# 1029
Delta R.T. -0.004 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:142 Resp: 136
Ion Ratio Lower Upper
142 100
141 85.6 69.4 104.2
115 81.7 35.7 53.5#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.217 min Scan# 1272
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampleId :
OB-9


Tgt Ion:	164	Resp:	1638
Ion	Ratio	Lower	Upper
164	100		
162	101.3	80.6	120.8
160	50.0	39.5	59.3



#14
2,4,6-Tribromophenol
Concen: 0.32 ng
RT: 15.727 min Scan# 1391
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:	330	Resp:	374
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	42.2	34.4	51.6









#15
2-Fluorobiphenyl
Concen: 0.33 ng
RT: 12.835 min Scan# 1163
Delta R.T. -0.012 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampleId :
OB-9


Tgt Ion:	172	Resp:	2040
Ion Ratio	Lower	Upper	
172	100		
171	38.1	28.2	42.4
170	29.3	20.0	30.0



#16
Acenaphthylene
Concen: 0.03 ng
RT: 13.938 min Scan# 1250
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:	152	Resp:	255
Ion Ratio	Lower	Upper	
152	100		
151	22.0	15.8	23.8
153	15.7	10.6	16.0









#17
Acenaphthene
Concen: 0.04 ng
RT: 14.281 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion:	154	Resp:	183
Ion	Ratio	Lower	Upper
154	100		
153	115.8	83.6	125.4
152	61.7	43.9	65.9



#18
Fluorene
Concen: 0.04 ng
RT: 15.271 min Scan# 1355
Delta R.T. -0.012 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:	166	Resp:	285
Ion	Ratio	Lower	Upper
166	100		
165	97.5	78.6	118.0
167	18.6	10.7	16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.968 min Scan# 1492
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion:	188	Resp:	3309
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	10.4	5.0	7.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.04 ng
RT: 15.397 min Scan# 1365
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:	198	Resp:	32
Ion Ratio	Lower	Upper	
198	100		
51	245.5	43.5	65.3#
105	132.7	27.2	40.8#









#21
4-Bromophenyl-phenylether
Concen: 0.04 ng
RT: 16.177 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion:248 Resp: 84
Ion Ratio Lower Upper
248 100
250 104.8 77.3 115.9
141 90.5 57.8 86.8#



#22
Hexachlorobenzene
Concen: 0.05 ng
RT: 16.275 min Scan# 1435
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:284 Resp: 115
Ion Ratio Lower Upper
284 100
142 53.0 32.0 48.0#
249 40.9 27.0 40.6#





#23

Atrazine

Concen: 0.02 ng

RT: 16.445 min Scan# 1449

Delta R.T. -0.012 min

Lab File: BN012438.D

Acq: 29 Oct 2020 15:08

Instrument :





BNA_N

ClientSampleId :

OB-9

Tgt Ion:200 Resp: 41

Ion	Ratio	Lower	Upper
200	100		
173	86.2	22.8	34.2#
215	84.6	38.1	57.1#



#24

Pentachlorophenol

Concen: 0.07 ng

RT: 16.628 min Scan# 1464




Delta R.T. -0.012 min


Lab File: BN012438.D

Acq: 29 Oct 2020 15:08

Tgt Ion:266 Resp: 79

Ion	Ratio	Lower	Upper
266	100		
264	81.0	50.9	76.3#
268	70.9	51.5	77.3









#25
Phenanthrene
Concen: 0.05 ng
RT: 17.017 min Scan# 1496
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion:	178	Resp:	525
Ion	Ratio	Lower	Upper
178	100		
176	20.8	14.8	22.2
179	16.4	12.4	18.6



#26
Anthracene
Concen: 0.05 ng
RT: 17.102 min Scan# 1503
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:	178	Resp:	424
Ion	Ratio	Lower	Upper
178	100		
176	20.3	14.8	22.2
179	15.1	12.5	18.7





#27

Fluoranthene-d10

Concen: 0.34 ng

RT: 19.008 min Scan# 1778

Delta R.T. -0.005 min

Lab File: BN012438.D

Acq: 29 Oct 2020 15:08





Tgt Ion:212	Resp:	3437
Ion Ratio	Lower	Upper
212	100	
106	8.9	6.8 10.2
104	5.0	4.1 6.1

Instrument :

BNA_N

ClientSampled :

OB-9



#28

Fluoranthene

Concen: 0.05 ng




RT: 19.043 min Scan# 1785


Delta R.T. 0.000 min

Lab File: BN012438.D

Acq: 29 Oct 2020 15:08

Tgt Ion:202	Resp:	618
Ion Ratio	Lower	Upper
202	100	
101	18.4	5.8 8.6#
203	18.9	13.7 20.5





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.174 min Scan# 2089

Delta R.T. 0.000 min

Lab File: BN012438.D

Acq: 29 Oct 2020 15:08





Instrument :

BNA_N

ClientSampled :

OB-9

Tgt Ion:240	Resp:	3728
Ion Ratio	Lower	Upper
240	100	
120	13.0	5.3 7.9#
236	28.9	21.8 32.8



#30

Pyrene

Concen: 0.05 ng




RT: 19.405 min Scan# 1858


Delta R.T. -0.005 min

Lab File: BN012438.D

Acq: 29 Oct 2020 15:08


Tgt Ion:202	Resp:	656
Ion Ratio	Lower	Upper
202	100	
200	21.3	16.6 24.8
203	22.6	13.8 20.8#








#31
Terphenyl-d14
Concen: 0.49 ng
RT: 19.618 min Scan# 1901
Delta R.T. -0.005 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Instrument :
BNA_N
ClientSampleId :
OB-9







Tgt Ion:244 Resp: 4301
Ion Ratio Lower Upper
244 100
212 10.5 7.3 10.9
122 10.5 7.3 10.9



#32
Benzo(a)anthracene
Concen: 0.05 ng
RT: 21.163 min Scan# 2088
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:228 Resp: 570
Ion Ratio Lower Upper
228 100
226 35.7 22.3 33.5#
229 32.5 15.7 23.5#









#33
Chrysene
Concen: 0.04 ng
RT: 21.216 min Scan# 2093
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion:228 Resp: 539
Ion Ratio Lower Upper
228 100
226 40.1 24.2 36.2#
229 31.2 15.7 23.5#



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.34 ng
RT: 21.099 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08


Tgt Ion:149 Resp: 2543
Ion Ratio Lower Upper
149 100
167 23.2 23.2 34.8#
279 4.3 3.9 5.9








#35
Indeno(1,2,3-cd)pyrene
Concen: 0.04 ng
RT: 25.517 min Scan# 3114
Delta R.T. 0.007 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Instrument :
BNA_N
ClientSampled :
OB-9







Tgt Ion:276 Resp: 757
Ion Ratio Lower Upper
276 100
138 7.7 13.3 19.9#
277 27.6 20.2 30.4



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.350 min Scan# 2481
Delta R.T. -0.003 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:264 Resp: 4359
Ion Ratio Lower Upper
264 100
260 25.4 19.8 29.6
265 67.9 51.4 77.0









#37
Benzo(b)fluoranthene
Concen: 0.05 ng
RT: 22.707 min Scan# 2293
Delta R.T. 0.004 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08




Instrument :
BNA_N
ClientSampled :
OB-9


Tgt Ion:252 Resp: 721
Ion Ratio Lower Upper
252 100
253 79.5 20.1 30.1#
125 52.9 7.2 10.8#



#38
Benzo(k)fluoranthene
Concen: 0.04 ng
RT: 22.748 min Scan# 2305
Delta R.T. 0.000 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Tgt Ion:252 Resp: 644
Ion Ratio Lower Upper
252 100
253 77.6 19.8 29.8#
125 48.5 7.1 10.7#





#39

Benzo(a)pyrene

Concen: 0.05 ng

RT: 23.255 min Scan# 2453

Delta R.T. -0.007 min

Lab File: BN012438.D

Acq: 29 Oct 2020 15:08

Tgt Ion:252 Resp: 615

Ion Ratio Lower Upper

252 100

253 89.7 21.3 31.9#





125 56.3 8.3 12.5#

Instrument :

BNA_N

ClientSampleId :

OB-9



#40

Dibenzo(a,h)anthracene

Concen: 0.04 ng

RT: 25.527 min Scan# 3117

Delta R.T. 0.000 min

Lab File: BN012438.D

Acq: 29 Oct 2020 15:08




Tgt Ion:278 Resp: 616


Ion Ratio Lower Upper

278 100

139 43.3 9.3 13.9#

279 91.8 21.6 32.4#





#41
Benzo(g,h,i)perylene
Concen: 0.05 ng
RT: 26.174 min Scan# 3306
Delta R.T. 0.004 min
Lab File: BN012438.D
Acq: 29 Oct 2020 15:08

Instrument :
BNA_N
ClientSampleId :
OB-9

Tgt Ion:276 Resp: 747
Ion Ratio Lower Upper
276 100
277 45.9 19.8 29.8#
138 34.6 11.7 17.5#

