

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102920\
 Data File : BN012454.D
 Acq On : 30 Oct 2020 02:01
 Operator : CG/JU
 Sample : SP5332
 Misc : 8270 SIM SURROGATE
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP5332

Quant Time: Oct 30 10:26:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 28 10:32:42 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.578	152	779	0.40	ng	# 0.00
7) Naphthalene-d8	10.352	136	3260	0.40	ng	# 0.00
13) Acenaphthene-d10	14.217	164	1933	0.40	ng	0.00
19) Phenanthrene-d10	16.968	188	3689	0.40	ng	# 0.00
29) Chrysene-d12	21.174	240	3556	0.40	ng	# 0.00
36) Perylene-d12	23.343	264	4229	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.186	112	995	0.37	ng	0.00
5) Phenol-d6	6.765	99	1043	0.32	ng	0.00
8) Nitrobenzene-d5	8.742	82	1547	0.43	ng	0.01
11) 2-Methylnaphthalene-d10	11.953	152	1618	0.31	ng	0.00
14) 2,4,6-Tribromophenol	15.727	330	394	0.29	ng	0.00
15) 2-Fluorobiphenyl	12.834	172	3282	0.46	ng	-0.01
27) Fluoranthene-d10	19.013	212	3711	0.33	ng	0.00
31) Terphenyl-d14	19.618	244	4840	0.58	ng	0.00

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102920\
Data File : BN012454.D
Acq On : 30 Oct 2020 02:01
Operator : CG/JU
Sample : SP5332
Misc : 8270 SIM SURROGATE
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP5332

Quant Time: Oct 30 10:26:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 28 10:32:42 2020
Response via : Initial Calibration

