

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102920\
 Data File : BN012431.D
 Acq On : 29 Oct 2020 10:54
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 29 12:04:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 28 10:32:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.651	0.620	4.8	111	0.00
3	n-Nitrosodimethylamine	1.581	1.352	14.5	94	0.00
4 S	2-Fluorophenol	1.398	1.247	10.8	98	0.00
5 S	Phenol-d6	1.655	1.456	12.0	98	0.00
6	bis(2-Chloroethyl)ether	1.127	0.938	16.8	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	97	-0.01
8 S	Nitrobenzene-d5	0.441	0.380	13.8	98	0.00
9	Naphthalene	1.120	1.018	9.1	100	-0.01
10	Hexachlorobutadiene	0.237	0.219	7.6	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.643	0.569	11.5	98	0.00
12	2-Methylnaphthalene	0.727	0.652	10.3	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
14 S	2,4,6-Tribromophenol	0.284	0.247	13.0	96	0.00
15 S	2-Fluorobiphenyl	1.488	1.363	8.4	99	-0.01
16	Acenaphthylene	1.975	1.772	10.3	99	0.00
17	Acenaphthene	1.262	1.133	10.2	97	0.00
18	Fluorene	1.608	1.454	9.6	99	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
20	4,6-Dinitro-2-methylphenol	0.093	0.078	16.1	91	-0.01
21	4-Bromophenyl-phenylether	0.243	0.209	14.0	94	0.00
22	Hexachlorobenzene	0.294	0.263	10.5	94	0.00
23	Atrazine	0.232	0.208	10.3	96	-0.01
24	Pentachlorophenol	0.137	0.116	15.3	95	-0.01
25	Phenanthrene	1.167	1.024	12.3	96	-0.01
26	Anthracene	1.081	0.922	14.7	94	0.00
27 SURR	Fluoranthene-d10	1.235	1.103	10.7	97	0.00
28	Fluoranthene	1.388	1.235	11.0	97	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	90	0.01
30	Pyrene	1.353	1.252	7.5	97	0.00
31 S	Terphenyl-d14	0.939	0.838	10.8	94	0.00
32	Benzo(a)anthracene	1.217	1.086	10.8	93	0.00
33	Chrysene	1.365	1.212	11.2	95	0.00
34	Bis(2-ethylhexyl)phthalate	0.797	0.709	11.0	93	0.00
35	Indeno(1,2,3-cd)pyrene	1.961	1.706	13.0	88	0.00
36 I	Perylene-d12	1.000	1.000	0.0	86	0.00
37	Benzo(b)fluoranthene	1.262	1.120	11.3	90	0.00
38	Benzo(k)fluoranthene	1.375	1.220	11.3	89	0.00
39 C	Benzo(a)pyrene	1.225	1.088	11.2	88	0.00
40	Dibenzo(a,h)anthracene	1.383	1.213	12.3	86	0.00
41	Benzo(g,h,i)perylene	1.453	1.270	12.6	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4