

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102922\
 Data File : BN022382.D
 Acq On : 28 Oct 2022 13:27
 Operator : CG/JU
 Sample : PB148509BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS509

Quant Time: Oct 29 01:09:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 29 00:50:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	167256	20.000	ng/u1	0.00
20) Naphthalene-d8	10.757	136	713655	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.604	164	390884	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.363	188	673580	20.000	ng/u1	0.00
79) Chrysene-d12	21.563	240	515921	20.000	ng/u1	0.00
88) Perylene-d12	24.021	264	510541	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	27680	6.272	ng/uL	0.00
4) Pyridine-d5	3.675	84	70702	5.151	ng/u1	0.00
7) Phenol-d5	7.087	99	509182	31.141	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.258	67	308576	30.116	ng/u1	0.00
11) 2-Chlorophenol-d4	7.452	132	378196	33.359	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	380030	29.250	ng/u1	0.00
21) Nitrobenzene-d5	9.110	128	190788	36.459	ng/u1	0.00
24) 2-Nitrophenol-d4	9.834	143	207613	38.708	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.375	165	353029	35.065	ng/u1	0.00
31) 4-Chloroaniline-d4	10.904	131	53433	3.246	ng/u1	0.00
46) Dimethylphthalate-d6	14.016	166	880196	32.303	ng/u1	0.00
49) Acenaphthylene-d8	14.298	160	1096074	36.022	ng/u1	0.00
54) 4-Nitrophenol-d4	14.816	143	165612	28.475	ng/u1	0.00
60) Fluorene-d10	15.598	176	742303	32.299	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.728	200	135023	35.137	ng/u1	0.00
73) Anthracene-d10	17.463	188	1017478	34.760	ng/u1	0.00
81) Pyrene-d10	19.763	212	978748	38.103	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.857	264	871995	34.827	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.275	88	55800	11.602	ng/uL	95
5) Pyridine	3.693	79	313352	23.011	ng/u1	91
6) Benzaldehyde	7.058	77	249228	30.866	ng/u1	97
8) Phenol	7.116	94	492200	28.361	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.352	93	393129	28.473	ng/u1	100
12) 2-Chlorophenol	7.487	128	381531	30.478	ng/u1	97
13) 2-Methylphenol	8.381	108	362432	27.931	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.469	45	527840	26.716	ng/u1	99
16) Acetophenone	8.769	105	568240	27.374	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.757	70	298763	27.022	ng/u1	98
18) 4-Methylphenol	8.716	108	397562	27.363	ng/u1	97
19) Hexachloroethane	9.010	117	155094	31.734	ng/u1	99
22) Nitrobenzene	9.152	77	439974	31.841	ng/u1	97
23) Isophorone	9.681	82	830538	30.444	ng/u1	98
25) 2-Nitrophenol	9.869	139	214452	33.879	ng/u1	96
26) 2,4-Dimethylphenol	9.928	107	412247	30.889	ng/u1	94
27) Bis(2-Chloroethoxy)met...	10.169	93	521510	30.571	ng/u1	98
29) 2,4-Dichlorophenol	10.404	162	334584	31.866	ng/u1	97
30) Naphthalene	10.810	128	1191491	31.166	ng/u1	99
32) 4-Chloroaniline	10.928	127	333369	19.537	ng/u1	98
33) Hexachlorobutadiene	11.087	225	191172	33.182	ng/u1	98
34) Caprolactam	11.716	113	61566	15.091	ng/u1	100
35) 4-Chloro-3-methylphenol	12.057	107	365465	28.871	ng/u1	97
36) 2-Methylnaphthalene	12.428	142	764256	28.948	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102922\
 Data File : BN022382.D
 Acq On : 28 Oct 2022 13:27
 Operator : CG/JU
 Sample : PB148509BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SLCS509


Quant Time: Oct 29 01:09:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 29 00:50:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.645	142	760784	28.776	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.793	216	353415	34.854	ng/ul	93
40) Hexachlorocyclopentadiene	12.769	237	200322	35.338	ng/ul	90
41) 2,4,6-Trichlorophenol	13.040	196	237355	34.329	ng/ul	96
42) 2,4,5-Trichlorophenol	13.110	196	254681	33.076	ng/ul	97
43) 1,1'-Biphenyl	13.440	154	960044	33.644	ng/ul	99
44) 2-Chloronaphthalene	13.481	162	744525	33.393	ng/ul	94
45) 2-Nitroaniline	13.692	65	225102	32.589	ng/ul	99
47) Dimethylphthalate	14.063	163	877052	29.945	ng/ul	100
48) 2,6-Dinitrotoluene	14.192	165	181495	33.082	ng/ul	94
50) Acenaphthylene	14.328	152	1149001	33.066	ng/ul	99
51) 3-Nitroaniline	14.522	138	176751	27.286	ng/ul	93
52) Acenaphthene	14.669	153	771976	31.328	ng/ul	100
53) 2,4-Dinitrophenol	14.734	184	100381	27.010	ng/ul	98
55) 4-Nitrophenol	14.828	109	121620	25.790	ng/ul	99
56) Dibenzofuran	15.004	168	1030424	30.644	ng/ul	98
57) 2,4-Dinitrotoluene	14.981	165	242994	30.234	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.234	232	185697	28.978	ng/ul	97
59) Diethylphthalate	15.428	149	868418	29.191	ng/ul	99
61) Fluorene	15.657	166	800586	29.617	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.651	204	370987	29.290	ng/ul	98
63) 4-Nitroaniline	15.686	138	184523	30.529	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.745	198	131192	31.800	ng/ul	100
67) N-Nitrosodiphenylamine	15.869	169	681345	34.978	ng/ul	98
68) 4-Bromophenyl-phenylether	16.545	248	217721	36.590	ng/ul	91
69) Hexachlorobenzene	16.663	284	246235	33.996	ng/ul	92
70) Atrazine	16.822	200	62292	9.496	ng/ul	98
71) Pentachlorophenol	17.004	266	131044	29.258	ng/ul	90
72) Phenanthrene	17.404	178	1159998	31.924	ng/ul	99
74) Anthracene	17.498	178	1152763	31.940	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.398	216	354389	41.830	ng/ul	98
76) Pentachlorobenzene	14.922	250	321580	35.364	ng/ul	97
77) Carbazole	17.769	167	1028265	30.803	ng/ul	99
78) Di-n-butylphthalate	18.333	149	1333095	32.222	ng/ul	99
80) Fluoranthene	19.427	202	1171705	36.339	ng/ul	98
82) Pyrene	19.792	202	1202868	35.658	ng/ul	100
83) Butylbenzylphthalate	20.692	149	547616	37.417	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.480	252	280504	25.140	ng/ul	96
85) Benzo(a)anthracene	21.545	228	1039291	31.526	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.469	149	809840	38.258	ng/ul	100
87) Chrysene	21.604	228	983724	31.051	ng/ul	97
89) Di-n-octyl phthalate	22.410	149	1309150	36.192	ng/ul	100
90) Benzo(b)fluoranthene	23.263	252	1045357	32.464	ng/ul	98
91) Benzo(k)fluoranthene	23.316	252	962372	30.251	ng/ul	97
93) Benzo(a)pyrene	23.910	252	944766	32.828	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.598	276	1203616	33.461	ng/ul	98
95) Dibenzo(a,h)anthracene	26.627	278	1036348	32.203	ng/ul	97
96) Benzo(g,h,i)perylene	27.392	276	1046292	32.331	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 29 01:09:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 29 00:50:42 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 11.602 ng/uL
RT: 3.275 min Scan# 49
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion: 88 Resp: 55800
Ion Ratio Lower Upper
88 100
43 26.6 22.6 34.0
58 70.3 59.9 89.9



#5
Pyridine
Concen: 23.011 ng/uL
RT: 3.693 min Scan# 120
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion: 79 Resp: 313352
Ion Ratio Lower Upper
79 100
52 67.7 61.7 92.5
51 32.5 27.8 41.6









#6
Benzaldehyde
Concen: 30.866 ng/ul
RT: 7.058 min Scan# 692
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion: 77 Resp: 249228
Ion Ratio Lower Upper
77 100
105 95.0 78.6 118.0
106 91.5 75.4 113.2



#8
Phenol
Concen: 28.361 ng/ul
RT: 7.116 min Scan# 702
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion: 94 Resp: 492200
Ion Ratio Lower Upper
94 100
65 28.0 22.1 33.1
66 35.0 29.0 43.6









#10
Bis(2-Chloroethyl)ether
Concen: 28.473 ng/ul
RT: 7.352 min Scan# 742
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion: 93 Resp: 393129
Ion Ratio Lower Upper
93 100
63 76.4 61.1 91.7
95 32.4 26.4 39.6



#12
2-Chlorophenol
Concen: 30.478 ng/ul
RT: 7.487 min Scan# 765
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:128 Resp: 381531
Ion Ratio Lower Upper
128 100
64 46.9 39.7 59.5
130 31.8 26.7 40.1









#13
2-Methylphenol
Concen: 27.931 ng/ul
RT: 8.381 min Scan# 917
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:108 Resp: 362432
Ion Ratio Lower Upper
108 100
107 86.8 72.4 108.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 26.716 ng/ul
RT: 8.469 min Scan# 932
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion: 45 Resp: 527840
Ion Ratio Lower Upper
45 100
77 14.8 11.7 17.5
79 11.5 8.7 13.1









#16
Acetophenone
Concen: 27.374 ng/ul
RT: 8.769 min Scan# 983
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:105 Resp: 568240
Ion Ratio Lower Upper
105 100
77 77.4 64.9 97.3
51 25.4 21.8 32.6



#17
N-Nitroso-di-n-propylamine
Concen: 27.022 ng/ul
RT: 8.757 min Scan# 981
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion: 70 Resp: 298763
Ion Ratio Lower Upper
70 100
42 55.8 45.6 68.4
101 10.0 7.4 11.2
130 21.3 15.4 23.0





#18

4-Methylphenol

Concen: 27.363 ng/ul

RT: 8.716 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :





SLCS509

Tgt Ion:108 Resp: 397562

Ion Ratio Lower Upper

108 100

107 105.5 87.2 130.8



#19

Hexachloroethane

Concen: 31.734 ng/ul

RT: 9.010 min Scan# 1024

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27




Tgt Ion:117 Resp: 155094


Ion Ratio Lower Upper

117 100

201 86.4 69.8 104.6

199 56.4 45.2 67.8









#22
Nitrobenzene
Concen: 31.841 ng/ul
RT: 9.152 min Scan# 1048
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion: 77 Resp: 439974
Ion Ratio Lower Upper
77 100
123 45.5 34.9 52.3
65 13.1 10.2 15.2



#23
Isophorone
Concen: 30.444 ng/ul
RT: 9.681 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion: 82 Resp: 830538
Ion Ratio Lower Upper
82 100
95 6.5 5.4 8.0
138 16.0 13.8 20.8









#25
2-Nitrophenol
Concen: 33.879 ng/ul
RT: 9.869 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:139 Resp: 214452
Ion Ratio Lower Upper
139 100
65 50.6 42.7 64.1
109 27.8 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 30.889 ng/ul
RT: 9.928 min Scan# 1180
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:107 Resp: 412247
Ion Ratio Lower Upper
107 100
121 48.5 41.0 61.6
122 83.2 71.5 107.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 30.571 ng/ul
 RT: 10.169 min Scan# 1221
 Delta R.T. 0.000 min
 Lab File: BN022382.D
 Acq: 28 Oct 2022 13:27




Instrument :
 BNA_N
 ClientSampleId :
 SLCS509


Tgt Ion: 93 Resp: 521510
 Ion Ratio Lower Upper
 93 100
 95 32.7 25.2 37.8
 123 10.9 8.6 12.8



#29
 2,4-Dichlorophenol
 Concen: 31.866 ng/ul
 RT: 10.404 min Scan# 1261
 Delta R.T. 0.006 min
 Lab File: BN022382.D
 Acq: 28 Oct 2022 13:27

Tgt Ion:162 Resp: 334584
 Ion Ratio Lower Upper
 162 100
 164 62.8 52.9 79.3
 98 37.0 30.5 45.7





#30

Naphthalene

Concen: 31.166 ng/ul

RT: 10.810 min Scan# 11

Delta R.T. 0.006 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :

SLCS509





Tgt Ion:128 Resp: 1191491

Ion Ratio Lower Upper

128 100

129 10.9 8.5 12.7

127 12.9 10.8 16.2



#32

4-Chloroaniline

Concen: 19.537 ng/ul

RT: 10.928 min Scan# 1350

Delta R.T. 0.000 min

Lab File: BN022382.D




Acq: 28 Oct 2022 13:27


Tgt Ion:127 Resp: 333369

Ion Ratio Lower Upper

127 100

129 31.9 26.2 39.4









#33
Hexachlorobutadiene
Concen: 33.182 ng/ul
RT: 11.087 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:225 Resp: 191172
Ion Ratio Lower Upper
225 100
223 59.9 47.7 71.5
227 61.1 51.4 77.0



#34
Caprolactam
Concen: 15.091 ng/ul
RT: 11.716 min Scan# 1484
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:113 Resp: 61566
Ion Ratio Lower Upper
113 100
55 173.9 138.7 208.1
56 127.0 101.4 152.0









#35
 4-Chloro-3-methylphenol
 Concen: 28.871 ng/ul
 RT: 12.057 min Scan# 1542
 Delta R.T. 0.006 min
 Lab File: BN022382.D
 Acq: 28 Oct 2022 13:27




Instrument :
 BNA_N
 ClientSampleId :
 SLCS509


Tgt Ion:107 Resp: 365465
 Ion Ratio Lower Upper
 107 100
 144 26.6 20.3 30.5
 142 78.1 64.4 96.6



#36
 2-Methylnaphthalene
 Concen: 28.948 ng/ul
 RT: 12.428 min Scan# 1605
 Delta R.T. 0.006 min
 Lab File: BN022382.D
 Acq: 28 Oct 2022 13:27

Tgt Ion:142 Resp: 764256
 Ion Ratio Lower Upper
 142 100
 141 85.9 69.7 104.5









#37
 1-Methylnaphthalene
 Concen: 28.776 ng/ul
 RT: 12.645 min Scan# 1642
 Delta R.T. 0.000 min
 Lab File: BN022382.D
 Acq: 28 Oct 2022 13:27




Instrument :
 BNA_N
 ClientSampleId :
 SLCS509


Tgt Ion:142 Resp: 760784
 Ion Ratio Lower Upper
 142 100
 141 93.9 76.7 115.1
 116 3.8 3.1 4.7



#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 34.854 ng/ul
 RT: 12.793 min Scan# 1667
 Delta R.T. 0.000 min
 Lab File: BN022382.D
 Acq: 28 Oct 2022 13:27

Tgt Ion:216 Resp: 353415
 Ion Ratio Lower Upper
 216 100
 214 72.5 65.4 98.2
 179 21.6 17.4 26.0
 108 19.6 16.4 24.6









#40
Hexachlorocyclopentadiene
Concen: 35.338 ng/ul
RT: 12.769 min Scan# 1662
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:237 Resp: 200322
Ion Ratio Lower Upper
237 100
235 62.5 43.4 65.0
272 13.2 10.2 15.2



#41
2,4,6-Trichlorophenol
Concen: 34.329 ng/ul
RT: 13.040 min Scan# 1709
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:196 Resp: 237355
Ion Ratio Lower Upper
196 100
198 96.4 73.4 110.2
200 28.9 24.3 36.5









#42
2,4,5-Trichlorophenol
Concen: 33.076 ng/ul
RT: 13.110 min Scan# 1721
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:196 Resp: 254681
Ion Ratio Lower Upper
196 100
198 100.2 77.0 115.6
200 30.6 24.7 37.1



#43
1,1'-Biphenyl
Concen: 33.644 ng/ul
RT: 13.440 min Scan# 1777
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:154 Resp: 960044
Ion Ratio Lower Upper
154 100
153 39.8 30.9 46.3
76 14.5 11.6 17.4





#44

2-Chloronaphthalene

Concen: 33.393 ng/ul

RT: 13.481 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :

SLCS509





Tgt Ion:162 Resp: 744525

Ion Ratio Lower Upper

162 100

164 33.1 23.7 35.5

127 41.9 30.8 46.2



#45

2-Nitroaniline

Concen: 32.589 ng/ul

RT: 13.692 min Scan# 1820

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27




Tgt Ion: 65 Resp: 225102


Ion Ratio Lower Upper

65 100

92 68.0 56.2 84.2

138 105.7 85.0 127.6









#47
Dimethylphthalate
Concen: 29.945 ng/ul
RT: 14.063 min Scan# 1883
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:163 Resp: 877052
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 10.3 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 33.082 ng/ul
RT: 14.192 min Scan# 1905
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:165 Resp: 181495
Ion Ratio Lower Upper
165 100
89 59.8 52.1 78.1
121 21.8 19.0 28.4









#50
Acenaphthylene
Concen: 33.066 ng/ul
RT: 14.328 min Scan# 1928
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:152 Resp: 1149001
Ion Ratio Lower Upper
152 100
151 19.9 15.5 23.3
153 12.8 9.8 14.8



#51
3-Nitroaniline
Concen: 27.286 ng/ul
RT: 14.522 min Scan# 1961
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:138 Resp: 176751
Ion Ratio Lower Upper
138 100
108 10.4 9.0 13.6
92 112.7 96.4 144.6





#52

Acenaphthene

Concen: 31.328 ng/ul

RT: 14.669 min Scan# 1986

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :

SLCS509





Tgt Ion:153 Resp: 771976

Ion Ratio Lower Upper

153 100

152 47.6 38.0 57.0

154 89.4 71.1 106.7



#53

2,4-Dinitrophenol

Concen: 27.010 ng/ul

RT: 14.734 min Scan# 1997

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27




Tgt Ion:184 Resp: 100381


Ion Ratio Lower Upper

184 100

63 69.4 56.8 85.2

154 62.2 49.0 73.6









#55
4-Nitrophenol
Concen: 25.790 ng/ul
RT: 14.828 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:109 Resp: 121620
Ion Ratio Lower Upper
109 100
139 136.0 108.3 162.5
65 137.1 107.9 161.9



#56
Dibenzofuran
Concen: 30.644 ng/ul
RT: 15.004 min Scan# 2043
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:168 Resp: 1030424
Ion Ratio Lower Upper
168 100
139 37.2 30.6 45.8









#57
2,4-Dinitrotoluene
Concen: 30.234 ng/ul
RT: 14.981 min Scan# 2039
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:165 Resp: 242994
Ion Ratio Lower Upper
165 100
63 46.0 36.2 54.4
182 2.8 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 28.978 ng/ul
RT: 15.234 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:232 Resp: 185697
Ion Ratio Lower Upper
232 100
131 53.2 44.4 66.6
130 2.6 2.2 3.4
166 33.8 28.8 43.2





#59

Diethylphthalate

Concen: 29.191 ng/ul

RT: 15.428 min Scan# 2115

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :

SLCS509





Tgt Ion:149 Resp: 868418

Ion Ratio Lower Upper

149 100

177 21.1 17.5 26.3

150 11.9 9.8 14.8



#61

Fluorene

Concen: 29.617 ng/ul

RT: 15.657 min Scan# 2154

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27




Tgt Ion:166 Resp: 800586


Ion Ratio Lower Upper

166 100

165 96.6 72.8 109.2

167 13.7 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 29.290 ng/ul
RT: 15.651 min Scan# 2153
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:204 Resp: 370987
Ion Ratio Lower Upper
204 100
206 34.7 26.2 39.2
141 66.0 52.0 78.0



#63
4-Nitroaniline
Concen: 30.529 ng/ul
RT: 15.686 min Scan# 2159
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:138 Resp: 184523
Ion Ratio Lower Upper
138 100
92 54.2 41.8 62.8
108 78.1 60.4 90.6









#66
4,6-Dinitro-2-methylphenol
Concen: 31.800 ng/ul
RT: 15.745 min Scan# 2169
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:198 Resp: 131192
Ion Ratio Lower Upper
198 100
182 4.6 3.8 5.8
77 30.4 24.4 36.6



#67
N-Nitrosodiphenylamine
Concen: 34.978 ng/ul
RT: 15.869 min Scan# 2190
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:169 Resp: 681345
Ion Ratio Lower Upper
169 100
168 65.5 50.9 76.3
167 36.0 28.2 42.4









#68
4-Bromophenyl-phenylether
Concen: 36.590 ng/ul
RT: 16.545 min Scan# 2177
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:248 Resp: 217721
Ion Ratio Lower Upper
248 100
250 98.4 86.3 129.5
141 84.4 74.2 111.4



#69
Hexachlorobenzene
Concen: 33.996 ng/ul
RT: 16.663 min Scan# 2325
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:284 Resp: 246235
Ion Ratio Lower Upper
284 100
142 38.0 35.7 53.5
249 30.2 27.0 40.6





#70

Atrazine

Concen: 9.496 ng/ul

RT: 16.822 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :

SLCS509





Tgt Ion:200 Resp: 62292

Ion Ratio Lower Upper

200 100

173 27.2 22.8 34.2

215 49.0 38.1 57.1



#71

Pentachlorophenol

Concen: 29.258 ng/ul

RT: 17.004 min Scan# 2383

Delta R.T. -0.006 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27




Tgt Ion:266 Resp: 131044


Ion Ratio Lower Upper

266 100

264 60.5 52.7 79.1

268 57.2 54.6 82.0





#72

Phenanthrene

Concen: 31.924 ng/ul

RT: 17.404 min Scan# 2451

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :

SLCS509





Tgt Ion:178 Resp: 1159998

Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 20.1 16.5 24.7



#74

Anthracene

Concen: 31.940 ng/ul

RT: 17.498 min Scan# 2467

Delta R.T. 0.006 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27




Tgt Ion:178 Resp: 1152763


Ion Ratio Lower Upper

178 100

179 15.5 13.0 19.6

176 18.4 15.8 23.6









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 41.830 ng/uL
 RT: 13.398 min Scan# 1770
 Delta R.T. 0.000 min
 Lab File: BN022382.D
 Acq: 28 Oct 2022 13:27




Instrument :
 BNA_N
 ClientSampleId :
 SLCS509


Tgt Ion:	216	Resp:	354389
Ion Ratio	Lower	Upper	
216	100		
214	78.5	61.2	91.8
218	47.8	39.4	59.0



#76
 Pentachlorobenzene
 Concen: 35.364 ng/uL
 RT: 14.922 min Scan# 2029
 Delta R.T. 0.000 min
 Lab File: BN022382.D
 Acq: 28 Oct 2022 13:27

Tgt Ion:	250	Resp:	321580
Ion Ratio	Lower	Upper	
250	100		
248	57.9	46.8	70.2
252	62.8	53.9	80.9





#77

Carbazole

Concen: 30.803 ng/ul

RT: 17.769 min Scan# 2513

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :

SLCS509





Tgt Ion:167 Resp: 1028265

Ion Ratio Lower Upper

167 100

166 21.0 17.3 25.9

139 13.4 10.7 16.1



#78

Di-n-butylphthalate

Concen: 32.222 ng/ul

RT: 18.333 min Scan# 2609

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27




Tgt Ion:149 Resp: 1333095


Ion Ratio Lower Upper

149 100

150 8.8 7.1 10.7

104 5.0 4.2 6.2









#80
Fluoranthene
Concen: 36.339 ng/ul
RT: 19.427 min Scan# 2795
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:202 Resp: 1171705
Ion Ratio Lower Upper
202 100
101 13.8 10.3 15.5
100 10.5 7.8 11.8



#82
Pyrene
Concen: 35.658 ng/ul
RT: 19.792 min Scan# 2857
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:202 Resp: 1202868
Ion Ratio Lower Upper
202 100
101 16.7 13.4 20.0
100 13.7 10.9 16.3





#83

Butylbenzylphthalate

Concen: 37.417 ng/ul

RT: 20.692 min Scan# 3010

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :

SLCS509





Tgt Ion:149 Resp: 547616

Ion Ratio Lower Upper

149 100

91 75.7 61.7 92.5

206 19.3 14.7 22.1



#84

3,3'-Dichlorobenzidine

Concen: 25.140 ng/ul

RT: 21.480 min Scan# 3144

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27




Tgt Ion:252 Resp: 280504


Ion Ratio Lower Upper

252 100

254 62.7 52.9 79.3

126 15.2 11.2 16.8









#85
Benzo(a)anthracene
Concen: 31.526 ng/ul
RT: 21.545 min Scan# 3155
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:228 Resp: 1039291
Ion Ratio Lower Upper
228 100
229 19.8 13.5 20.3
226 26.8 21.0 31.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 38.258 ng/ul
RT: 21.469 min Scan# 3142
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:149 Resp: 809840
Ion Ratio Lower Upper
149 100
167 26.5 21.3 31.9
279 3.9 2.9 4.3









#87
Chrysene
Concen: 31.051 ng/ul
RT: 21.604 min Scan# 3164
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27




Instrument :
BNA_N
ClientSampleId :
SLCS509


Tgt Ion:228 Resp: 983724
Ion Ratio Lower Upper
228 100
226 28.6 23.8 35.8
229 18.6 16.5 24.7



#89
Di-n-octyl phthalate
Concen: 36.192 ng/ul
RT: 22.410 min Scan# 3302
Delta R.T. 0.000 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:149 Resp: 1309150





#90

Benzo(b)fluoranthene

Concen: 32.464 ng/ul

RT: 23.263 min Scan# 3447

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :

SLCS509





Tgt Ion:252 Resp: 1045357

Ion Ratio Lower Upper

252 100

253 23.2 17.6 26.4

125 13.4 10.7 16.1



#91

Benzo(k)fluoranthene

Concen: 30.251 ng/ul

RT: 23.316 min Scan# 3456

Delta R.T. 0.006 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27




Tgt Ion:252 Resp: 962372


Ion Ratio Lower Upper

252 100

253 23.7 17.7 26.5

125 13.7 10.6 15.8





#93

Benzo(a)pyrene

Concen: 32.828 ng/ul

RT: 23.910 min Scan# 3

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27

Instrument :

BNA_N

ClientSampleId :

SLCS509





Tgt Ion:252 Resp: 944766

Ion Ratio Lower Upper

252 100

253 20.4 17.2 25.8

125 13.7 10.8 16.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 33.461 ng/ul

RT: 26.598 min Scan# 4014

Delta R.T. 0.000 min

Lab File: BN022382.D

Acq: 28 Oct 2022 13:27




Tgt Ion:276 Resp: 1203616


Ion Ratio Lower Upper

276 100

138 30.1 24.6 36.8

277 24.8 18.4 27.6









#95
Dibenzo(a,h)anthracene
Concen: 32.203 ng/ul
RT: 26.627 min Scan# 4016
Delta R.T. 0.018 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Instrument :
BNA_N
ClientSampleId :
SLCS509

Tgt Ion:278 Resp: 1036348
Ion Ratio Lower Upper
278 100
139 17.0 15.7 23.5
279 20.2 16.5 24.7



#96
Benzo(g,h,i)perylene
Concen: 32.331 ng/ul
RT: 27.392 min Scan# 4149
Delta R.T. 0.006 min
Lab File: BN022382.D
Acq: 28 Oct 2022 13:27

Tgt Ion:276 Resp: 1046292
Ion Ratio Lower Upper
276 100
138 24.9 21.2 31.8
277 24.2 17.9 26.9

