


Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN102922\  
 Data File : BN022362.D  
 Acq On : 27 Oct 2022 15:31  
 Operator : CG/JU  
 Sample : DFTPP257  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP257

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\SFAM-EPA-BN101222.M  
 Title : SVOA CALIBRATION  
 Last Update : Fri Oct 28 00:29:18 2022



AutoFind: Scans 2518, 2519, 2520; Background Corrected with Scan 2512

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 38.7         | 131920     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.5          | 2141       | PASS                |
| 69             | 198             | 0.00            | 100             | 41.8         | 142460     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 652        | PASS                |
| 127            | 198             | 10              | 80              | 50.7         | 172701     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.5          | 1716       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 340651     | PASS                |
| 199            | 198             | 5               | 9               | 7.0          | 23741      | PASS                |
| 275            | 198             | 10              | 60              | 26.6         | 90480      | PASS                |
| 365            | 198             | 1               | 100             | 3.4          | 11634      | PASS                |
| 441            | 198             | 0.01            | 100             | 12.8         | 43613      | PASS                |
| 442            | 198             | 50              | 100             | 82.0         | 279275     | PASS                |
| 443            | 442             | 15              | 24              | 19.7         | 55120      | PASS                |

| m/z    | Abundance |
|--------|-----------|
| 39.00  | 245.0     |
| 40.00  | 1276.0    |
| 41.00  | 774.0     |
| 42.10  | 166.0     |
| 43.00  | 679.0     |
| 44.00  | 773.0     |
| 48.90  | 256.0     |
| 54.00  | 166.0     |
| 55.00  | 521.0     |
| 56.00  | 324.0     |
| 57.00  | 725.0     |
| 67.00  | 223.0     |
| 68.10  | 150.0     |
| 69.00  | 361.0     |
| 70.10  | 264.0     |
| 71.10  | 329.0     |
| 73.00  | 562.0     |
| 76.90  | 181.0     |
| 81.00  | 205.0     |
| 82.00  | 209.0     |
| 83.00  | 393.0     |
| 83.90  | 246.0     |
| 85.10  | 369.0     |
| 90.90  | 169.0     |
| 95.90  | 182.0     |
| 97.10  | 345.0     |
| 133.00 | 193.0     |
| 149.00 | 159.0     |
| 207.00 | 1247.0    |
| 208.00 | 290.0     |
| 281.00 | 347.0     |

Instrument :

BNA\_N

ClientSampleId :

DFTPP257

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 669.0     |
| 38.00  | 1831.0    |
| 39.00  | 11380.0   |
| 40.00  | 1361.0    |
| 41.00  | 711.0     |
| 43.00  | 442.0     |
| 44.00  | 722.0     |
| 45.00  | 289.0     |
| 49.00  | 1222.0    |
| 50.00  | 41744.0   |
| 51.00  | 159872.0  |
| 52.00  | 8041.0    |
| 53.00  | 398.0     |
| 55.00  | 1169.0    |
| 56.00  | 4925.0    |
| 57.00  | 11541.0   |
| 58.00  | 623.0     |
| 59.00  | 207.0     |
| 59.90  | 169.0     |
| 61.00  | 2108.0    |
| 62.00  | 2279.0    |
| 63.00  | 6486.0    |
| 64.00  | 951.0     |
| 65.00  | 3499.0    |
| 66.00  | 295.0     |
| 67.00  | 346.0     |
| 68.00  | 2692.0    |
| 69.00  | 169984.0  |
| 70.00  | 1027.0    |
| 71.00  | 338.0     |
| 72.00  | 182.0     |
| 73.00  | 1918.0    |
| 74.00  | 17280.0   |
| 75.00  | 27744.0   |
| 76.00  | 10236.0   |
| 77.00  | 206976.0  |
| 78.00  | 14662.0   |
| 79.00  | 12303.0   |
| 80.00  | 9405.0    |
| 81.00  | 13522.0   |
| 82.00  | 3702.0    |
| 83.00  | 3516.0    |
| 83.90  | 595.0     |
| 85.00  | 2802.0    |
| 86.00  | 4229.0    |
| 87.00  | 1719.0    |
| 88.00  | 830.0     |
| 88.90  | 344.0     |
| 91.00  | 2907.0    |
| 92.00  | 3087.0    |
| 93.00  | 21944.0   |
| 94.00  | 1566.0    |
| 95.00  | 468.0     |
| 96.00  | 1180.0    |
| 97.00  | 600.0     |
| 98.00  | 17832.0   |
| 99.00  | 14324.0   |
| 100.00 | 1222.0    |
| 101.00 | 7520.0    |
| 102.00 | 499.0     |
| 103.00 | 2445.0    |
| 104.00 | 4410.0    |
| 105.00 | 4853.0    |
| 106.00 | 1672.0    |
| 107.00 | 59272.0   |
| 108.00 | 8679.0    |
| 109.00 | 1563.0    |
| 110.00 | 106968.0  |
| 111.00 | 15913.0   |
| 112.00 | 2132.0    |
| 112.90 | 722.0     |
| 114.80 | 291.0     |
| 116.00 | 3029.0    |
| 117.00 | 52520.0   |
| 118.00 | 4205.0    |
| 118.90 | 454.0     |
| 120.00 | 676.0     |
| 120.90 | 303.0     |
| 122.00 | 3821.0    |
| 123.00 | 6026.0    |
| 123.90 | 2782.0    |
| 125.00 | 2422.0    |
| 126.00 | 674.0     |
| 127.00 | 195968.0  |
| 128.00 | 15735.0   |
| 129.00 | 76344.0   |
| 130.00 | 6676.0    |
| 131.00 | 1219.0    |
| 132.00 | 581.0     |
| 133.00 | 436.0     |
| 133.90 | 2054.0    |
| 135.00 | 6393.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP257

|        |          |
|--------|----------|
| 135.90 | 2503.0   |
| 137.00 | 3715.0   |
| 137.80 | 694.0    |
| 139.00 | 399.0    |
| 140.00 | 1003.0   |
| 141.00 | 9850.0   |
| 142.00 | 3137.0   |
| 142.90 | 2437.0   |
| 144.00 | 655.0    |
| 144.90 | 531.0    |
| 146.00 | 1732.0   |
| 147.00 | 5290.0   |
| 148.00 | 12756.0  |
| 149.00 | 2742.0   |
| 150.00 | 765.0    |
| 151.00 | 1262.0   |
| 151.80 | 716.0    |
| 153.00 | 3064.0   |
| 154.00 | 2247.0   |
| 155.00 | 5313.0   |
| 156.00 | 8265.0   |
| 157.00 | 1642.0   |
| 157.90 | 1790.0   |
| 159.00 | 1502.0   |
| 160.00 | 2901.0   |
| 161.00 | 5161.0   |
| 161.90 | 1311.0   |
| 163.00 | 362.0    |
| 164.00 | 401.0    |
| 165.00 | 3705.0   |
| 166.00 | 2739.0   |
| 167.00 | 21736.0  |
| 168.00 | 12118.0  |
| 169.00 | 2003.0   |
| 170.00 | 598.0    |
| 170.90 | 893.0    |
| 171.90 | 1688.0   |
| 173.00 | 2136.0   |
| 174.00 | 4384.0   |
| 175.00 | 7631.0   |
| 176.00 | 2338.0   |
| 177.00 | 3213.0   |
| 178.00 | 1075.0   |
| 179.00 | 14700.0  |
| 180.00 | 10644.0  |
| 181.00 | 4864.0   |
| 182.00 | 730.0    |
| 182.80 | 294.0    |
| 184.00 | 1060.0   |
| 185.00 | 7180.0   |
| 186.00 | 52240.0  |
| 187.00 | 15160.0  |
| 188.00 | 1869.0   |
| 189.00 | 3433.0   |
| 189.90 | 494.0    |
| 191.10 | 1574.0   |
| 192.00 | 4353.0   |
| 193.00 | 5383.0   |
| 194.00 | 1141.0   |
| 195.00 | 395.0    |
| 196.00 | 10291.0  |
| 196.90 | 2074.0   |
| 198.00 | 371712.0 |
| 199.00 | 26224.0  |
| 200.00 | 2190.0   |
| 201.50 | 1437.0   |
| 203.00 | 2801.0   |
| 204.00 | 13267.0  |
| 205.00 | 24048.0  |
| 206.10 | 95400.0  |
| 207.00 | 13835.0  |
| 208.00 | 3279.0   |
| 209.00 | 1002.0   |
| 210.10 | 1462.0   |
| 211.10 | 4007.0   |
| 211.90 | 761.0    |
| 213.10 | 261.0    |
| 215.00 | 1003.0   |
| 216.00 | 1923.0   |
| 217.00 | 24296.0  |
| 218.00 | 3490.0   |
| 219.10 | 329.0    |
| 221.00 | 14004.0  |
| 223.00 | 5478.0   |
| 224.10 | 54048.0  |
| 225.00 | 14186.0  |
| 226.00 | 1657.0   |
| 227.00 | 22056.0  |
| 228.00 | 2786.0   |
| 229.00 | 4257.0   |
| 230.00 | 592.0    |
| 231.00 | 1593.0   |
| 231.90 | 296.0    |
| 233.00 | 398.0    |
| 234.00 | 1309.0   |
| 235.00 | 1609.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP257

|        |          |
|--------|----------|
| 236.00 | 1009.0   |
| 237.00 | 2349.0   |
| 238.00 | 313.0    |
| 239.00 | 887.0    |
| 239.90 | 711.0    |
| 240.90 | 1292.0   |
| 242.00 | 2777.0   |
| 243.00 | 3091.0   |
| 244.10 | 40552.0  |
| 245.10 | 5123.0   |
| 246.00 | 8153.0   |
| 247.00 | 1659.0   |
| 248.00 | 377.0    |
| 249.00 | 1353.0   |
| 249.90 | 221.0    |
| 251.00 | 457.0    |
| 252.00 | 425.0    |
| 253.00 | 812.0    |
| 254.10 | 1597.0   |
| 255.00 | 196800.0 |
| 256.00 | 29256.0  |
| 257.00 | 2587.0   |
| 258.00 | 11671.0  |
| 259.00 | 1853.0   |
| 260.00 | 395.0    |
| 260.90 | 413.0    |
| 264.00 | 299.0    |
| 265.00 | 4800.0   |
| 265.90 | 637.0    |
| 266.90 | 154.0    |
| 270.00 | 255.0    |
| 271.00 | 588.0    |
| 272.20 | 522.0    |
| 273.00 | 6059.0   |
| 274.00 | 16370.0  |
| 275.00 | 95488.0  |
| 276.10 | 12461.0  |
| 277.00 | 7972.0   |
| 277.90 | 1401.0   |
| 279.00 | 287.0    |
| 281.00 | 323.0    |
| 282.00 | 271.0    |
| 283.00 | 1010.0   |
| 284.00 | 579.0    |
| 285.00 | 1461.0   |
| 285.90 | 344.0    |
| 289.10 | 273.0    |
| 289.90 | 255.0    |
| 290.90 | 206.0    |
| 292.00 | 373.0    |
| 293.00 | 1825.0   |
| 294.00 | 547.0    |
| 296.00 | 25648.0  |
| 297.00 | 3624.0   |
| 297.90 | 216.0    |
| 301.00 | 332.0    |
| 302.00 | 460.0    |
| 303.00 | 3176.0   |
| 304.00 | 724.0    |
| 307.90 | 388.0    |
| 308.90 | 207.0    |
| 310.00 | 443.0    |
| 312.90 | 212.0    |
| 314.10 | 1135.0   |
| 315.00 | 2676.0   |
| 316.00 | 1744.0   |
| 316.90 | 274.0    |
| 320.90 | 850.0    |
| 322.10 | 366.0    |
| 323.10 | 8383.0   |
| 324.00 | 1700.0   |
| 325.00 | 160.0    |
| 327.00 | 1481.0   |
| 328.00 | 925.0    |
| 332.00 | 524.0    |
| 333.00 | 828.0    |
| 334.00 | 5030.0   |
| 335.10 | 1291.0   |
| 336.00 | 161.0    |
| 341.00 | 1033.0   |
| 342.00 | 330.0    |
| 345.90 | 2150.0   |
| 347.00 | 448.0    |
| 351.00 | 184.0    |
| 352.00 | 2604.0   |
| 353.00 | 1694.0   |
| 354.00 | 2850.0   |
| 355.00 | 691.0    |
| 359.00 | 268.0    |
| 365.00 | 10501.0  |
| 365.90 | 1770.0   |
| 369.90 | 246.0    |
| 371.10 | 555.0    |
| 372.00 | 4823.0   |
| 373.00 | 1198.0   |
| 383.00 | 1508.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP257

|        |          |
|--------|----------|
| 384.00 | 348.0    |
| 390.00 | 570.0    |
| 391.00 | 386.0    |
| 392.00 | 238.0    |
| 400.90 | 306.0    |
| 401.90 | 1475.0   |
| 403.00 | 2476.0   |
| 404.00 | 964.0    |
| 405.00 | 303.0    |
| 420.90 | 2026.0   |
| 422.00 | 1867.0   |
| 423.00 | 12768.0  |
| 424.10 | 4303.0   |
| 425.00 | 616.0    |
| 441.00 | 40120.0  |
| 442.10 | 237760.0 |
| 443.10 | 49744.0  |
| 444.10 | 3662.0   |
| 445.10 | 231.0    |

|                  |
|------------------|
| Instrument :     |
| BNA_N            |
| ClientSampleId : |
| DFTPP257         |

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 674.0     |
| 38.00  | 1764.0    |
| 39.10  | 10377.0   |
| 40.00  | 1223.0    |
| 41.00  | 920.0     |
| 43.00  | 513.0     |
| 44.00  | 813.0     |
| 45.00  | 281.0     |
| 48.00  | 156.0     |
| 49.00  | 1295.0    |
| 50.00  | 39512.0   |
| 51.00  | 150592.0  |
| 52.00  | 7770.0    |
| 53.00  | 404.0     |
| 55.00  | 1085.0    |
| 56.00  | 4712.0    |
| 57.00  | 10445.0   |
| 58.00  | 545.0     |
| 59.90  | 182.0     |
| 61.00  | 1982.0    |
| 62.00  | 2227.0    |
| 63.00  | 6395.0    |
| 64.00  | 815.0     |
| 65.00  | 3167.0    |
| 66.00  | 278.0     |
| 67.00  | 334.0     |
| 68.00  | 2662.0    |
| 69.00  | 165504.0  |
| 70.00  | 1099.0    |
| 71.00  | 327.0     |
| 72.00  | 155.0     |
| 73.00  | 1810.0    |
| 74.00  | 17544.0   |
| 75.00  | 26528.0   |
| 76.10  | 9531.0    |
| 77.00  | 196992.0  |
| 78.10  | 13716.0   |
| 79.00  | 12026.0   |
| 80.00  | 8747.0    |
| 81.00  | 13231.0   |
| 82.00  | 3709.0    |
| 83.00  | 3416.0    |
| 84.00  | 711.0     |
| 85.00  | 2462.0    |
| 86.00  | 3771.0    |
| 87.00  | 1750.0    |
| 87.90  | 833.0     |
| 89.00  | 325.0     |
| 90.00  | 161.0     |
| 91.00  | 3038.0    |
| 92.00  | 3453.0    |
| 93.00  | 22696.0   |
| 94.00  | 1315.0    |
| 95.00  | 450.0     |
| 96.00  | 1160.0    |
| 97.00  | 472.0     |
| 98.00  | 17568.0   |
| 99.00  | 13702.0   |
| 99.90  | 1236.0    |
| 101.00 | 6844.0    |
| 101.90 | 381.0     |
| 102.90 | 2342.0    |
| 104.00 | 4575.0    |
| 105.00 | 4559.0    |
| 106.00 | 1395.0    |
| 107.00 | 59488.0   |
| 108.00 | 8867.0    |
| 109.10 | 1704.0    |
| 110.00 | 104800.0  |
| 111.00 | 15852.0   |
| 112.00 | 2058.0    |
| 112.90 | 678.0     |
| 114.00 | 189.0     |
| 115.00 | 338.0     |
| 116.00 | 3164.0    |
| 117.00 | 52088.0   |
| 118.00 | 3979.0    |
| 118.90 | 430.0     |
| 120.00 | 693.0     |
| 121.00 | 269.0     |
| 122.00 | 3800.0    |
| 123.00 | 5680.0    |
| 124.00 | 2730.0    |
| 125.00 | 2603.0    |
| 126.00 | 695.0     |
| 127.00 | 200000.0  |
| 128.00 | 15429.0   |
| 129.00 | 77720.0   |
| 130.00 | 6724.0    |
| 131.00 | 1248.0    |
| 132.00 | 689.0     |
| 132.80 | 340.0     |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP257

|        |          |
|--------|----------|
| 134.00 | 2251.0   |
| 135.00 | 6098.0   |
| 136.00 | 2700.0   |
| 137.00 | 3973.0   |
| 137.90 | 928.0    |
| 139.00 | 406.0    |
| 139.90 | 893.0    |
| 141.00 | 10612.0  |
| 142.00 | 3288.0   |
| 142.90 | 2410.0   |
| 144.10 | 776.0    |
| 145.00 | 485.0    |
| 146.00 | 1898.0   |
| 147.00 | 5127.0   |
| 148.00 | 12440.0  |
| 149.00 | 2768.0   |
| 150.00 | 834.0    |
| 151.10 | 1152.0   |
| 153.00 | 3095.0   |
| 154.00 | 2473.0   |
| 155.00 | 5878.0   |
| 156.00 | 8036.0   |
| 157.10 | 1651.0   |
| 157.90 | 1605.0   |
| 159.00 | 1561.0   |
| 159.90 | 3143.0   |
| 161.00 | 5001.0   |
| 162.00 | 1423.0   |
| 163.10 | 473.0    |
| 164.00 | 520.0    |
| 165.00 | 3645.0   |
| 166.00 | 3005.0   |
| 167.00 | 21736.0  |
| 168.00 | 12949.0  |
| 169.00 | 1953.0   |
| 170.00 | 693.0    |
| 170.90 | 795.0    |
| 172.00 | 1754.0   |
| 173.00 | 2619.0   |
| 174.00 | 4556.0   |
| 175.00 | 8019.0   |
| 176.00 | 2241.0   |
| 177.00 | 3756.0   |
| 178.00 | 1193.0   |
| 179.00 | 15782.0  |
| 180.00 | 10958.0  |
| 181.00 | 4928.0   |
| 182.00 | 886.0    |
| 183.00 | 389.0    |
| 184.00 | 1400.0   |
| 185.00 | 7245.0   |
| 186.00 | 58288.0  |
| 187.00 | 15407.0  |
| 188.00 | 1702.0   |
| 189.00 | 3616.0   |
| 190.00 | 641.0    |
| 191.00 | 1834.0   |
| 192.00 | 4989.0   |
| 193.00 | 5693.0   |
| 194.00 | 929.0    |
| 195.00 | 633.0    |
| 196.00 | 9805.0   |
| 196.90 | 1935.0   |
| 198.00 | 389504.0 |
| 199.00 | 28376.0  |
| 200.00 | 2498.0   |
| 201.40 | 1540.0   |
| 203.00 | 2948.0   |
| 204.00 | 14893.0  |
| 205.00 | 24104.0  |
| 206.00 | 106280.0 |
| 207.10 | 13564.0  |
| 208.00 | 3775.0   |
| 209.00 | 1263.0   |
| 210.10 | 1738.0   |
| 210.90 | 4201.0   |
| 213.10 | 222.0    |
| 213.90 | 174.0    |
| 214.90 | 1155.0   |
| 216.00 | 2419.0   |
| 217.00 | 28096.0  |
| 218.00 | 3734.0   |
| 219.00 | 261.0    |
| 221.00 | 16664.0  |
| 222.00 | 3663.0   |
| 223.00 | 6738.0   |
| 224.10 | 58064.0  |
| 225.10 | 15676.0  |
| 226.00 | 1542.0   |
| 227.00 | 25200.0  |
| 228.00 | 3391.0   |
| 229.00 | 5602.0   |
| 230.00 | 655.0    |
| 231.00 | 2172.0   |
| 232.00 | 384.0    |
| 233.00 | 447.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP257



|        |          |
|--------|----------|
| 234.00 | 1120.0   |
| 235.00 | 1559.0   |
| 236.00 | 1079.0   |
| 237.00 | 2355.0   |
| 238.00 | 369.0    |
| 239.00 | 943.0    |
| 240.00 | 747.0    |
| 240.90 | 1260.0   |
| 242.00 | 2994.0   |
| 243.10 | 3758.0   |
| 244.00 | 45344.0  |
| 245.00 | 6587.0   |
| 246.00 | 9106.0   |
| 247.00 | 1884.0   |
| 248.00 | 340.0    |
| 249.00 | 1525.0   |
| 249.90 | 289.0    |
| 250.90 | 367.0    |
| 252.00 | 397.0    |
| 253.00 | 928.0    |
| 255.00 | 251584.0 |
| 256.00 | 33896.0  |
| 257.10 | 2715.0   |
| 258.00 | 14352.0  |
| 259.00 | 2230.0   |
| 259.90 | 431.0    |
| 260.90 | 387.0    |
| 264.10 | 348.0    |
| 265.00 | 5794.0   |
| 266.00 | 619.0    |
| 270.00 | 257.0    |
| 270.90 | 607.0    |
| 272.00 | 620.0    |
| 273.00 | 7347.0   |
| 274.00 | 19416.0  |
| 275.00 | 105232.0 |
| 276.10 | 12446.0  |
| 277.00 | 8973.0   |
| 278.00 | 1819.0   |
| 279.10 | 298.0    |
| 281.10 | 463.0    |
| 282.00 | 346.0    |
| 283.10 | 1061.0   |
| 284.00 | 703.0    |
| 285.00 | 1469.0   |
| 285.90 | 279.0    |
| 289.00 | 391.0    |
| 290.00 | 314.0    |
| 291.00 | 219.0    |
| 292.00 | 518.0    |
| 293.00 | 1884.0   |
| 293.90 | 599.0    |
| 296.00 | 31120.0  |
| 297.00 | 4596.0   |
| 298.00 | 278.0    |
| 301.00 | 467.0    |
| 302.00 | 597.0    |
| 303.00 | 4040.0   |
| 304.00 | 1055.0   |
| 305.00 | 153.0    |
| 308.00 | 403.0    |
| 309.00 | 259.0    |
| 310.00 | 477.0    |
| 313.00 | 248.0    |
| 314.10 | 1257.0   |
| 314.90 | 3271.0   |
| 316.00 | 1868.0   |
| 317.00 | 384.0    |
| 321.00 | 1033.0   |
| 322.00 | 552.0    |
| 323.10 | 10906.0  |
| 324.00 | 2128.0   |
| 325.00 | 242.0    |
| 327.00 | 1786.0   |
| 328.00 | 969.0    |
| 329.00 | 248.0    |
| 332.10 | 685.0    |
| 333.00 | 915.0    |
| 334.10 | 6995.0   |
| 335.00 | 1876.0   |
| 336.00 | 217.0    |
| 338.90 | 161.0    |
| 341.00 | 1270.0   |
| 342.00 | 295.0    |
| 345.90 | 2566.0   |
| 346.90 | 536.0    |
| 350.90 | 202.0    |
| 352.00 | 3268.0   |
| 353.00 | 2142.0   |
| 354.00 | 2924.0   |
| 355.10 | 676.0    |
| 359.00 | 277.0    |
| 365.00 | 14130.0  |
| 365.90 | 2043.0   |
| 367.00 | 154.0    |
| 369.90 | 288.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP257

|        |          |
|--------|----------|
| 371.00 | 976.0    |
| 372.00 | 5858.0   |
| 373.00 | 1481.0   |
| 377.00 | 214.0    |
| 383.00 | 1454.0   |
| 384.10 | 542.0    |
| 390.10 | 618.0    |
| 391.10 | 426.0    |
| 392.00 | 355.0    |
| 401.00 | 447.0    |
| 402.00 | 2337.0   |
| 403.00 | 3904.0   |
| 403.90 | 1150.0   |
| 404.90 | 203.0    |
| 421.00 | 3233.0   |
| 422.00 | 2414.0   |
| 423.00 | 19400.0  |
| 424.10 | 5461.0   |
| 424.90 | 492.0    |
| 441.10 | 47864.0  |
| 442.00 | 340608.0 |
| 443.00 | 55240.0  |
| 444.00 | 5696.0   |
| 445.00 | 428.0    |

**Instrument :**  
BNA\_N  
**ClientSampleId :**  
DFTPP257

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 431.0     |
| 38.00  | 1056.0    |
| 39.00  | 6666.0    |
| 40.00  | 1336.0    |
| 41.00  | 582.0     |
| 42.00  | 163.0     |
| 43.00  | 701.0     |
| 44.00  | 780.0     |
| 45.00  | 226.0     |
| 49.00  | 768.0     |
| 50.00  | 22296.0   |
| 51.00  | 85296.0   |
| 52.00  | 4477.0    |
| 53.00  | 248.0     |
| 55.00  | 1096.0    |
| 56.00  | 2416.0    |
| 57.00  | 6411.0    |
| 57.90  | 345.0     |
| 61.00  | 1247.0    |
| 62.00  | 1291.0    |
| 63.00  | 3495.0    |
| 64.00  | 518.0     |
| 65.00  | 1823.0    |
| 66.00  | 169.0     |
| 67.00  | 288.0     |
| 68.00  | 1519.0    |
| 69.00  | 92976.0   |
| 70.00  | 621.0     |
| 71.00  | 407.0     |
| 73.00  | 1238.0    |
| 74.00  | 9471.0    |
| 75.00  | 15451.0   |
| 76.00  | 5590.0    |
| 77.00  | 113368.0  |
| 78.00  | 8328.0    |
| 79.00  | 7272.0    |
| 80.00  | 5575.0    |
| 81.00  | 8184.0    |
| 82.00  | 1700.0    |
| 83.00  | 1972.0    |
| 84.00  | 490.0     |
| 85.00  | 1688.0    |
| 86.00  | 2399.0    |
| 87.00  | 1121.0    |
| 88.00  | 482.0     |
| 89.00  | 226.0     |
| 91.00  | 1703.0    |
| 92.00  | 1999.0    |
| 93.00  | 12466.0   |
| 93.90  | 831.0     |
| 95.00  | 327.0     |
| 96.00  | 711.0     |
| 96.90  | 429.0     |
| 98.00  | 9606.0    |
| 99.00  | 8382.0    |
| 100.00 | 713.0     |
| 101.00 | 4688.0    |
| 101.90 | 241.0     |
| 103.00 | 1363.0    |
| 104.00 | 2805.0    |
| 105.00 | 2644.0    |
| 106.00 | 709.0     |
| 107.00 | 34968.0   |
| 108.00 | 5087.0    |
| 109.00 | 930.0     |
| 110.00 | 62720.0   |
| 111.00 | 9185.0    |
| 112.00 | 1361.0    |
| 113.00 | 478.0     |
| 114.90 | 212.0     |
| 116.00 | 1630.0    |
| 117.00 | 29576.0   |
| 118.00 | 2251.0    |
| 118.90 | 318.0     |
| 120.00 | 527.0     |
| 121.00 | 182.0     |
| 122.00 | 2495.0    |
| 123.00 | 3387.0    |
| 124.00 | 1536.0    |
| 124.90 | 1540.0    |
| 126.10 | 536.0     |
| 127.00 | 122136.0  |
| 128.00 | 9161.0    |
| 129.00 | 45456.0   |
| 130.00 | 4166.0    |
| 131.00 | 905.0     |
| 131.90 | 348.0     |
| 133.00 | 340.0     |
| 133.90 | 1211.0    |
| 135.00 | 3828.0    |
| 136.00 | 1446.0    |
| 137.10 | 2161.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP257

|        |          |
|--------|----------|
| 137.90 | 466.0    |
| 139.00 | 329.0    |
| 139.90 | 588.0    |
| 141.00 | 5830.0   |
| 142.00 | 1836.0   |
| 143.00 | 1385.0   |
| 143.90 | 357.0    |
| 145.00 | 319.0    |
| 146.00 | 1027.0   |
| 147.00 | 3210.0   |
| 148.00 | 8051.0   |
| 149.00 | 1610.0   |
| 150.00 | 446.0    |
| 151.10 | 813.0    |
| 151.60 | 504.0    |
| 152.90 | 1967.0   |
| 153.90 | 1442.0   |
| 155.00 | 3501.0   |
| 156.00 | 5082.0   |
| 157.00 | 865.0    |
| 157.90 | 1135.0   |
| 159.00 | 821.0    |
| 160.00 | 2007.0   |
| 161.00 | 2928.0   |
| 162.00 | 870.0    |
| 162.90 | 280.0    |
| 163.90 | 383.0    |
| 165.00 | 2427.0   |
| 166.00 | 1878.0   |
| 167.00 | 14217.0  |
| 168.00 | 7797.0   |
| 169.00 | 1076.0   |
| 170.00 | 432.0    |
| 171.00 | 522.0    |
| 172.00 | 1055.0   |
| 173.00 | 1494.0   |
| 174.00 | 2578.0   |
| 175.00 | 5010.0   |
| 176.00 | 1396.0   |
| 176.90 | 2150.0   |
| 177.90 | 648.0    |
| 179.00 | 10069.0  |
| 180.00 | 6769.0   |
| 181.00 | 3071.0   |
| 182.00 | 495.0    |
| 182.90 | 307.0    |
| 183.90 | 632.0    |
| 185.00 | 4608.0   |
| 186.00 | 33320.0  |
| 187.00 | 10052.0  |
| 188.00 | 1149.0   |
| 189.00 | 2194.0   |
| 190.00 | 351.0    |
| 190.90 | 1244.0   |
| 192.00 | 2951.0   |
| 193.00 | 3959.0   |
| 193.90 | 774.0    |
| 195.00 | 421.0    |
| 196.00 | 7392.0   |
| 197.00 | 1138.0   |
| 198.00 | 260736.0 |
| 198.90 | 16624.0  |
| 200.00 | 1392.0   |
| 201.40 | 1010.0   |
| 203.00 | 1656.0   |
| 204.00 | 9127.0   |
| 205.00 | 15963.0  |
| 206.00 | 68128.0  |
| 207.10 | 9806.0   |
| 208.00 | 2190.0   |
| 209.00 | 873.0    |
| 210.20 | 1194.0   |
| 210.90 | 2429.0   |
| 213.00 | 206.0    |
| 215.00 | 768.0    |
| 216.00 | 1540.0   |
| 217.00 | 17856.0  |
| 218.00 | 2405.0   |
| 219.00 | 236.0    |
| 219.90 | 331.0    |
| 221.00 | 10376.0  |
| 221.80 | 2554.0   |
| 223.00 | 4389.0   |
| 224.00 | 36576.0  |
| 225.00 | 9807.0   |
| 226.00 | 1018.0   |
| 227.00 | 14375.0  |
| 228.00 | 1943.0   |
| 229.00 | 2975.0   |
| 230.00 | 472.0    |
| 231.00 | 1401.0   |
| 233.00 | 275.0    |
| 234.00 | 988.0    |
| 235.00 | 1120.0   |
| 235.90 | 771.0    |
| 237.10 | 1509.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP257

|        |          |
|--------|----------|
| 237.90 | 240.0    |
| 239.00 | 723.0    |
| 240.00 | 558.0    |
| 240.90 | 878.0    |
| 242.10 | 1886.0   |
| 243.00 | 2368.0   |
| 244.00 | 32032.0  |
| 245.00 | 4142.0   |
| 246.00 | 5719.0   |
| 247.00 | 1269.0   |
| 247.90 | 267.0    |
| 249.00 | 1096.0   |
| 249.90 | 196.0    |
| 251.00 | 250.0    |
| 251.90 | 180.0    |
| 253.00 | 697.0    |
| 254.00 | 1009.0   |
| 255.00 | 160000.0 |
| 256.00 | 22696.0  |
| 257.10 | 1896.0   |
| 258.00 | 8924.0   |
| 258.90 | 1404.0   |
| 260.10 | 289.0    |
| 260.90 | 270.0    |
| 263.90 | 217.0    |
| 265.00 | 3285.0   |
| 266.00 | 621.0    |
| 266.90 | 161.0    |
| 269.90 | 205.0    |
| 271.00 | 454.0    |
| 271.90 | 526.0    |
| 273.00 | 5103.0   |
| 274.00 | 12745.0  |
| 275.00 | 70720.0  |
| 276.10 | 9706.0   |
| 277.00 | 5651.0   |
| 278.00 | 1294.0   |
| 279.00 | 241.0    |
| 281.00 | 377.0    |
| 281.90 | 224.0    |
| 283.00 | 918.0    |
| 284.00 | 365.0    |
| 285.10 | 1093.0   |
| 285.90 | 203.0    |
| 288.90 | 254.0    |
| 289.90 | 221.0    |
| 292.00 | 279.0    |
| 293.00 | 1492.0   |
| 294.00 | 502.0    |
| 294.90 | 330.0    |
| 296.00 | 22568.0  |
| 297.00 | 2953.0   |
| 298.00 | 202.0    |
| 301.10 | 280.0    |
| 302.00 | 352.0    |
| 303.00 | 2573.0   |
| 304.20 | 629.0    |
| 307.90 | 365.0    |
| 308.90 | 196.0    |
| 310.10 | 301.0    |
| 313.00 | 200.0    |
| 314.10 | 1037.0   |
| 315.00 | 2374.0   |
| 316.00 | 1345.0   |
| 317.00 | 263.0    |
| 321.00 | 800.0    |
| 322.10 | 272.0    |
| 323.00 | 7411.0   |
| 324.00 | 1217.0   |
| 325.00 | 165.0    |
| 326.90 | 1285.0   |
| 327.90 | 868.0    |
| 329.00 | 215.0    |
| 332.00 | 544.0    |
| 333.00 | 636.0    |
| 334.00 | 4774.0   |
| 335.00 | 1278.0   |
| 336.00 | 173.0    |
| 341.10 | 748.0    |
| 342.00 | 247.0    |
| 346.00 | 1503.0   |
| 346.90 | 328.0    |
| 352.00 | 2137.0   |
| 353.00 | 1635.0   |
| 354.00 | 1853.0   |
| 355.00 | 578.0    |
| 359.00 | 223.0    |
| 365.00 | 10270.0  |
| 365.90 | 1461.0   |
| 370.90 | 574.0    |
| 372.00 | 4489.0   |
| 373.00 | 1073.0   |
| 374.00 | 175.0    |
| 383.10 | 1044.0   |
| 384.00 | 337.0    |
| 389.90 | 502.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP257

|        |          |
|--------|----------|
| 391.00 | 270.0    |
| 392.10 | 232.0    |
| 401.10 | 247.0    |
| 402.00 | 1806.0   |
| 403.00 | 2730.0   |
| 404.00 | 911.0    |
| 404.90 | 232.0    |
| 420.90 | 1825.0   |
| 422.00 | 2065.0   |
| 423.00 | 14561.0  |
| 424.10 | 4015.0   |
| 424.90 | 485.0    |
| 441.10 | 42856.0  |
| 442.00 | 259456.0 |
| 443.10 | 60376.0  |
| 444.00 | 4403.0   |
| 445.00 | 266.0    |

**Instrument :**  
BNA\_N  
**ClientSampleId :**  
DFTPP257