


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102922\
 Data File : BN022379.D
 Acq On : 28 Oct 2022 10:58
 Operator : CG/JU
 Sample : DFTPP258
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP258

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN101222.M
 Title : SVOA CALIBRATION
 Last Update : Sat Oct 29 00:50:42 2022



AutoFind: Scans 2517, 2518, 2519; Background Corrected with Scan 2511

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.8	191705	PASS
68	69	0.00	2	1.6	3423	PASS
69	198	0.00	100	39.2	210187	PASS
70	69	0.00	2	0.4	928	PASS
127	198	10	80	50.4	269888	PASS
197	198	0.00	2	0.3	1852	PASS
198	198	100	100	100.0	535872	PASS
199	198	5	9	6.6	35387	PASS
275	198	10	60	26.6	142720	PASS
365	198	1	100	3.9	20923	PASS
441	198	0.01	100	13.4	72069	PASS
442	198	50	100	86.8	465237	PASS
443	442	15	24	20.8	96707	PASS

m/z	Abundance
39.00	306.0
40.00	1245.0
41.00	739.0
42.10	262.0
43.00	1154.0
44.00	1292.0
45.00	365.0
48.90	156.0
51.00	188.0
55.00	765.0
56.00	343.0
57.00	872.0
60.00	253.0
67.10	215.0
68.00	156.0
69.00	522.0
70.00	367.0
71.00	500.0
73.00	692.0
74.00	172.0
75.00	231.0
76.90	233.0
77.90	170.0
81.00	187.0
82.10	153.0
83.00	392.0
83.90	437.0
85.00	281.0
91.00	193.0
95.00	185.0
95.80	262.0
97.10	415.0
111.00	162.0
133.00	258.0
134.90	186.0
147.00	227.0
191.00	251.0
192.90	188.0
207.00	1402.0
208.10	277.0
209.00	253.0
281.00	455.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

m/z	Abundance
37.10	917.0
38.00	2352.0
39.10	14021.0
40.00	1299.0
41.00	862.0
42.00	223.0
43.00	782.0
44.00	949.0
44.90	604.0
49.00	1425.0
50.00	53784.0
51.00	203456.0
52.00	10136.0
53.00	542.0
55.00	1419.0
56.00	6222.0
57.00	13755.0
58.00	684.0
58.90	189.0
60.00	249.0
61.00	2671.0
62.00	3074.0
63.00	8466.0
64.00	1063.0
65.00	4190.0
66.00	345.0
66.90	382.0
68.00	3621.0
69.00	223680.0
70.00	1289.0
71.00	381.0
72.00	218.0
73.00	2336.0
74.00	23280.0
75.00	36608.0
76.00	12819.0
77.10	263360.0
78.00	19216.0
79.00	16872.0
80.00	12595.0
81.00	18608.0
82.00	4478.0
83.00	4335.0
84.00	610.0
85.00	3395.0
86.00	5495.0
87.00	2387.0
88.00	1038.0
89.00	522.0
91.00	4127.0
92.00	4745.0
93.00	30064.0
94.00	2216.0
94.90	567.0
96.00	1301.0
97.00	754.0
98.00	23944.0
99.00	19280.0
100.00	1804.0
101.00	10437.0
102.00	571.0
103.00	3410.0
104.00	6427.0
105.00	5771.0
106.00	2158.0
107.00	80816.0
108.00	12313.0
109.00	2449.0
110.00	142528.0
111.00	22344.0
112.00	2801.0
113.00	970.0
114.20	213.0
114.90	369.0
116.00	4297.0
117.00	71288.0
118.00	5527.0
119.00	687.0
119.90	1066.0
121.10	416.0
122.00	5123.0
123.00	7793.0
124.00	3576.0
125.00	3327.0
126.00	1093.0
127.00	272640.0
128.00	21208.0
129.00	102248.0
130.00	9021.0
131.00	1714.0
132.00	959.0
133.10	474.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

134.00	2866.0
135.00	8298.0
136.00	3495.0
137.00	4616.0
137.90	961.0
138.90	571.0
140.00	1332.0
141.00	13687.0
142.00	4658.0
143.00	3153.0
144.00	979.0
145.10	810.0
146.00	2440.0
147.00	7053.0
148.00	17208.0
149.00	3595.0
150.00	1081.0
151.10	1781.0
151.70	1171.0
153.00	4435.0
154.00	3437.0
155.00	7676.0
156.00	11107.0
157.10	1969.0
158.00	2454.0
159.00	1950.0
160.00	3809.0
161.00	6725.0
162.00	1944.0
163.00	610.0
164.00	611.0
165.00	5071.0
166.00	4023.0
167.00	29632.0
168.00	16800.0
169.00	2756.0
169.90	827.0
171.00	964.0
171.90	2460.0
173.00	3131.0
174.00	5910.0
175.00	10812.0
176.00	3079.0
177.00	4910.0
178.00	1802.0
179.00	19864.0
180.00	14271.0
181.00	6540.0
182.00	982.0
182.90	535.0
184.00	1654.0
185.00	9621.0
186.00	75208.0
187.00	20968.0
188.00	2127.0
189.00	5056.0
190.00	708.0
191.00	2168.0
192.00	6521.0
193.00	7363.0
194.10	1461.0
195.10	682.0
196.00	14962.0
197.00	3074.0
198.00	523968.0
199.00	34840.0
200.00	2824.0
201.40	2169.0
203.00	3486.0
204.10	19352.0
205.00	30600.0
206.10	130976.0
207.00	18784.0
208.00	4941.0
209.00	1613.0
210.10	2098.0
211.00	4948.0
212.90	375.0
215.00	1519.0
216.10	3091.0
217.00	34984.0
218.00	4466.0
219.00	432.0
221.10	18896.0
221.80	5603.0
223.00	8178.0
224.00	70128.0
225.00	19352.0
226.10	2249.0
227.00	32624.0
228.00	4766.0
229.00	6647.0
230.00	837.0
231.00	2542.0
232.00	374.0
233.00	627.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

234.00	2000.0
235.00	2375.0
236.00	1618.0
237.00	2957.0
238.00	425.0
239.00	1254.0
240.00	1136.0
241.00	1617.0
242.00	3877.0
243.10	4158.0
244.00	62600.0
245.10	7893.0
246.00	11640.0
247.00	2496.0
248.00	441.0
249.00	1920.0
250.00	332.0
251.00	402.0
252.00	583.0
253.00	1161.0
254.00	2048.0
255.00	298112.0
256.00	43112.0
257.00	3569.0
258.00	17608.0
259.00	2951.0
259.90	523.0
261.00	422.0
263.10	166.0
264.00	429.0
265.00	7154.0
266.00	1059.0
267.10	151.0
270.00	367.0
271.00	573.0
272.00	758.0
273.00	8548.0
274.00	24080.0
275.00	130688.0
276.00	16672.0
277.10	12132.0
278.00	2236.0
278.90	387.0
281.00	471.0
282.00	325.0
283.10	1143.0
284.00	809.0
285.10	2128.0
286.00	414.0
289.00	561.0
290.00	344.0
291.00	364.0
292.00	433.0
293.00	2415.0
294.00	632.0
295.10	896.0
296.00	38016.0
297.00	5696.0
298.00	359.0
301.00	462.0
302.00	569.0
303.00	4608.0
304.10	1194.0
304.90	185.0
307.90	576.0
308.90	297.0
310.00	448.0
311.90	167.0
313.00	228.0
314.00	1572.0
315.00	3860.0
316.00	2457.0
317.10	458.0
321.00	1046.0
322.10	625.0
323.00	12578.0
324.00	2334.0
325.00	274.0
326.00	253.0
327.00	2479.0
328.00	1301.0
329.00	295.0
332.00	795.0
333.10	1043.0
334.00	8533.0
335.00	2102.0
336.00	270.0
338.90	234.0
341.10	1309.0
342.00	399.0
345.90	2531.0
347.00	582.0
351.00	174.0
352.00	3880.0
353.00	2581.0
354.10	3693.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

355.10	900.0
358.90	250.0
364.10	154.0
365.00	16960.0
366.00	3024.0
367.00	191.0
370.10	250.0
370.80	758.0
371.10	713.0
372.10	6408.0
373.00	1718.0
374.00	163.0
383.00	1988.0
384.00	541.0
390.00	809.0
390.90	521.0
392.00	498.0
401.00	329.0
402.10	2485.0
403.00	3560.0
404.10	1229.0
405.00	341.0
415.10	213.0
421.10	3292.0
422.00	3376.0
423.10	24056.0
424.00	5888.0
425.00	916.0
441.00	62256.0
442.10	345152.0
443.10	64472.0
444.10	6183.0
445.10	340.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

m/z	Abundance
35.90	169.0
37.00	996.0
38.10	2598.0
39.10	14506.0
40.00	1287.0
41.10	870.0
41.90	183.0
43.00	610.0
44.00	1004.0
45.00	600.0
48.00	188.0
49.00	1616.0
50.00	57096.0
51.00	213952.0
52.00	10719.0
53.00	528.0
55.00	1680.0
56.00	6527.0
57.00	15030.0
58.00	699.0
59.00	282.0
60.00	312.0
61.00	3021.0
62.00	3361.0
63.00	8817.0
64.00	1172.0
65.00	4414.0
65.90	359.0
67.00	393.0
68.00	4139.0
69.00	235904.0
70.00	1501.0
71.00	349.0
72.00	171.0
73.00	2559.0
74.00	25512.0
75.00	39360.0
76.00	13669.0
77.10	287552.0
78.00	19760.0
79.00	18616.0
80.00	13577.0
81.00	20424.0
82.00	5040.0
83.00	4755.0
84.00	675.0
85.00	4015.0
86.00	6109.0
87.00	2421.0
88.00	1114.0
89.00	486.0
91.00	4332.0
92.00	5233.0
93.00	33320.0
94.00	2281.0
95.00	495.0
96.00	1484.0
97.10	737.0
98.00	26144.0
99.00	19368.0
100.00	1840.0
101.00	11353.0
102.00	651.0
103.00	3527.0
104.00	7451.0
105.00	7110.0
106.00	2376.0
107.00	88232.0
108.00	13889.0
109.00	2524.0
110.00	161920.0
111.00	24232.0
112.00	3035.0
113.00	930.0
114.10	277.0
114.90	329.0
116.00	5094.0
117.00	79120.0
118.00	5861.0
119.00	883.0
120.00	1156.0
121.00	393.0
122.00	6097.0
123.00	9371.0
123.90	4005.0
125.00	4004.0
126.10	1600.0
127.00	306560.0
128.00	23360.0
129.00	115136.0
130.00	10308.0
131.00	2018.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

132.00	991.0
132.90	568.0
134.00	3430.0
135.00	9994.0
136.00	3686.0
137.00	5104.0
138.00	1044.0
139.00	588.0
140.00	1447.0
141.00	15250.0
142.00	4783.0
143.00	3576.0
143.90	848.0
145.00	798.0
146.00	3093.0
147.00	8360.0
148.00	18744.0
149.00	3963.0
150.00	1131.0
151.10	1887.0
151.80	1200.0
153.00	5252.0
154.00	3613.0
155.00	8522.0
156.00	12978.0
157.10	2511.0
157.90	2720.0
159.00	2408.0
160.00	4960.0
161.00	7310.0
162.00	2126.0
162.90	589.0
164.00	717.0
165.00	5616.0
166.00	4451.0
167.00	31592.0
168.00	17272.0
169.00	2756.0
170.00	976.0
171.00	1265.0
172.00	2851.0
173.00	3838.0
174.00	6420.0
175.10	12132.0
176.10	3339.0
177.00	5471.0
178.00	2107.0
179.00	23728.0
180.00	16592.0
181.00	7399.0
182.10	1191.0
182.90	631.0
183.90	1787.0
185.10	11247.0
186.10	87064.0
187.00	24816.0
188.00	2581.0
189.00	5249.0
189.90	970.0
191.00	2824.0
192.00	7390.0
193.10	8569.0
194.00	1668.0
196.00	17600.0
198.00	609408.0
199.00	41120.0
200.00	3453.0
201.50	2190.0
203.00	4135.0
204.00	22664.0
205.00	37336.0
206.10	155776.0
207.10	22160.0
208.00	5603.0
209.00	1892.0
210.20	2577.0
211.00	6092.0
211.70	1298.0
213.00	469.0
214.10	207.0
215.00	1823.0
216.10	3451.0
217.00	42096.0
218.00	5544.0
219.10	567.0
221.00	23016.0
221.90	5639.0
223.00	8965.0
224.10	88792.0
225.10	23024.0
226.10	2426.0
227.00	39832.0
228.00	5415.0
229.00	7230.0
230.00	1130.0
231.00	3053.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

231.90	426.0
233.00	687.0
234.00	2316.0
235.00	2765.0
236.00	1791.0
237.00	3108.0
238.00	501.0
239.00	1450.0
240.00	1192.0
240.90	2146.0
242.10	5296.0
243.00	5272.0
244.10	73376.0
245.10	9131.0
246.00	14044.0
247.00	2873.0
248.00	699.0
249.00	2482.0
250.00	383.0
251.00	560.0
252.00	565.0
253.10	1469.0
255.00	363776.0
256.10	52352.0
257.00	4060.0
258.00	20792.0
259.00	3452.0
260.00	729.0
261.00	522.0
262.00	155.0
263.10	242.0
264.10	645.0
265.10	9078.0
266.00	1428.0
267.00	192.0
270.00	429.0
270.90	703.0
272.10	1087.0
273.10	10597.0
274.10	28160.0
275.00	152512.0
276.00	21528.0
277.00	14138.0
278.00	2348.0
279.00	486.0
281.00	430.0
282.10	516.0
283.00	1671.0
284.00	1096.0
285.10	2401.0
286.10	385.0
287.90	166.0
288.90	622.0
290.00	499.0
291.00	366.0
292.00	708.0
293.00	3402.0
294.00	1007.0
296.00	45624.0
297.10	7226.0
298.00	407.0
301.00	574.0
302.00	719.0
303.10	5736.0
304.10	1591.0
305.00	180.0
308.00	739.0
309.00	543.0
310.00	669.0
312.10	225.0
313.00	431.0
314.00	2145.0
315.00	4760.0
316.10	2873.0
317.10	629.0
320.00	177.0
321.00	1606.0
322.00	746.0
323.00	15092.0
324.00	2450.0
325.00	287.0
326.10	399.0
327.00	3382.0
328.00	1414.0
329.10	300.0
332.10	1038.0
333.00	1583.0
334.00	10161.0
335.00	2476.0
336.10	407.0
339.00	343.0
340.00	293.0
341.00	1921.0
342.00	349.0
346.00	3628.0
347.00	845.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

351.00	306.0
352.00	5440.0
353.00	3669.0
354.10	5024.0
355.00	1087.0
359.10	403.0
365.00	26104.0
366.00	3880.0
367.00	330.0
370.00	464.0
371.00	1482.0
372.10	8286.0
373.00	2210.0
374.10	281.0
377.00	205.0
383.00	2663.0
384.00	767.0
385.10	188.0
390.00	1395.0
391.00	655.0
391.90	508.0
401.00	488.0
402.00	3716.0
403.00	4733.0
404.10	1936.0
405.10	361.0
415.00	190.0
421.10	5249.0
422.10	5004.0
423.00	32096.0
424.00	8255.0
425.00	800.0
440.10	386.0
441.00	67760.0
442.00	502656.0
443.10	120432.0
444.00	8441.0
445.00	536.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

m/z	Abundance
35.90	170.0
37.10	771.0
38.10	1939.0
39.10	10898.0
40.00	1264.0
41.10	854.0
42.00	160.0
43.00	794.0
44.00	1151.0
45.00	515.0
49.00	1249.0
50.00	39784.0
51.00	158272.0
52.00	8108.0
53.00	389.0
55.00	1331.0
56.00	4675.0
57.00	11135.0
58.00	688.0
58.90	205.0
59.90	278.0
61.00	2111.0
62.00	2442.0
63.00	6569.0
64.00	884.0
65.00	3349.0
66.00	262.0
67.00	352.0
68.00	2976.0
69.00	172544.0
70.00	1094.0
71.00	427.0
71.90	194.0
73.00	2138.0
74.00	19136.0
75.00	28720.0
76.00	9908.0
77.00	214464.0
78.10	14681.0
79.00	13241.0
80.00	9801.0
81.00	14821.0
82.00	3639.0
83.00	3646.0
83.90	697.0
85.00	3164.0
86.00	4495.0
87.00	2083.0
87.90	817.0
88.90	377.0
91.00	3300.0
92.00	3675.0
93.00	24208.0
94.00	1814.0
95.00	548.0
96.10	1258.0
96.90	581.0
98.00	18608.0
99.00	14726.0
100.00	1364.0
101.00	8342.0
102.00	437.0
103.00	2848.0
104.00	5181.0
105.00	4818.0
106.00	1643.0
107.00	65784.0
108.00	9683.0
109.00	1879.0
110.00	116856.0
111.00	17160.0
111.90	2059.0
113.00	745.0
114.10	188.0
115.00	319.0
116.00	3610.0
117.00	56392.0
118.00	4104.0
119.00	580.0
120.00	906.0
120.90	389.0
122.00	4283.0
123.00	6725.0
124.00	3334.0
125.00	2921.0
127.00	230464.0
128.00	16792.0
129.00	89864.0
130.00	7303.0
131.00	1558.0
132.00	656.0
132.90	510.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

134.00	2567.0
135.00	7398.0
136.00	2803.0
137.00	4202.0
138.00	861.0
139.00	565.0
140.00	1136.0
141.00	11493.0
142.00	3697.0
143.00	2765.0
144.00	739.0
144.90	725.0
146.00	2066.0
147.00	5988.0
148.00	14712.0
149.00	3071.0
150.00	966.0
151.10	1552.0
151.90	898.0
153.00	3723.0
154.00	2880.0
155.00	6852.0
156.00	10516.0
157.00	1875.0
157.90	2240.0
158.90	1554.0
160.00	3615.0
161.00	5742.0
162.00	1641.0
163.00	458.0
164.00	662.0
165.00	4183.0
166.00	3490.0
167.00	24648.0
168.00	14733.0
169.10	2143.0
169.90	770.0
171.00	1162.0
171.90	2069.0
173.00	2899.0
174.00	4990.0
175.00	9098.0
176.00	2555.0
177.00	4411.0
178.00	1642.0
179.00	17736.0
180.00	12199.0
181.00	5593.0
182.10	1029.0
182.90	562.0
184.00	1469.0
185.00	8380.0
186.10	63904.0
187.00	19576.0
188.00	1993.0
189.00	4374.0
190.00	731.0
191.00	1905.0
192.00	5909.0
193.00	6753.0
194.10	1384.0
195.00	636.0
196.10	13808.0
197.00	2483.0
198.00	474240.0
199.00	30200.0
200.00	2359.0
201.50	2385.0
203.00	3619.0
204.10	17920.0
205.00	29024.0
206.10	128976.0
207.00	17712.0
208.00	4531.0
209.00	1562.0
210.20	2143.0
211.00	5122.0
213.00	369.0
214.00	182.0
215.00	1711.0
216.00	2685.0
217.00	32368.0
218.00	4450.0
219.10	435.0
221.10	18752.0
223.00	7949.0
224.10	72552.0
225.10	17608.0
226.10	1817.0
227.00	30440.0
228.00	3884.0
229.00	6342.0
229.90	756.0
231.00	2570.0
231.90	373.0
233.00	495.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

234.00	1940.0
235.00	2142.0
236.00	1364.0
237.00	2749.0
238.00	421.0
239.00	1277.0
240.00	934.0
241.00	1709.0
242.10	3714.0
243.10	4017.0
244.10	57112.0
245.10	7194.0
246.00	11880.0
247.00	2293.0
247.90	437.0
249.00	1896.0
250.00	414.0
251.00	611.0
252.00	522.0
253.00	1236.0
254.10	2192.0
255.00	283648.0
256.00	42920.0
257.00	3032.0
258.00	17968.0
259.00	2655.0
260.00	561.0
260.90	563.0
264.00	400.0
265.00	7839.0
266.00	1019.0
267.00	179.0
269.90	312.0
271.00	620.0
272.10	756.0
273.00	9417.0
274.00	24800.0
275.10	144960.0
276.00	17448.0
277.00	12852.0
278.00	2153.0
279.00	359.0
281.00	399.0
281.90	307.0
283.00	1583.0
284.00	1048.0
285.10	1978.0
286.00	437.0
289.00	473.0
290.10	417.0
291.00	327.0
292.00	619.0
293.00	2732.0
294.00	781.0
296.00	44080.0
297.10	5736.0
298.00	455.0
301.00	520.0
302.00	636.0
303.10	4761.0
304.00	1506.0
304.90	186.0
308.00	568.0
309.00	498.0
310.00	569.0
312.00	159.0
313.10	520.0
314.00	1974.0
315.00	4723.0
316.00	2888.0
317.00	535.0
319.90	165.0
321.00	1363.0
322.00	735.0
323.10	14287.0
324.00	2398.0
325.00	264.0
325.90	233.0
327.00	2667.0
328.00	1452.0
328.90	308.0
332.00	1114.0
333.10	1377.0
334.00	8588.0
335.10	2264.0
335.90	279.0
339.10	210.0
341.00	1697.0
342.00	515.0
346.00	3567.0
347.00	581.0
350.90	208.0
352.10	4147.0
353.00	2845.0
354.10	4607.0
355.00	935.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258

358.90	403.0
363.90	195.0
365.00	19704.0
366.00	3057.0
367.00	189.0
370.00	380.0
371.10	1280.0
372.00	7870.0
373.00	1776.0
374.00	252.0
376.90	262.0
383.00	2263.0
384.00	708.0
385.00	245.0
390.00	1057.0
390.90	838.0
392.00	595.0
401.00	511.0
402.00	3851.0
403.00	5604.0
403.90	1864.0
405.00	350.0
410.00	167.0
414.90	275.0
421.10	3671.0
421.90	3700.0
423.10	35128.0
424.10	6049.0
425.00	761.0
440.10	467.0
441.10	86192.0
442.10	547904.0
443.00	105216.0
444.00	8886.0
445.00	472.0

Instrument :
BNA_N
ClientSampleId :
DFTPP258