

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103020\
 Data File : BN012484.D
 Acq On : 31 Oct 2020 02:15
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 31 05:39:00 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN103020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 14:49:10 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.570	152	1206	0.40	ng	0.00
7) Naphthalene-d8	10.339	136	5256	0.40	ng	# 0.00
13) Acenaphthene-d10	14.204	164	2900	0.40	ng	-0.01
19) Phenanthrene-d10	16.956	188	5930	0.40	ng	#-0.01
29) Chrysene-d12	21.163	240	5709	0.40	ng	#-0.01
36) Perylene-d12	23.333	264	6533	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.186	112	1562	0.37	ng	0.00
5) Phenol-d6	6.757	99	1971	0.38	ng	0.00
8) Nitrobenzene-d5	8.716	82	2046	0.34	ng	-0.01
11) 2-Methylnaphthalene-d10	11.935	152	3033	0.35	ng	0.00
14) 2,4,6-Tribromophenol	15.714	330	659	0.34	ng	0.00
15) 2-Fluorobiphenyl	12.822	172	3861	0.35	ng	-0.01
27) Fluoranthene-d10	19.003	212	5992	0.33	ng	0.00
31) Terphenyl-d14	19.608	244	4640	0.35	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.165	88	740	0.37	ng	96
3) n-Nitrosodimethylamine	3.479	42	1536	0.31	ng	# 77
6) bis(2-Chloroethyl)ether	7.014	93	1218	0.36	ng	# 86
9) Naphthalene	10.389	128	5274	0.36	ng	98
10) Hexachlorobutadiene	10.668	225	1082	0.35	ng	# 98
12) 2-Methylnaphthalene	12.011	142	3440	0.36	ng	# 91
16) Acenaphthylene	13.925	152	4953	0.34	ng	99
17) Acenaphthene	14.268	154	3216	0.35	ng	91
18) Fluorene	15.270	166	3928	0.34	ng	99
20) 4,6-Dinitro-2-methylph...	15.372	198	384	0.34	ng	# 28
21) 4-Bromophenyl-phenylether	16.165	248	1187	0.33	ng	96
22) Hexachlorobenzene	16.274	284	1476	0.33	ng	# 85
23) Atrazine	16.445	200	1081	0.31	ng	94
24) Pentachlorophenol	16.615	266	640	0.31	ng	98
25) Phenanthrene	17.005	178	5966	0.34	ng	99
26) Anthracene	17.090	178	5372	0.33	ng	98
28) Fluoranthene	19.032	202	6847	0.33	ng	99
30) Pyrene	19.395	202	6952	0.36	ng	99
32) Benzo(a)anthracene	21.152	228	6068	0.34	ng	99
33) Chrysene	21.205	228	6778	0.35	ng	99
34) Bis(2-ethylhexyl)phtha...	21.089	149	4278	0.37	ng	91
35) Indeno(1,2,3-cd)pyrene	25.489	276	9401	0.33	ng	98
37) Benzo(b)fluoranthene	22.690	252	7254	0.34	ng	# 84
38) Benzo(k)fluoranthene	22.731	252	7418	0.34	ng	# 91
39) Benzo(a)pyrene	23.241	252	7041	0.35	ng	# 80
40) Dibenzo(a,h)anthracene	25.500	278	7639	0.34	ng	96
41) Benzo(g,h,i)perylene	26.147	276	7884	0.33	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103020\
Data File : BN012484.D
Acq On : 31 Oct 2020 02:15
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Oct 31 05:39:00 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN103020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 30 14:49:10 2020
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.570 min Scan# 570
 Delta R.T. -0.000 min
 Lab File: BN012484.D
 Acq: 31 Oct 2020 02:15




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:152 Resp: 1206
 Ion Ratio Lower Upper
 152 100
 150 146.5 116.6 174.8
 115 77.5 52.9 79.3



#2
 1,4-Dioxane
 Concen: 0.37 ng
 RT: 3.165 min Scan# 23
 Delta R.T. -0.000 min
 Lab File: BN012484.D
 Acq: 31 Oct 2020 02:15

Tgt Ion: 88 Resp: 740
 Ion Ratio Lower Upper
 88 100
 58 75.0 59.4 89.2
 43 46.5 32.3 48.5









#3
n-Nitrosodimethylamine
Concen: 0.31 ng
RT: 3.479 min Scan# 62
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 1536
Ion Ratio Lower Upper
42 100
74 59.5 64.4 96.6#
44 6.6 2.8 4.2#



#4
2-Fluorophenol
Concen: 0.37 ng
RT: 5.186 min Scan# 274
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion: 112 Resp: 1562
Ion Ratio Lower Upper
112 100
64 69.3 49.5 74.3
63 49.9 32.0 48.0#









#5
Phenol-d6
Concen: 0.38 ng
RT: 6.757 min Scan# 469
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	99	Resp:	1971
Ion	Ratio	Lower	Upper
99	100		
42	38.6	22.4	33.6#
71	57.7	37.8	56.6#



#6
bis(2-Chloroethyl)ether
Concen: 0.36 ng
RT: 7.014 min Scan# 501
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:	93	Resp:	1218
Ion	Ratio	Lower	Upper
93	100		
63	95.0	63.1	94.7#
95	33.7	26.1	39.1









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.339 min Scan# 822
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	5256
Ion	Ratio	Lower	Upper
136	100		
137	12.3	10.6	15.8
54	9.3	8.6	12.8
68	4.8	5.7	8.5



#8
Nitrobenzene-d5
Concen: 0.34 ng
RT: 8.716 min Scan# 694
Delta R.T. -0.013 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:	82	Resp:	2046
Ion	Ratio	Lower	Upper
82	100		
128	33.9	30.6	46.0
54	70.0	48.3	72.5









#9
Naphthalene
Concen: 0.36 ng
RT: 10.389 min Scan# 826
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	5274
Ion Ratio	Lower	Upper	
128	100		
129	12.9	9.8	14.8
127	14.1	12.2	18.4



#10
Hexachlorobutadiene
Concen: 0.35 ng
RT: 10.668 min Scan# 848
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:	225	Resp:	1082
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	62.0	51.0	76.6









#11
2-Methylnaphthalene-d10
Concen: 0.35 ng
RT: 11.935 min Scan# 1009
Delta R.T. -0.005 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 3033
Ion Ratio Lower Upper
152 100
151 19.8 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.36 ng
RT: 12.011 min Scan# 1026
Delta R.T. -0.005 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:142 Resp: 3440
Ion Ratio Lower Upper
142 100
141 89.4 69.4 104.2
115 57.7 35.7 53.5#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.204 min Scan# 1271
Delta R.T. -0.013 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 2900
Ion Ratio Lower Upper
164 100
162 108.2 80.6 120.8
160 51.3 39.5 59.3



#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.714 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:330 Resp: 659
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 48.1 34.4 51.6









#15
2-Fluorobiphenyl
Concen: 0.35 ng
RT: 12.822 min Scan# 1162
Delta R.T. -0.013 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	3861
Ion	Ratio	Lower	Upper
172	100		
171	37.9	28.2	42.4
170	26.9	20.0	30.0



#16
Acenaphthylene
Concen: 0.34 ng
RT: 13.925 min Scan# 1249
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:	152	Resp:	4953
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.8	23.8
153	13.6	10.6	16.0









#17
Acenaphthene
Concen: 0.35 ng
RT: 14.268 min Scan# 1276
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 3216
Ion Ratio Lower Upper
154 100
153 117.2 83.6 125.4
152 57.6 43.9 65.9



#18
Fluorene
Concen: 0.34 ng
RT: 15.270 min Scan# 1355
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:166 Resp: 3928
Ion Ratio Lower Upper
166 100
165 99.6 78.6 118.0
167 13.8 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.956 min Scan# 1491
Delta R.T. -0.012 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 5930
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.7 5.0 7.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.34 ng
RT: 15.372 min Scan# 1363
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:198 Resp: 384
Ion Ratio Lower Upper
198 100
51 116.6 43.5 65.3#
105 62.7 27.2 40.8#









#21
4-Bromophenyl-phenylether
Concen: 0.33 ng
RT: 16.165 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 1187
Ion Ratio Lower Upper
248 100
250 102.3 77.3 115.9
141 70.7 57.8 86.8



#22
Hexachlorobenzene
Concen: 0.33 ng
RT: 16.274 min Scan# 1435
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:284 Resp: 1476
Ion Ratio Lower Upper
284 100
142 55.0 32.0 48.0#
249 31.2 27.0 40.6









#23
Atrazine
Concen: 0.31 ng
RT: 16.445 min Scan# 1449
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 1081
Ion Ratio Lower Upper
200 100
173 25.6 22.8 34.2
215 51.6 38.1 57.1



#24
Pentachlorophenol
Concen: 0.31 ng
RT: 16.615 min Scan# 1463
Delta R.T. -0.012 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:266 Resp: 640
Ion Ratio Lower Upper
266 100
264 63.3 50.9 76.3
268 66.6 51.5 77.3









#25
Phenanthrene
Concen: 0.34 ng
RT: 17.005 min Scan# 1495
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 5966
Ion Ratio Lower Upper
178 100
176 19.5 14.8 22.2
179 15.4 12.4 18.6



#26
Anthracene
Concen: 0.33 ng
RT: 17.090 min Scan# 1502
Delta R.T. -0.012 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:178 Resp: 5372
Ion Ratio Lower Upper
178 100
176 19.0 14.8 22.2
179 14.5 12.5 18.7









#27
Fluoranthene-d10
Concen: 0.33 ng
RT: 19.003 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 5992
Ion Ratio Lower Upper
212 100
106 7.9 6.8 10.2
104 5.5 4.1 6.1



#28
Fluoranthene
Concen: 0.33 ng
RT: 19.032 min Scan# 1783
Delta R.T. -0.005 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:202 Resp: 6847
Ion Ratio Lower Upper
202 100
101 7.7 5.8 8.6
203 16.5 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.163 min Scan# 2088
Delta R.T. -0.011 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 5709
Ion Ratio Lower Upper
240 100
120 10.3 5.3 7.9#
236 27.5 21.8 32.8



#30
Pyrene
Concen: 0.36 ng
RT: 19.395 min Scan# 1856
Delta R.T. -0.005 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:202 Resp: 6952
Ion Ratio Lower Upper
202 100
200 21.3 16.6 24.8
203 17.9 13.8 20.8









#31
Terphenyl-d14
Concen: 0.35 ng
RT: 19.608 min Scan# 1899
Delta R.T. -0.010 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 4640
Ion Ratio Lower Upper
244 100
212 11.0 7.3 10.9#
122 12.3 7.3 10.9#



#32
Benzo(a)anthracene
Concen: 0.34 ng
RT: 21.152 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:228 Resp: 6068
Ion Ratio Lower Upper
228 100
226 27.4 22.3 33.5
229 20.6 15.7 23.5





#33

Chrysene

Concen: 0.35 ng

RT: 21.205 min Scan# 2092

Delta R.T. -0.000 min

Lab File: BN012484.D

Acq: 31 Oct 2020 02:15

Tgt Ion:228 Resp: 6778

Ion Ratio Lower Upper

228 100

226 29.9 24.2 36.2





229 20.9 15.7 23.5

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.37 ng

RT: 21.089 min Scan# 2081

Delta R.T. -0.011 min

Lab File: BN012484.D

Acq: 31 Oct 2020 02:15




Tgt Ion:149 Resp: 4278


Ion Ratio Lower Upper

149 100

167 23.6 23.2 34.8

279 4.7 3.9 5.9









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.33 ng
RT: 25.489 min Scan# 3106
Delta R.T. -0.010 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:276 Resp: 9401
Ion Ratio Lower Upper
276 100
138 15.6 13.3 19.9
277 24.1 20.2 30.4



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.333 min Scan# 2476
Delta R.T. -0.010 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:264 Resp: 6533
Ion Ratio Lower Upper
264 100
260 24.9 19.8 29.6
265 53.3 51.4 77.0









#37
Benzo(b)fluoranthene
Concen: 0.34 ng
RT: 22.690 min Scan# 2288
Delta R.T. -0.007 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 7254
Ion Ratio Lower Upper
252 100
253 30.2 20.1 30.1#
125 20.4 7.2 10.8#



#38
Benzo(k)fluoranthene
Concen: 0.34 ng
RT: 22.731 min Scan# 2300
Delta R.T. -0.010 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:252 Resp: 7418
Ion Ratio Lower Upper
252 100
253 29.6 19.8 29.8
125 10.8 7.1 10.7#









#39
Benzo(a)pyrene
Concen: 0.35 ng
RT: 23.241 min Scan# 2449
Delta R.T. -0.007 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252	Resp:	7041
Ion Ratio	Lower	Upper
252	100	
253	32.3	21.3 31.9#
125	26.1	8.3 12.5#



#40
Dibenzo(a,h)anthracene
Concen: 0.34 ng
RT: 25.500 min Scan# 3109
Delta R.T. -0.014 min
Lab File: BN012484.D
Acq: 31 Oct 2020 02:15

Tgt Ion:278	Resp:	7639
Ion Ratio	Lower	Upper
278	100	
139	13.6	9.3 13.9
279	29.1	21.6 32.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.33 ng
 RT: 26.147 min Scan# 3298
 Delta R.T. -0.010 min
 Lab File: BN012484.D
 Acq: 31 Oct 2020 02:15

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	276	Resp:	7884
Ion	Ratio	Lower	Upper
276	100		
277	26.3	19.8	29.8
138	16.3	11.7	17.5

