

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103020\
 Data File : BN012469.D
 Acq On : 30 Oct 2020 14:43
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN103020

Quant Time: Oct 30 15:30:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN103020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 14:49:10 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.672	0.587	12.6	120	0.00
3	n-Nitrosodimethylamine	1.653	1.429	13.6	101	0.00
4 S	2-Fluorophenol	1.406	1.267	9.9	111	0.00
5 S	Phenol-d6	1.730	1.562	9.7	112	0.00
6	bis(2-Chloroethyl)ether	1.130	0.975	13.7	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.461	0.383	16.9	106	-0.01
9	Naphthalene	1.129	1.010	10.5	111	0.00
10	Hexachlorobutadiene	0.235	0.207	11.9	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.656	0.591	9.9	113	0.00
12	2-Methylnaphthalene	0.732	0.653	10.8	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	-0.01
14 S	2,4,6-Tribromophenol	0.270	0.239	11.5	109	0.00
15 S	2-Fluorobiphenyl	1.503	1.355	9.8	110	0.00
16	Acenaphthylene	1.991	1.791	10.0	112	0.00
17	Acenaphthene	1.272	1.141	10.3	111	0.00
18	Fluorene	1.585	1.429	9.8	111	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4,6-Dinitro-2-methylphenol	0.083	0.069	16.9	116	0.01
21	4-Bromophenyl-phenylether	0.243	0.216	11.1	111	0.00
22	Hexachlorobenzene	0.305	0.269	11.8	109	0.00
23	Atrazine	0.233	0.200	14.2	109	0.00
24	Pentachlorophenol	0.138	0.117	15.2	109	0.00
25	Phenanthrene	1.192	1.053	11.7	111	0.00
26	Anthracene	1.107	0.931	15.9	108	0.00
27 SURR	Fluoranthene-d10	1.239	1.090	12.0	110	0.00
28	Fluoranthene	1.395	1.199	14.1	109	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
30	Pyrene	1.367	1.281	6.3	109	0.00
31 S	Terphenyl-d14	0.930	0.857	7.8	108	0.00
32	Benzo(a)anthracene	1.247	1.096	12.1	104	0.00
33	Chrysene	1.361	1.254	7.9	106	0.00
34	Bis(2-ethylhexyl)phthalate	0.815	0.736	9.7	104	0.00
35	Indeno(1,2,3-cd)pyrene	1.990	1.741	12.5	94	0.00
36 I	Perylene-d12	1.000	1.000	0.0	97	0.00
37	Benzo(b)fluoranthene	1.290	1.151	10.8	101	0.00
38	Benzo(k)fluoranthene	1.352	1.249	7.6	100	0.00
39 C	Benzo(a)pyrene	1.231	1.112	9.7	100	0.00
40	Dibenzo(a,h)anthracene	1.394	1.260	9.6	95	0.00
41	Benzo(g,h,i)perylene	1.455	1.306	10.2	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN103020