

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103020\
 Data File : BN012469.D
 Acq On : 30 Oct 2020 14:43
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN103020

Quant Time: Oct 30 15:30:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN103020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 14:49:10 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	0.00
2	1,4-Dioxane	0.400	0.349	12.8	120	0.00
3	n-Nitrosodimethylamine	0.400	0.346	13.5	101	0.00
4 S	2-Fluorophenol	0.400	0.360	10.0	111	0.00
5 S	Phenol-d6	0.400	0.361	9.8	112	0.00
6	bis(2-Chloroethyl)ether	0.400	0.345	13.8	109	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	113	0.00
8 S	Nitrobenzene-d5	0.400	0.332	17.0	106	-0.01
9	Naphthalene	0.400	0.358	10.5	111	0.00
10	Hexachlorobutadiene	0.400	0.353	11.8	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.360	10.0	113	0.00
12	2-Methylnaphthalene	0.400	0.357	10.8	110	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	111	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.354	11.5	109	0.00
15 S	2-Fluorobiphenyl	0.400	0.360	10.0	110	0.00
16	Acenaphthylene	0.400	0.360	10.0	112	0.00
17	Acenaphthene	0.400	0.359	10.3	111	0.00
18	Fluorene	0.400	0.360	10.0	111	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	112	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.353	11.8	116	0.01
21	4-Bromophenyl-phenylether	0.400	0.356	11.0	111	0.00
22	Hexachlorobenzene	0.400	0.353	11.8	109	0.00
23	Atrazine	0.400	0.343	14.2	109	0.00
24	Pentachlorophenol	0.400	0.338	15.5	109	0.00
25	Phenanthrene	0.400	0.353	11.8	111	0.00
26	Anthracene	0.400	0.336	16.0	108	0.00
27 SURR	Fluoranthene-d10	0.400	0.352	12.0	110	0.00
28	Fluoranthene	0.400	0.344	14.0	109	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	106	0.00
30	Pyrene	0.400	0.375	6.3	109	0.00
31 S	Terphenyl-d14	0.400	0.368	8.0	108	0.00
32	Benzo(a)anthracene	0.400	0.351	12.3	104	0.00
33	Chrysene	0.400	0.369	7.8	106	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.361	9.8	104	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.350	12.5	94	0.00
36 I	Perylene-d12	0.400	0.400	0.0	97	0.00
37	Benzo(b)fluoranthene	0.400	0.357	10.8	101	0.00
38	Benzo(k)fluoranthene	0.400	0.370	7.5	100	0.00
39 C	Benzo(a)pyrene	0.400	0.362	9.5	100	0.00
40	Dibenzo(a,h)anthracene	0.400	0.361	9.8	95	0.00
41	Benzo(g,h,i)perylene	0.400	0.359	10.3	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN103020