

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103020\
 Data File : BN012477.D
 Acq On : 30 Oct 2020 19:37
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 31 05:30:59 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN103020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 14:49:10 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	141	0.00
2	1,4-Dioxane	0.400	0.408	-2.0	178	0.00
3	n-Nitrosodimethylamine	0.400	0.329	17.8	122	0.00
4 S	2-Fluorophenol	0.400	0.365	8.8	142	0.00
5 S	Phenol-d6	0.400	0.366	8.5	144	0.00
6	bis(2-Chloroethyl)ether	0.400	0.360	10.0	144	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	144	0.00
8 S	Nitrobenzene-d5	0.400	0.346	13.5	141	-0.01
9	Naphthalene	0.400	0.361	9.8	144	0.00
10	Hexachlorobutadiene	0.400	0.368	8.0	140	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.354	11.5	142	0.00
12	2-Methylnaphthalene	0.400	0.368	8.0	145	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	146	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.343	14.2	140	0.00
15 S	2-Fluorobiphenyl	0.400	0.358	10.5	144	0.00
16	Acenaphthylene	0.400	0.360	10.0	147	0.00
17	Acenaphthene	0.400	0.364	9.0	148	0.00
18	Fluorene	0.400	0.352	12.0	142	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	144	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.343	14.2	143	0.00
21	4-Bromophenyl-phenylether	0.400	0.343	14.2	138	0.00
22	Hexachlorobenzene	0.400	0.352	12.0	140	0.00
23	Atrazine	0.400	0.341	14.8	139	0.00
24	Pentachlorophenol	0.400	0.332	17.0	137	0.00
25	Phenanthrene	0.400	0.356	11.0	144	0.00
26	Anthracene	0.400	0.342	14.5	142	0.00
27 SURR	Fluoranthene-d10	0.400	0.343	14.2	137	0.00
28	Fluoranthene	0.400	0.341	14.8	139	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	133	0.00
30	Pyrene	0.400	0.380	5.0	138	0.00
31 S	Terphenyl-d14	0.400	0.371	7.3	136	0.00
32	Benzo(a)anthracene	0.400	0.366	8.5	136	0.00
33	Chrysene	0.400	0.358	10.5	129	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.366	8.5	132	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.341	14.8	115	0.00
36 I	Perylene-d12	0.400	0.400	0.0	121	0.00
37	Benzo(b)fluoranthene	0.400	0.356	11.0	125	0.00
38	Benzo(k)fluoranthene	0.400	0.355	11.3	119	0.00
39 C	Benzo(a)pyrene	0.400	0.363	9.3	124	0.00
40	Dibenzo(a,h)anthracene	0.400	0.349	12.8	114	0.00
41	Benzo(g,h,i)perylene	0.400	0.353	11.8	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4