

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103020\
 Data File : BN012479.D
 Acq On : 30 Oct 2020 23:14
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 02 08:44:55 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN103020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 14:49:10 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	0.00
2	1,4-Dioxane	0.400	0.314	21.5	110	0.00
3	n-Nitrosodimethylamine	0.400	0.364	9.0	109	0.00
4 S	2-Fluorophenol	0.400	0.357	10.8	112	0.00
5 S	Phenol-d6	0.400	0.365	8.8	116	0.00
6	bis(2-Chloroethyl)ether	0.400	0.328	18.0	106	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	115	0.00
8 S	Nitrobenzene-d5	0.400	0.338	15.5	110	-0.01
9	Naphthalene	0.400	0.359	10.3	114	0.00
10	Hexachlorobutadiene	0.400	0.369	7.8	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.359	10.3	115	0.00
12	2-Methylnaphthalene	0.400	0.349	12.8	110	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.337	15.8	109	0.00
15 S	2-Fluorobiphenyl	0.400	0.360	10.0	115	0.00
16	Acenaphthylene	0.400	0.352	12.0	114	0.00
17	Acenaphthene	0.400	0.360	10.0	116	0.00
18	Fluorene	0.400	0.356	11.0	114	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	115	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.328	18.0	108	0.01
21	4-Bromophenyl-phenylether	0.400	0.344	14.0	110	0.00
22	Hexachlorobenzene	0.400	0.360	10.0	114	0.00
23	Atrazine	0.400	0.337	15.8	109	0.00
24	Pentachlorophenol	0.400	0.338	15.5	112	0.00
25	Phenanthrene	0.400	0.353	11.8	114	0.00
26	Anthracene	0.400	0.341	14.8	113	0.00
27 SURR	Fluoranthene-d10	0.400	0.348	13.0	111	0.00
28	Fluoranthene	0.400	0.344	14.0	112	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	111	-0.01
30	Pyrene	0.400	0.368	8.0	112	0.00
31 S	Terphenyl-d14	0.400	0.363	9.3	111	0.00
32	Benzo(a)anthracene	0.400	0.363	9.3	113	0.00
33	Chrysene	0.400	0.347	13.3	105	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.358	10.5	108	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.377	5.8	106	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	108	0.00
37	Benzo(b)fluoranthene	0.400	0.342	14.5	107	0.00
38	Benzo(k)fluoranthene	0.400	0.363	9.3	108	0.00
39 C	Benzo(a)pyrene	0.400	0.356	11.0	108	0.00
40	Dibenzo(a,h)anthracene	0.400	0.363	9.3	105	0.00
41	Benzo(g,h,i)perylene	0.400	0.360	10.0	104	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4