

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103020\
 Data File : BN012484.D
 Acq On : 31 Oct 2020 02:15
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 31 04:39:40 2020
 Quant Title :
 QLast Update : Fri Oct 30 14:49:10 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	177#	0.00
2	1,4-Dioxane	0.672	0.614	8.6	199#	0.00
3	n-Nitrosodimethylamine	1.653	1.274	22.9	143	0.00
4 S	2-Fluorophenol	1.406	1.295	7.9	179#	0.00
5 S	Phenol-d6	1.730	1.634	5.5	185#	0.00
6	bis(2-Chloroethyl)ether	1.130	1.010	10.6	179#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	183#	0.00
8 S	Nitrobenzene-d5	0.461	0.389	15.6	175#	-0.01
9	Naphthalene	1.129	1.003	11.2	180#	0.00
10	Hexachlorobutadiene	0.235	0.206	12.3	170#	0.00
11 SURR	2-Methylnaphthalene-d10	0.656	0.577	12.0	180#	0.00
12	2-Methylnaphthalene	0.732	0.654	10.7	179#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	185#	-0.01
14 S	2,4,6-Tribromophenol	0.270	0.227	15.9	173#	0.00
15 S	2-Fluorobiphenyl	1.503	1.331	11.4	180#	-0.01
16	Acenaphthylene	1.991	1.708	14.2	177#	0.00
17	Acenaphthene	1.272	1.109	12.8	180#	0.00
18	Fluorene	1.585	1.354	14.6	174#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	187#	-0.01
20	4,6-Dinitro-2-methylphenol	0.083	0.065	21.7	183#	0.00
21	4-Bromophenyl-phenylether	0.243	0.200	17.7	172#	0.00
22	Hexachlorobenzene	0.305	0.249	18.4	168#	0.00
23	Atrazine	0.233	0.182	21.9	166#	0.00
24	Pentachlorophenol	0.138	0.108	21.7	168#	-0.01
25	Phenanthrene	1.192	1.006	15.6	178#	0.00
26	Anthracene	1.107	0.906	18.2	176#	-0.01
27 SURR	Fluoranthene-d10	1.239	1.010	18.5	170#	0.00
28	Fluoranthene	1.395	1.155	17.2	176#	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	177#	-0.01
30	Pyrene	1.367	1.218	10.9	173#	0.00
31 S	Terphenyl-d14	0.930	0.813	12.6	171#	0.00
32	Benzo(a)anthracene	1.247	1.063	14.8	169#	0.00
33	Chrysene	1.361	1.187	12.8	168#	0.00
34	Bis(2-ethylhexyl)phthalate	0.815	0.749	8.1	177#	-0.01
35	Indeno(1,2,3-cd)pyrene	1.990	1.647	17.2	148	-0.01
36 I	Perylene-d12	1.000	1.000	0.0	165#	-0.01
37	Benzo(b)fluoranthene	1.290	1.110	14.0	164#	0.00
38	Benzo(k)fluoranthene	1.352	1.135	16.1	154#	-0.01
39 C	Benzo(a)pyrene	1.231	1.078	12.4	163#	0.00
40	Dibenzo(a,h)anthracene	1.394	1.169	16.1	149	-0.01
41	Benzo(g,h,i)perylene	1.455	1.207	17.0	147	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0