

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN103020\  
 Data File : BN012484.D  
 Acq On : 31 Oct 2020 02:15  
 Operator : CG/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Oct 31 04:39:40 2020  
 Quant Title :  
 QLast Update : Fri Oct 30 14:49:10 2020  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	177	0.00
2	1,4-Dioxane	0.400	0.365	8.8	199	0.00
3	n-Nitrosodimethylamine	0.400	0.308	23.0	143	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	179	0.00
5 S	Phenol-d6	0.400	0.378	5.5	185	0.00
6	bis(2-Chloroethyl)ether	0.400	0.357	10.8	179	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	183	0.00
8 S	Nitrobenzene-d5	0.400	0.338	15.5	175	-0.01
9	Naphthalene	0.400	0.355	11.3	180	0.00
10	Hexachlorobutadiene	0.400	0.351	12.3	170	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.352	12.0	180	0.00
12	2-Methylnaphthalene	0.400	0.358	10.5	179	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	185	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.337	15.8	173	0.00
15 S	2-Fluorobiphenyl	0.400	0.354	11.5	180	-0.01
16	Acenaphthylene	0.400	0.343	14.2	177	0.00
17	Acenaphthene	0.400	0.349	12.8	180	0.00
18	Fluorene	0.400	0.342	14.5	174	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	187	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.338	15.5	183	0.00
21	4-Bromophenyl-phenylether	0.400	0.329	17.8	172	0.00
22	Hexachlorobenzene	0.400	0.326	18.5	168	0.00
23	Atrazine	0.400	0.313	21.8	166	0.00
24	Pentachlorophenol	0.400	0.313	21.8	168	-0.01
25	Phenanthrene	0.400	0.338	15.5	178	0.00
26	Anthracene	0.400	0.327	18.3	176	-0.01
27 SURR	Fluoranthene-d10	0.400	0.326	18.5	170	0.00
28	Fluoranthene	0.400	0.331	17.3	176	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	177	-0.01
30	Pyrene	0.400	0.356	11.0	173	0.00
31 S	Terphenyl-d14	0.400	0.349	12.8	171	0.00
32	Benzo(a)anthracene	0.400	0.341	14.8	169	0.00
33	Chrysene	0.400	0.349	12.8	168	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.368	8.0	177	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.331	17.3	148	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	165	-0.01
37	Benzo(b)fluoranthene	0.400	0.344	14.0	164	0.00
38	Benzo(k)fluoranthene	0.400	0.336	16.0	154	-0.01
39 C	Benzo(a)pyrene	0.400	0.350	12.5	163	0.00
40	Dibenzo(a,h)anthracene	0.400	0.335	16.3	149	-0.01
41	Benzo(g,h,i)perylene	0.400	0.332	17.0	147	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0