

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103020\
 Data File : BN012467.D
 Acq On : 30 Oct 2020 13:22
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Manual Integrations
 APPROVED

mohammad
 11/2/2020 12:37:17 PM

Quant Time: Oct 30 14:10:14 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN103020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 12:11:25 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.570	152	696	0.40	ng	# 0.00
7) Naphthalene-d8	10.339	136	2913	0.40	ng	# 0.00
13) Acenaphthene-d10	14.205	164	1628	0.40	ng	-0.01
19) Phenanthrene-d10	16.956	188	3305	0.40	ng	#-0.01
29) Chrysene-d12	21.174	240	3609	0.40	ng	# 0.00
36) Perylene-d12	23.340	264	3976	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.186	112	7523	3.09	ng	0.00
5) Phenol-d6	6.749	99	9920	3.44	ng	0.00
8) Nitrobenzene-d5	8.717	82	11910	3.71	ng	-0.01
11) 2-Methylnaphthalene-d10	11.935	152	16128	3.44	ng	0.00
14) 2,4,6-Tribromophenol	15.714	330	3738	3.23	ng	0.00
15) 2-Fluorobiphenyl	12.822	172	20784	3.43	ng	-0.01
27) Fluoranthene-d10	19.003	212	34279	3.36	ng	0.00
31) Terphenyl-d14	19.613	244	27474	3.24	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.157	88	3309	2.92	ng	# 82
3) n-Nitrosodimethylamine	3.463	42	9131	3.32	ng	# 72
6) bis(2-Chloroethyl)ether	7.006	93	7062	3.60	ng	88
9) Naphthalene	10.390	128	26726	3.28	ng	96
10) Hexachlorobutadiene	10.668	225	5373	3.12	ng	# 100
12) 2-Methylnaphthalene	12.011	142	18145	3.43	ng	# 92
16) Acenaphthylene	13.926	152	27504	3.42	ng	100
17) Acenaphthene	14.268	154	17131	3.34	ng	92
18) Fluorene	15.270	166	21671	3.31	ng	98
20) 4,6-Dinitro-2-methylph...	15.359	198	2645	3.44	ng	93
21) 4-Bromophenyl-phenylether	16.165	248	6887	3.44	ng	# 89
22) Hexachlorobenzene	16.275	284	7971	3.28	ng	# 84
23) Atrazine	16.433	200	6437	3.36	ng	98
24) Pentachlorophenol	16.615	266	3909	3.47	ng	99
25) Phenanthrene	17.005	178	33364	3.46	ng	98
26) Anthracene	17.090	178	31897	3.57	ng	99
28) Fluoranthene	19.032	202	38804	3.38	ng	100
30) Pyrene	19.395	202	39688	3.25	ng	100
32) Benzo(a)anthracene	21.152	228	39055	3.56	ng	99
33) Chrysene	21.206	228	38256	3.11	ng	99
34) Bis(2-ethylhexyl)phtha...	21.099	149	23691	3.30	ng	91
35) Indeno(1,2,3-cd)pyrene	25.493	276	55685	3.15	ng	98
37) Benzo(b)fluoranthene	22.693	252	43600m	3.48	ng	
38) Benzo(k)fluoranthene	22.734	252	44218	3.24	ng	96
39) Benzo(a)pyrene	23.244	252	40783	3.35	ng	94
40) Dibenzo(a,h)anthracene	25.507	278	45894	3.34	ng	95
41) Benzo(g,h,i)perylene	26.150	276	45939	3.18	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103020\
Data File : BN012467.D
Acq On : 30 Oct 2020 13:22
Operator : CG/JU
Sample : SSTDIC3.2
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDIC3.2

Manual Integrations
APPROVED

mohammad
11/2/2020 12:37:17 PM


Quant Time: Oct 30 14:10:14 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN103020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 30 12:11:25 2020
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.570 min Scan# 570
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22




Instrument :
BNA_N
ClientSampled :
SSTDICC3.2




Tgt Ion:152 Resp: 696
Ion Ratio Lower Upper
152 100
150 153.1 116.6 174.8
115 81.1 52.9 79.3#

Manual Integrations
APPROVED



mohammad
11/2/2020 12:37:17 PM




#2
1,4-Dioxane
Concen: 2.92 ng
RT: 3.157 min Scan# 22
Delta R.T. -0.008 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22



Tgt Ion: 88 Resp: 3309
Ion Ratio Lower Upper
88 100
58 83.1 59.4 89.2
43 60.3 32.3 48.5#









#3
n-Nitrosodimethylamine
Concen: 3.32 ng
RT: 3.463 min Scan# 60
Delta R.T. -0.016 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion	Ratio	Lower	Upper
42	100		
74	54.7	64.4	96.6#
44	3.5	2.8	4.2

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#4
2-Fluorophenol
Concen: 3.09 ng
RT: 5.186 min Scan# 274
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion	Ratio	Lower	Upper
112	100		
64	72.3	49.5	74.3
63	53.4	32.0	48.0#









#5
Phenol-d6
Concen: 3.44 ng
RT: 6.749 min Scan# 468
Delta R.T. -0.008 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion: 99 Resp: 9920
Ion Ratio Lower Upper
99 100
42 47.0 22.4 33.6#
71 63.0 37.8 56.6#

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#6
bis(2-Chloroethyl)ether
Concen: 3.60 ng
RT: 7.006 min Scan# 500
Delta R.T. -0.008 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22


Tgt Ion: 93 Resp: 7062
Ion Ratio Lower Upper
93 100
63 93.3 63.1 94.7
95 33.0 26.1 39.1





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.339 min Scan# 822
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22



Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:	136	Resp:	2913
Ion	Ratio	Lower	Upper
136	100		
137	13.4	10.6	15.8
54	10.2	8.6	12.8
68	5.6	5.7	8.5#

Manual Integrations
APPROVED



mohammad
11/2/2020 12:37:17 PM




#8
Nitrobenzene-d5
Concen: 3.71 ng
RT: 8.717 min Scan# 694
Delta R.T. -0.013 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22



Tgt Ion:	82	Resp:	11910
Ion	Ratio	Lower	Upper
82	100		
128	33.0	30.6	46.0
54	64.8	48.3	72.5









#9
Naphthalene
Concen: 3.28 ng
RT: 10.390 min Scan# 826
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:	128	Resp:	26726
Ion Ratio	Lower	Upper	
128	100		
129	11.3	9.8	14.8
127	13.3	12.2	18.4

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#10
Hexachlorobutadiene
Concen: 3.12 ng
RT: 10.668 min Scan# 848
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:	225	Resp:	5373
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.4	51.0	76.6









#11
2-Methylnaphthalene-d10
Concen: 3.44 ng
RT: 11.935 min Scan# 1009
Delta R.T. -0.004 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:152 Resp: 16128
Ion Ratio Lower Upper
152 100
151 20.2 16.7 25.1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM




#12
2-Methylnaphthalene
Concen: 3.43 ng
RT: 12.011 min Scan# 1026
Delta R.T. -0.004 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:142 Resp: 18145
Ion Ratio Lower Upper
142 100
141 89.4 69.4 104.2
115 57.1 35.7 53.5#








#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.205 min Scan# 1271
Delta R.T. -0.013 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22




Tgt Ion:164 Resp: 1628

Ion	Ratio	Lower	Upper
164	100		
162	105.8	80.6	120.8
160	52.0	39.5	59.3





#14
2,4,6-Tribromophenol
Concen: 3.23 ng
RT: 15.714 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22



Tgt Ion:330 Resp: 3738


Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	52.0	34.4	51.6



Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED

mohammad
11/2/2020 12:37:17 PM







#15
2-Fluorobiphenyl
Concen: 3.43 ng
RT: 12.822 min Scan# 1162
Delta R.T. -0.013 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:	172	Resp:	20784
Ion Ratio	Lower	Upper	
172	100		
171	36.3	28.2	42.4
170	25.2	20.0	30.0

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM




#16
Acenaphthylene
Concen: 3.42 ng
RT: 13.926 min Scan# 1249
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:	152	Resp:	27504
Ion Ratio	Lower	Upper	
152	100		
151	19.9	15.8	23.8
153	13.0	10.6	16.0







#17
Acenaphthene
Concen: 3.34 ng
RT: 14.268 min Scan# 1276
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22




Tgt Ion:154 Resp: 17131
Ion Ratio Lower Upper
154 100
153 115.2 83.6 125.4
152 56.8 43.9 65.9




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED



mohammad
11/2/2020 12:37:17 PM




#18
Fluorene
Concen: 3.31 ng
RT: 15.270 min Scan# 1355
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22




Tgt Ion:166 Resp: 21671
Ion Ratio Lower Upper
166 100
165 100.2 78.6 118.0
167 13.3 10.7 16.1





#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.956 min Scan# 1491
Delta R.T. -0.012 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:188 Resp: 3305
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.9 5.0 7.4#

Manual Integrations
APPROVED



mohammad
11/2/2020 12:37:17 PM




#20
4,6-Dinitro-2-methylphenol
Concen: 3.44 ng
RT: 15.359 min Scan# 1362
Delta R.T. -0.013 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22




Tgt Ion:198 Resp: 2645
Ion Ratio Lower Upper
198 100
51 60.4 43.5 65.3
105 36.7 27.2 40.8








#21
4-Bromophenyl-phenylether
Concen: 3.44 ng
RT: 16.165 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2





Tgt Ion:248 Resp: 6887
Ion Ratio Lower Upper
248 100
250 102.5 77.3 115.9
141 57.8 57.8 86.8#




#22
Hexachlorobenzene
Concen: 3.28 ng
RT: 16.275 min Scan# 1435
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22



Tgt Ion:284 Resp: 7971
Ion Ratio Lower Upper
284 100
142 56.5 32.0 48.0#
249 31.8 27.0 40.6









#23
Atrazine
Concen: 3.36 ng
RT: 16.433 min Scan# 1448
Delta R.T. -0.012 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:	200	Resp:	6437
Ion Ratio	Lower	Upper	
200	100		
173	28.4	22.8	34.2
215	45.7	38.1	57.1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#24
Pentachlorophenol
Concen: 3.47 ng
RT: 16.615 min Scan# 1463
Delta R.T. -0.012 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:	266	Resp:	3909
Ion Ratio	Lower	Upper	
266	100		
264	62.2	50.9	76.3
268	64.2	51.5	77.3









#25
Phenanthrene
Concen: 3.46 ng
RT: 17.005 min Scan# 1495
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:178 Resp: 33364
Ion Ratio Lower Upper
178 100
176 19.5 14.8 22.2
179 15.2 12.4 18.6

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#26
Anthracene
Concen: 3.57 ng
RT: 17.090 min Scan# 1502
Delta R.T. -0.012 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:178 Resp: 31897
Ion Ratio Lower Upper
178 100
176 18.6 14.8 22.2
179 15.1 12.5 18.7









#27
Fluoranthene-d10
Concen: 3.36 ng
RT: 19.003 min Scan# 1777
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:212 Resp: 34279
Ion Ratio Lower Upper
212 100
106 7.8 6.8 10.2
104 4.9 4.1 6.1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#28
Fluoranthene
Concen: 3.38 ng
RT: 19.032 min Scan# 1783
Delta R.T. -0.005 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:202 Resp: 38804
Ion Ratio Lower Upper
202 100
101 7.2 5.8 8.6
203 17.3 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.174 min Scan# 2089
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:240 Resp: 3609
Ion Ratio Lower Upper
240 100
120 8.2 5.3 7.9#
236 27.9 21.8 32.8

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#30
Pyrene
Concen: 3.25 ng
RT: 19.395 min Scan# 1856
Delta R.T. -0.005 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22


Tgt Ion:202 Resp: 39688
Ion Ratio Lower Upper
202 100
200 20.9 16.6 24.8
203 17.5 13.8 20.8





#31
Terphenyl-d14
Concen: 3.24 ng
RT: 19.613 min Scan# 1900
Delta R.T. -0.005 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22



Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Tgt Ion:244 Resp: 27474
Ion Ratio Lower Upper
244 100
212 9.2 7.3 10.9
122 8.9 7.3 10.9

Manual Integrations
APPROVED



mohammad
11/2/2020 12:37:17 PM




#32
Benzo(a)anthracene
Concen: 3.56 ng
RT: 21.152 min Scan# 2087
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22



Tgt Ion:228 Resp: 39055
Ion Ratio Lower Upper
228 100
226 26.9 22.3 33.5
229 19.5 15.7 23.5









#33
Chrysene
Concen: 3.11 ng
RT: 21.206 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion: 228 Resp: 38256
Ion Ratio Lower Upper
228 100
226 29.6 24.2 36.2
229 19.6 15.7 23.5

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.30 ng
RT: 21.099 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion: 149 Resp: 23691
Ion Ratio Lower Upper
149 100
167 23.6 23.2 34.8
279 4.2 3.9 5.9









#35
Indeno(1,2,3-cd)pyrene
Concen: 3.15 ng
RT: 25.493 min Scan# 3107
Delta R.T. -0.007 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion: 276 Resp: 55685
Ion Ratio Lower Upper
276 100
138 15.4 13.3 19.9
277 24.8 20.2 30.4

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.340 min Scan# 2478
Delta R.T. -0.003 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion: 264 Resp: 3976
Ion Ratio Lower Upper
264 100
260 26.3 19.8 29.6
265 69.8 51.4 77.0









#37
Benzo(b)fluoranthene
Concen: 3.48 ng m
RT: 22.693 min Scan# 2289
Delta R.T. -0.003 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:252 Resp: 43600
Ion Ratio Lower Upper
252 100
253 22.7 20.1 30.1
125 7.1 7.2 10.8#

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#38
Benzo(k)fluoranthene
Concen: 3.24 ng
RT: 22.734 min Scan# 2301
Delta R.T. -0.007 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:252 Resp: 44218
Ion Ratio Lower Upper
252 100
253 22.6 19.8 29.8
125 7.3 7.1 10.7









#39
Benzo(a)pyrene
Concen: 3.35 ng
RT: 23.244 min Scan# 2450
Delta R.T. -0.003 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:	252	Resp:	40783
Ion Ratio	Lower	Upper	
252	100		
253	23.5	21.3	31.9
125	8.4	8.3	12.5

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2




Manual Integrations
APPROVED


mohammad
11/2/2020 12:37:17 PM



#40
Dibenzo(a,h)anthracene
Concen: 3.34 ng
RT: 25.507 min Scan# 3111
Delta R.T. -0.007 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22


Tgt Ion:	278	Resp:	45894
Ion Ratio	Lower	Upper	
278	100		
139	10.3	9.3	13.9
279	24.2	21.6	32.4





#41
Benzo(g,h,i)perylene
Concen: 3.18 ng
RT: 26.150 min Scan# 3299
Delta R.T. -0.007 min
Lab File: BN012467.D
Acq: 30 Oct 2020 13:22

Tgt Ion:	276	Resp:	45939
Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.8	29.8
138	13.5	11.7	17.5



Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Manual Integrations
APPROVED

mohammad
11/2/2020 12:37:17 PM