

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103024\
 Data File : BN034747.D
 Acq On : 30 Oct 2024 15:21
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN103024

Quant Time: Oct 30 15:57:34 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN103024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 30 13:28:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.604	152	7993	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	24332	0.400	ng	0.00
13) Acenaphthene-d10	14.240	164	13847	0.400	ng	0.00
19) Phenanthrene-d10	16.976	188	28842	0.400	ng	# 0.00
29) Chrysene-d12	21.176	240	20211	0.400	ng	0.00
35) Perylene-d12	23.362	264	19052	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.220	112	9901	0.409	ng	0.00
5) Phenol-d6	6.780	99	12725	0.404	ng	0.00
8) Nitrobenzene-d5	8.738	82	6502	0.340	ng	0.00
11) 2-Methylnaphthalene-d10	11.969	152	12746	0.362	ng	0.00
14) 2,4,6-Tribromophenol	15.723	330	2296	0.336	ng	0.00
15) 2-Fluorobiphenyl	12.860	172	18636	0.342	ng	0.00
27) Fluoranthene-d10	19.017	212	25699	0.380	ng	0.00
31) Terphenyl-d14	19.626	244	14610	0.337	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.205	88	3241	0.370	ng	97
3) n-Nitrosodimethylamine	3.501	42	3778	0.382	ng	94
6) bis(2-Chloroethyl)ether	7.040	93	8441	0.368	ng	99
9) Naphthalene	10.415	128	23988	0.356	ng	99
10) Hexachlorobutadiene	10.714	225	3960	0.357	ng	# 100
12) 2-Methylnaphthalene	12.041	142	15573	0.354	ng	99
16) Acenaphthylene	13.951	152	22661	0.326	ng	100
17) Acenaphthene	14.293	154	15571	0.336	ng	99
18) Fluorene	15.287	166	19700	0.341	ng	100
20) 4,6-Dinitro-2-methylph...	15.362	198	1194	0.396	ng	88
21) 4-Bromophenyl-phenylether	16.182	248	5801	0.348	ng	# 81
22) Hexachlorobenzene	16.294	284	6660	0.356	ng	99
23) Atrazine	16.467	200	4984	0.347	ng	# 91
24) Pentachlorophenol	16.629	266	2972	0.369	ng	97
25) Phenanthrene	17.026	178	31124	0.367	ng	100
26) Anthracene	17.113	178	28137	0.354	ng	100
28) Fluoranthene	19.045	202	34832	0.373	ng	100
30) Pyrene	19.407	202	36026	0.341	ng	100
32) Benzo(a)anthracene	21.158	228	28785	0.358	ng	99
33) Chrysene	21.212	228	28844	0.366	ng	100
34) Bis(2-ethylhexyl)phtha...	21.122	149	26293	0.391	ng	100
36) Indeno(1,2,3-cd)pyrene	25.537	276	24079	0.344	ng	100
37) Benzo(b)fluoranthene	22.710	252	27161	0.354	ng	98
38) Benzo(k)fluoranthene	22.754	252	27081	0.362	ng	98
39) Benzo(a)pyrene	23.262	252	22298	0.352	ng	97
40) Dibenzo(a,h)anthracene	25.558	278	18731	0.343	ng	98
41) Benzo(g,h,i)perylene	26.195	276	20248	0.343	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103024\
Data File : BN034747.D
Acq On : 30 Oct 2024 15:21
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Quant Time: Oct 30 15:57:34 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN103024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 30 13:28:24 2024
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.604 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN034747.D
 Acq: 30 Oct 2024 15:21

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN103024



Tgt Ion:152 Resp: 7993




Ion	Ratio	Lower	Upper
152	100		
150	154.4	124.9	187.3
115	61.0	50.2	75.2




#2
 1,4-Dioxane
 Concen: 0.370 ng
 RT: 3.205 min Scan# 45
 Delta R.T. -0.000 min
 Lab File: BN034747.D
 Acq: 30 Oct 2024 15:21

Tgt Ion: 88 Resp: 3241

Ion	Ratio	Lower	Upper
88	100		
43	31.3	26.7	40.1
58	76.7	59.3	88.9









#3
n-Nitrosodimethylamine
Concen: 0.382 ng
RT: 3.501 min Scan# 86
Delta R.T. -0.000 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21




Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Tgt Ion: 42 Resp: 3778
Ion Ratio Lower Upper
42 100
74 121.8 103.1 154.7
44 8.7 6.7 10.1



#4
2-Fluorophenol
Concen: 0.409 ng
RT: 5.220 min Scan# 324
Delta R.T. -0.000 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion: 112 Resp: 9901
Ion Ratio Lower Upper
112 100
64 56.0 44.9 67.3
63 29.7 23.5 35.3









#5
Phenol-d6
Concen: 0.404 ng
RT: 6.780 min Scan# 540
Delta R.T. -0.000 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21




Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Tgt Ion: 99 Resp: 12725
Ion Ratio Lower Upper
99 100
42 19.9 15.9 23.9
71 31.0 24.3 36.5



#6
bis(2-Chloroethyl)ether
Concen: 0.368 ng
RT: 7.040 min Scan# 576
Delta R.T. -0.000 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion: 93 Resp: 8441
Ion Ratio Lower Upper
93 100
63 75.9 61.2 91.8
95 32.0 25.8 38.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.372 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21




Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Tgt Ion:136	Resp:	24332
Ion Ratio	Lower	Upper
136	100	
137	11.2	9.2 13.8
54	7.5	7.0 10.4
68	6.1	5.6 8.4



#8
Nitrobenzene-d5
Concen: 0.340 ng
RT: 8.738 min Scan# 792
Delta R.T. -0.000 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion: 82	Resp:	6502
Ion Ratio	Lower	Upper
82	100	
128	41.4	33.7 50.5
54	56.2	44.4 66.6









#9
Naphthalene
Concen: 0.356 ng
RT: 10.415 min Scan# 949
Delta R.T. -0.011 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21




Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Tgt Ion:128 Resp: 23988
Ion Ratio Lower Upper
128 100
129 11.4 9.4 14.0
127 13.5 10.6 15.8



#10
Hexachlorobutadiene
Concen: 0.357 ng
RT: 10.714 min Scan# 977
Delta R.T. -0.011 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion:225 Resp: 3960
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.3 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 0.362 ng
RT: 11.969 min Scan# 1174
Delta R.T. -0.000 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21




Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Tgt Ion:152 Resp: 12746
Ion Ratio Lower Upper
152 100
151 20.8 16.6 24.8



#12
2-Methylnaphthalene
Concen: 0.354 ng
RT: 12.041 min Scan# 1193
Delta R.T. -0.004 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion:142 Resp: 15573
Ion Ratio Lower Upper
142 100
141 88.7 70.6 106.0
115 36.5 28.6 42.8









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.240 min Scan# 1495
Delta R.T. 0.007 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21




Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Tgt Ion:164 Resp: 13847
Ion Ratio Lower Upper
164 100
162 101.7 84.3 126.5
160 48.6 41.7 62.5



#14
2,4,6-Tribromophenol
Concen: 0.336 ng
RT: 15.723 min Scan# 1631
Delta R.T. -0.006 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion:330 Resp: 2296
Ion Ratio Lower Upper
330 100
332 96.6 77.8 116.6
141 51.8 41.0 61.4









#15
2-Fluorobiphenyl
Concen: 0.342 ng
RT: 12.860 min Scan# 1367
Delta R.T. -0.004 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21




Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Tgt Ion:172 Resp: 18636
Ion Ratio Lower Upper
172 100
171 33.9 27.2 40.8
170 23.1 18.0 27.0



#16
Acenaphthylene
Concen: 0.326 ng
RT: 13.951 min Scan# 1468
Delta R.T. -0.004 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion:152 Resp: 22661
Ion Ratio Lower Upper
152 100
151 19.3 15.4 23.2
153 12.9 10.2 15.4





#17

Acenaphthene

Concen: 0.336 ng

RT: 14.293 min Scan# 1500

Delta R.T. -0.004 min

Lab File: BN034747.D

Acq: 30 Oct 2024 15:21

Instrument :

BNA_N

ClientSampleId :

ICVBN103024





Tgt Ion:154 Resp: 15571

Ion Ratio Lower Upper

154 100

153 110.8 87.8 131.6

152 56.4 44.6 67.0



#18

Fluorene

Concen: 0.341 ng

RT: 15.287 min Scan# 1593

Delta R.T. -0.004 min

Lab File: BN034747.D

Acq: 30 Oct 2024 15:21




Tgt Ion:166 Resp: 19700


Ion Ratio Lower Upper

166 100

165 98.3 78.5 117.7

167 13.4 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.976 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21




Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Tgt Ion:188 Resp: 28842
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.9 9.8 14.6#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.396 ng
RT: 15.362 min Scan# 1600
Delta R.T. -0.004 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion:198 Resp: 1194
Ion Ratio Lower Upper
198 100
51 84.5 58.0 87.0
105 67.7 49.1 73.7









#21
4-Bromophenyl-phenylether
Concen: 0.348 ng
RT: 16.182 min Scan# 1668
Delta R.T. -0.006 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21




Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Tgt Ion:248 Resp: 5801
Ion Ratio Lower Upper
248 100
250 93.9 78.7 118.1
141 89.8 46.4 69.6#



#22
Hexachlorobenzene
Concen: 0.356 ng
RT: 16.294 min Scan# 1677
Delta R.T. 0.007 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion:284 Resp: 6660
Ion Ratio Lower Upper
284 100
142 45.6 35.4 53.0
249 32.0 25.7 38.5





#23

Atrazine

Concen: 0.347 ng

RT: 16.467 min Scan# 1

Delta R.T. 0.007 min

Lab File: BN034747.D

Acq: 30 Oct 2024 15:21

Instrument :

BNA_N

ClientSampleId :

ICVBN103024





Tgt Ion:200 Resp: 4984

Ion Ratio Lower Upper

200 100

173 23.0 25.8 38.6#

215 44.9 34.2 51.4



#24

Pentachlorophenol

Concen: 0.369 ng

RT: 16.629 min Scan# 1704

Delta R.T. -0.005 min

Lab File: BN034747.D

Acq: 30 Oct 2024 15:21




Tgt Ion:266 Resp: 2972


Ion Ratio Lower Upper

266 100

264 63.0 48.7 73.1

268 65.2 50.2 75.4









#25
 Phenanthrene
 Concen: 0.367 ng
 RT: 17.026 min Scan# 1736
 Delta R.T. 0.007 min
 Lab File: BN034747.D
 Acq: 30 Oct 2024 15:21




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN103024


Tgt Ion:178 Resp: 31124
 Ion Ratio Lower Upper
 178 100
 176 19.4 15.4 23.0
 179 15.2 12.2 18.4



#26
 Anthracene
 Concen: 0.354 ng
 RT: 17.113 min Scan# 1743
 Delta R.T. 0.007 min
 Lab File: BN034747.D
 Acq: 30 Oct 2024 15:21

Tgt Ion:178 Resp: 28137
 Ion Ratio Lower Upper
 178 100
 176 18.7 15.0 22.6
 179 15.1 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.380 ng
RT: 19.017 min Scan# 1965
Delta R.T. 0.002 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21




Instrument :
BNA_N
ClientSampleId :
ICVBN103024


Tgt Ion:212 Resp: 25699
Ion Ratio Lower Upper
212 100
106 20.0 16.1 24.1
104 11.4 9.2 13.8



#28
Fluoranthene
Concen: 0.373 ng
RT: 19.045 min Scan# 1970
Delta R.T. -0.003 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion:202 Resp: 34832
Ion Ratio Lower Upper
202 100
101 14.0 11.4 17.0
203 17.2 13.7 20.5





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.176 min Scan# 2411

Delta R.T. -0.003 min

Lab File: BN034747.D

Acq: 30 Oct 2024 15:21

Instrument :

BNA_N

ClientSampleId :

ICVBN103024





Tgt Ion:240 Resp: 20211

Ion Ratio Lower Upper

240 100

120 15.9 11.5 17.3

236 28.5 23.2 34.8



#30

Pyrene

Concen: 0.341 ng

RT: 19.407 min Scan# 2048

Delta R.T. -0.003 min

Lab File: BN034747.D

Acq: 30 Oct 2024 15:21




Tgt Ion:202 Resp: 36026


Ion Ratio Lower Upper

202 100

200 20.9 17.0 25.4

203 17.6 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.337 ng
 RT: 19.626 min Scan# 2097
 Delta R.T. -0.003 min
 Lab File: BN034747.D
 Acq: 30 Oct 2024 15:21




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN103024


Tgt Ion:244 Resp: 14610
 Ion Ratio Lower Upper
 244 100
 212 10.7 8.2 12.4
 122 26.2 19.8 29.6



#32
 Benzo(a)anthracene
 Concen: 0.358 ng
 RT: 21.158 min Scan# 2409
 Delta R.T. -0.003 min
 Lab File: BN034747.D
 Acq: 30 Oct 2024 15:21

Tgt Ion:228 Resp: 28785
 Ion Ratio Lower Upper
 228 100
 226 27.6 22.6 33.8
 229 19.6 15.7 23.5





#33

Chrysene

Concen: 0.366 ng

RT: 21.212 min Scan# 2417

Delta R.T. -0.003 min

Lab File: BN034747.D

Acq: 30 Oct 2024 15:21

Instrument :

BNA_N

ClientSampleId :

ICVBN103024





Tgt Ion:228 Resp: 28844

Ion Ratio Lower Upper

228 100

226 30.5 24.4 36.6

229 19.4 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.391 ng

RT: 21.122 min Scan# 2405

Delta R.T. -0.003 min

Lab File: BN034747.D

Acq: 30 Oct 2024 15:21




Tgt Ion:149 Resp: 26293


Ion Ratio Lower Upper

149 100

167 25.7 20.6 30.8

279 1.9 1.5 2.3









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.362 min Scan# 21
 Delta R.T. 0.001 min
 Lab File: BN034747.D
 Acq: 30 Oct 2024 15:21




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN103024


Tgt Ion:264 Resp: 19052
 Ion Ratio Lower Upper
 264 100
 260 27.0 21.5 32.3
 265 51.6 47.5 71.3



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.344 ng
 RT: 25.537 min Scan# 3618
 Delta R.T. 0.001 min
 Lab File: BN034747.D
 Acq: 30 Oct 2024 15:21


Tgt Ion:276 Resp: 24079
 Ion Ratio Lower Upper
 276 100
 138 39.6 31.5 47.3
 277 24.7 19.7 29.5








#37
Benzo(b)fluoranthene
Concen: 0.354 ng
RT: 22.710 min Scan# 2651
Delta R.T. 0.001 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Instrument :
BNA_N
ClientSampleId :
ICVBN103024







Tgt Ion:252 Resp: 27161
Ion Ratio Lower Upper
252 100
253 23.9 19.7 29.5
125 20.1 17.1 25.7



#38
Benzo(k)fluoranthene
Concen: 0.362 ng
RT: 22.754 min Scan# 2666
Delta R.T. 0.001 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Tgt Ion:252 Resp: 27081
Ion Ratio Lower Upper
252 100
253 23.6 19.8 29.6
125 21.9 18.2 27.2





#39

Benzo(a)pyrene

Concen: 0.352 ng

RT: 23.262 min Scan# 21

Delta R.T. -0.002 min

Lab File: BN034747.D


Acq: 30 Oct 2024 15:21

Instrument :

BNA_N

ClientSampleId :

ICVBN103024






Tgt Ion:252 Resp: 22298

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

252	100		
-----	-----	--	--

253	24.9	20.7	31.1
-----	------	------	------

125	23.3	20.3	30.5
-----	------	------	------



#40

Dibenzo(a,h)anthracene

Concen: 0.343 ng

RT: 25.558 min Scan# 3625

Delta R.T. -0.002 min

Lab File: BN034747.D

Acq: 30 Oct 2024 15:21




Tgt Ion:278 Resp: 18731


Ion	Ratio	Lower	Upper
-----	-------	-------	-------

278	100		
-----	-----	--	--

139	29.6	24.9	37.3
-----	------	------	------


279	25.2	21.0	31.4
-----	------	------	------





#41
Benzo(g,h,i)perylene
Concen: 0.343 ng
RT: 26.195 min Scan# 3843
Delta R.T. -0.002 min
Lab File: BN034747.D
Acq: 30 Oct 2024 15:21

Instrument :
BNA_N
ClientSampleId :
ICVBN103024



Tgt Ion:276 Resp: 20248
Ion Ratio Lower Upper
276 100
277 25.1 20.2 30.2
138 36.0 28.7 43.1

