

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN103122\  
 Data File : BN022410.D  
 Acq On : 31 Oct 2022 21:15  
 Operator : CG/JU  
 Sample : PB148647BL  
 Misc :  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 SBLK647

Quant Time: Nov 01 01:12:28 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN102022.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Nov 01 01:09:32 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.928  | 152  | 3976     | 0.400 | ng/u1 | 0.00     |
| 4) Naphthalene-d8           | 10.750 | 136  | 12454    | 0.400 | ng/u1 | 0.00     |
| 9) Acenaphthene-d10         | 14.600 | 164  | 6266     | 0.400 | ng/u1 | 0.00     |
| 13) Phenanthrene-d10        | 17.353 | 188  | 11182    | 0.400 | ng/u1 | 0.00     |
| 17) Chrysene-d12            | 21.559 | 240  | 7002     | 0.400 | ng/u1 | 0.00     |
| 23) Perylene-d12            | 24.002 | 264  | 6604     | 0.400 | ng/u1 | 0.00     |
| System Monitoring Compounds |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8           | 3.240  | 96   | 33819    | 6.545 | ng/u1 | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.356 | 152  | 5771     | 0.306 | ng/u1 | 0.00     |
| 18) Fluoranthene-d10        | 19.394 | 212  | 8168     | 0.347 | ng/u1 | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN103122\  
Data File : BN022410.D  
Acq On : 31 Oct 2022 21:15  
Operator : CG/JU  
Sample : PB148647BL  
Misc :  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
SBLK647

Quant Time: Nov 01 01:12:28 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN102022.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Nov 01 01:09:32 2022  
Response via : Initial Calibration

