

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103122\
 Data File : BN022422.D
 Acq On : 01 Nov 2022 04:44
 Operator : CG/JU
 Sample : N5277-11
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4X2

Quant Time: Nov 04 10:46:14 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 01 01:09:32 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.927	152	5062	0.400	ng/ul	0.00
4) Naphthalene-d8	10.750	136	16969	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.600	164	9619	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.353	188	17887	0.400	ng/ul	0.00
17) Chrysene-d12	21.550	240	10669	0.400	ng/ul	0.00
23) Perylene-d12	23.996	264	8154	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.236	96	24946	3.792	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.351	152	7486	0.292	ng/ul	0.00
18) Fluoranthene-d10	19.389	212	14934	0.417	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.274	88	1379	0.204	ng/ul#	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 04 10:46:14 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 01 01:09:32 2022
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.204 ng/ul
 RT: 3.274 min Scan# 61
 Delta R.T. -0.009 min
 Lab File: BN022422.D
 Acq: 01 Nov 2022 04:44

Instrument :
 BNA_N
 ClientSampleId :
 EW4X2

Tgt Ion: 88 Resp: 1379

Ion	Ratio	Lower	Upper
88	100		
43	64.6	27.4	41.2#
58	66.2	48.2	72.2

