

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103122\
 Data File : BN022440.D
 Acq On : 01 Nov 2022 16:38
 Operator : CG/JU
 Sample : N5277-17
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4X8

Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2022
 Supervised By :mohammad ahmed 11/04/2022

Quant Time: Nov 02 00:54:41 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 01 01:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.928	152	4707	0.400	ng/ul	0.00
4) Naphthalene-d8	10.750	136	16571m	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.600	164	9899m	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.353	188	18007	0.400	ng/ul	0.00
17) Chrysene-d12	21.556	240	9054	0.400	ng/ul	0.00
23) Perylene-d12	23.999	264	7262	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.236	96	23385	3.823	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.351	152	8412m	0.336	ng/ul	0.00
18) Fluoranthene-d10	19.393	212	13870	0.456	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.278	88	1496	0.238	ng/ul#	25

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 0.238 ng/ul
 RT: 3.278 min Scan# 70
 Delta R.T. -0.004 min
 Lab File: BN022440.D
 Acq: 01 Nov 2022 16:38

Instrument :
 BNA_N
 ClientSampleId :
 EW4X8



Tgt Ion: 88 Resp: 1496

Ion	Ratio	Lower	Upper
88	100		
43	133.6	27.4	41.2
58	74.7	48.2	72.2

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/02/2022
 Supervised By :mohammad ahmed 11/04/2022

