


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103122\
 Data File : BN022395.D
 Acq On : 31 Oct 2022 09:11
 Operator : CG/JU
 Sample : DFTPP292
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP292

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102022.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 01 01:09:32 2022



AutoFind: Scans 2515, 2516, 2517; Background Corrected with Scan 2508

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.9	167488	PASS
68	69	0.00	2	1.6	2941	PASS
69	198	0.00	100	42.3	182144	PASS
70	69	0.00	2	0.5	997	PASS
127	198	10	80	53.2	228736	PASS
197	198	0.00	2	0.6	2517	PASS
198	198	100	100	100.0	430101	PASS
199	198	5	9	6.8	29280	PASS
275	198	10	60	25.3	108872	PASS
365	198	1	100	3.9	16662	PASS
441	198	0.01	100	12.8	55149	PASS
442	198	50	100	76.2	327595	PASS
443	442	15	24	20.6	67448	PASS

m/z	Abundance
39.00	155.0
39.90	1037.0
40.90	204.0
43.00	172.0
43.90	560.0
55.00	180.0
73.00	493.0
83.80	176.0
96.00	185.0
146.90	197.0
190.90	168.0
207.00	1339.0
208.00	296.0
209.00	195.0
281.00	300.0

Instrument :

BNA_N

ClientSampleId :

DFTPP292

m/z	Abundance
37.10	657.0
38.00	1816.0
39.10	11089.0
40.00	1137.0
41.00	422.0
43.90	507.0
45.00	331.0
49.00	1273.0
50.00	42120.0
51.00	162368.0
52.00	8475.0
53.00	376.0
55.00	852.0
56.00	4595.0
57.00	10952.0
57.90	568.0
58.90	159.0
61.00	2066.0
62.00	2365.0
63.00	6866.0
64.00	990.0
65.00	3273.0
66.00	267.0
67.10	244.0
68.00	2713.0
69.00	175488.0
70.00	963.0
73.00	1833.0
74.00	18296.0
75.00	27792.0
76.00	9502.0
77.10	208704.0
78.00	15097.0
79.00	13022.0
80.00	9692.0
81.00	13670.0
82.00	3430.0
83.00	3315.0
84.00	520.0
85.00	2521.0
86.00	3983.0
87.00	1745.0
87.90	848.0
88.90	403.0
91.00	2806.0
92.00	3606.0
93.00	23864.0
94.00	1721.0
95.00	428.0
96.00	1117.0
97.10	365.0
98.00	18760.0
99.00	15702.0
100.00	1517.0
101.00	8020.0
102.00	380.0
103.00	2447.0
104.00	5028.0
105.00	4661.0
106.00	1323.0
107.00	59624.0
108.00	9691.0
109.00	1668.0
110.00	110952.0
111.00	17008.0
112.00	2255.0
113.00	652.0
113.80	172.0
115.00	190.0
116.00	3184.0
117.00	53192.0
118.00	4152.0
118.90	527.0
120.00	820.0
121.10	354.0
122.00	3877.0
123.00	6150.0
124.00	2852.0
125.00	2559.0
127.00	208640.0
128.00	15956.0
129.00	78768.0
130.00	6786.0
131.00	1469.0
132.00	692.0
132.90	411.0
133.90	2299.0
135.00	6377.0
136.00	2560.0
137.00	4038.0
137.90	754.0
138.90	493.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

140.00	1019.0
141.00	10350.0
142.00	3531.0
143.00	2263.0
144.00	702.0
145.00	601.0
146.00	1835.0
147.00	5184.0
148.00	13153.0
149.00	2804.0
150.00	681.0
151.10	1479.0
151.80	670.0
153.00	3218.0
154.00	2276.0
155.00	5255.0
156.00	8523.0
157.10	1609.0
157.90	1891.0
158.90	1478.0
160.00	2920.0
161.00	4575.0
162.00	1469.0
163.00	346.0
164.00	444.0
165.00	3639.0
166.00	2938.0
167.00	21496.0
168.00	13784.0
169.00	1971.0
170.00	683.0
170.90	916.0
172.00	1820.0
173.00	2266.0
174.00	4111.0
175.00	8165.0
176.10	1935.0
177.00	3588.0
178.00	1153.0
179.00	14405.0
180.00	10984.0
181.10	5012.0
182.00	710.0
183.00	280.0
184.00	1270.0
185.00	6941.0
186.00	53648.0
187.00	15567.0
188.00	1947.0
189.00	3623.0
190.00	698.0
191.00	1619.0
192.00	4463.0
193.00	5155.0
194.00	1163.0
195.00	641.0
196.00	11625.0
197.00	2241.0
198.00	391808.0
199.00	26464.0
200.00	2104.0
201.50	1514.0
203.00	2843.0
204.00	14509.0
205.00	23712.0
206.10	99848.0
207.00	14523.0
208.00	3426.0
209.00	1312.0
210.00	1486.0
211.00	4058.0
211.90	716.0
213.00	322.0
215.10	1123.0
216.10	2199.0
217.00	24696.0
218.00	3644.0
219.00	278.0
221.00	15031.0
223.00	5763.0
224.00	51104.0
225.00	13641.0
226.00	1712.0
227.00	22160.0
228.00	3417.0
229.00	4788.0
230.00	726.0
231.00	1826.0
232.00	392.0
233.00	375.0
234.00	1366.0
235.00	1722.0
236.00	1092.0
237.00	2562.0
238.00	424.0
238.90	923.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

240.00	751.0
241.00	1198.0
242.00	2870.0
243.00	3155.0
244.00	39800.0
245.10	5686.0
246.00	7922.0
247.00	1642.0
248.00	449.0
248.90	1355.0
249.90	304.0
250.90	340.0
252.00	380.0
253.00	930.0
255.00	210816.0
256.00	29216.0
257.00	2713.0
258.00	11802.0
259.00	1943.0
260.10	392.0
261.00	322.0
264.00	317.0
265.00	4781.0
266.00	715.0
270.00	303.0
270.90	629.0
271.90	555.0
273.00	6228.0
274.10	15017.0
275.00	87608.0
276.00	13219.0
277.00	7356.0
278.00	1382.0
279.00	293.0
281.00	305.0
282.10	236.0
283.00	984.0
284.00	545.0
285.10	1392.0
286.00	244.0
289.00	305.0
289.90	233.0
291.10	188.0
291.90	352.0
293.00	1844.0
294.00	566.0
295.00	552.0
296.00	25272.0
297.00	3797.0
298.00	362.0
301.00	369.0
302.00	518.0
303.00	3198.0
304.00	783.0
305.00	239.0
308.00	438.0
308.90	261.0
310.00	349.0
313.00	155.0
314.00	1242.0
315.00	2291.0
316.00	1725.0
317.00	452.0
321.00	712.0
322.00	488.0
323.00	8086.0
324.00	1391.0
325.00	162.0
325.90	216.0
327.00	1554.0
327.90	716.0
332.00	653.0
333.00	782.0
334.00	5020.0
335.00	1267.0
336.00	224.0
338.90	150.0
341.10	976.0
342.00	303.0
346.00	1798.0
347.10	406.0
350.90	160.0
352.00	2376.0
353.00	1793.0
354.10	2689.0
355.00	681.0
359.00	259.0
365.00	12274.0
366.00	1500.0
367.10	191.0
369.90	201.0
371.00	697.0
372.00	4064.0
373.00	1213.0
383.00	1307.0
384.00	346.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

390.00	558.0
390.90	455.0
392.10	321.0
401.00	415.0
402.00	1612.0
403.00	2561.0
404.00	986.0
405.00	266.0
421.00	2093.0
422.00	2046.0
423.10	15883.0
424.10	4126.0
425.10	737.0
441.00	35048.0
442.00	215808.0
443.10	43160.0
444.10	4207.0
444.90	273.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

m/z	Abundance
35.00	153.0
37.00	903.0
38.00	2301.0
39.10	12861.0
40.00	1233.0
41.00	380.0
43.00	182.0
44.00	538.0
44.90	325.0
49.00	1469.0
50.00	48456.0
51.00	188032.0
52.00	9768.0
53.00	321.0
55.00	978.0
56.00	5580.0
57.00	12477.0
58.00	696.0
59.00	158.0
61.00	2488.0
62.00	2631.0
63.00	7627.0
64.00	977.0
65.00	4046.0
66.00	243.0
66.90	255.0
68.00	3455.0
69.00	204032.0
70.00	1064.0
70.90	184.0
73.00	2056.0
74.00	21024.0
75.00	33272.0
76.00	11418.0
77.00	249088.0
78.00	17584.0
79.00	15370.0
80.00	11589.0
81.00	17008.0
82.00	4188.0
83.00	3863.0
84.00	516.0
85.00	3115.0
86.00	5211.0
87.00	2147.0
87.90	899.0
89.00	391.0
91.00	3598.0
92.00	4359.0
93.00	27920.0
94.00	1779.0
95.00	363.0
96.00	1238.0
97.00	445.0
98.00	21672.0
99.00	18088.0
100.00	1576.0
101.00	9251.0
102.00	530.0
103.00	3160.0
104.00	5490.0
105.00	5459.0
106.00	1699.0
107.00	73160.0
108.00	11031.0
109.00	1977.0
110.00	130944.0
111.00	20488.0
112.00	2507.0
113.00	778.0
114.00	181.0
115.00	312.0
116.00	4242.0
117.00	67384.0
118.00	4841.0
119.00	579.0
120.00	905.0
121.00	345.0
122.00	4867.0
123.00	7401.0
124.00	3288.0
125.00	3119.0
127.00	255936.0
128.00	19696.0
129.00	98568.0
130.00	8225.0
131.00	1664.0
132.00	1021.0
133.00	490.0
134.00	2719.0
135.00	7800.0
136.00	3039.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

137.00	4716.0
137.90	861.0
138.90	555.0
139.90	1283.0
141.00	12346.0
142.00	3946.0
143.00	2884.0
144.00	766.0
145.00	744.0
146.00	2440.0
147.00	6771.0
148.00	15898.0
149.00	3131.0
150.00	1057.0
151.10	1725.0
151.90	1032.0
153.00	4031.0
154.00	3023.0
155.00	7262.0
156.00	10567.0
157.10	1959.0
157.90	2254.0
159.00	1821.0
160.00	3950.0
161.00	5979.0
162.00	1616.0
163.00	497.0
164.00	664.0
165.00	4981.0
166.00	3685.0
167.00	28808.0
168.00	14909.0
169.00	2278.0
170.00	919.0
171.00	1046.0
172.00	2286.0
173.00	3078.0
174.00	5877.0
175.00	10173.0
176.10	2604.0
177.00	4973.0
178.00	1667.0
179.00	18152.0
180.00	13988.0
181.00	6501.0
182.00	922.0
183.00	368.0
184.00	1586.0
185.00	8916.0
186.00	68304.0
187.00	20720.0
188.00	2249.0
189.00	4510.0
189.90	684.0
190.90	2024.0
192.00	6117.0
193.10	6888.0
194.00	1464.0
195.10	748.0
196.00	13947.0
197.00	2892.0
198.00	478720.0
199.00	32568.0
200.00	2514.0
201.60	2220.0
203.00	3362.0
204.00	17768.0
205.00	29784.0
206.10	133248.0
207.10	17336.0
208.00	4661.0
209.00	1519.0
210.20	2019.0
211.00	5152.0
213.00	366.0
215.00	1558.0
216.10	2796.0
217.00	33496.0
218.00	4659.0
219.00	386.0
221.00	19008.0
223.00	7295.0
224.10	70880.0
225.00	19672.0
226.10	1914.0
227.00	31384.0
228.00	4531.0
229.00	6351.0
230.10	864.0
231.00	2621.0
232.00	440.0
233.00	533.0
234.00	1805.0
235.00	2351.0
236.00	1333.0
237.00	2879.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

238.00	343.0
239.00	1105.0
240.00	1043.0
241.00	1649.0
242.00	3753.0
243.10	3664.0
244.00	48520.0
245.10	7540.0
246.00	11449.0
247.10	2239.0
247.90	509.0
249.00	2073.0
249.90	344.0
251.00	409.0
252.00	582.0
253.00	1148.0
255.00	277888.0
256.00	42608.0
257.00	3392.0
258.00	16231.0
259.00	2391.0
260.00	485.0
260.90	503.0
264.00	425.0
265.00	7639.0
266.00	1138.0
267.00	232.0
270.10	250.0
271.00	752.0
272.00	766.0
273.00	8643.0
274.00	21984.0
275.00	122896.0
276.00	17672.0
277.00	11332.0
278.00	2065.0
279.00	338.0
281.00	321.0
281.90	302.0
283.00	1070.0
284.00	746.0
285.00	1769.0
286.10	356.0
288.00	165.0
289.00	467.0
290.00	379.0
291.00	255.0
292.00	579.0
293.00	2650.0
294.00	762.0
296.00	39928.0
297.10	5270.0
298.00	339.0
301.00	571.0
302.00	643.0
303.00	3572.0
304.00	1400.0
304.90	159.0
307.90	486.0
309.00	311.0
309.90	596.0
313.00	383.0
314.00	1787.0
315.00	4225.0
316.00	2302.0
317.10	450.0
321.10	1261.0
322.00	595.0
323.00	12870.0
324.00	2455.0
325.10	205.0
326.10	240.0
327.00	2587.0
328.00	1309.0
329.00	271.0
332.00	740.0
333.00	1258.0
334.00	8152.0
335.00	2090.0
336.00	237.0
339.00	240.0
339.90	178.0
341.00	1602.0
342.00	503.0
346.00	2827.0
347.00	592.0
351.00	179.0
352.10	3851.0
353.10	2605.0
354.10	4079.0
355.10	960.0
359.00	273.0
365.00	17784.0
366.00	2550.0
367.10	205.0
370.00	435.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

371.00	1077.0
372.00	7020.0
373.10	1759.0
374.10	185.0
376.80	190.0
383.00	1784.0
384.00	587.0
384.90	189.0
390.10	758.0
391.00	559.0
392.00	454.0
400.90	519.0
402.00	2587.0
403.00	4316.0
404.00	1259.0
405.00	328.0
415.00	199.0
421.00	3449.0
421.90	3348.0
423.10	23920.0
424.00	6818.0
425.10	957.0
440.00	277.0
441.10	60088.0
442.10	359744.0
443.00	67840.0
444.10	8345.0
445.00	522.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

m/z	Abundance
37.00	600.0
38.00	1709.0
39.00	10613.0
40.00	1214.0
41.00	416.0
43.00	180.0
44.00	540.0
45.00	311.0
49.00	1075.0
50.00	39912.0
51.00	152064.0
52.00	7805.0
53.10	306.0
55.00	729.0
56.00	4534.0
57.00	10747.0
58.00	449.0
58.90	155.0
61.00	1868.0
62.00	2126.0
63.00	6392.0
64.00	840.0
65.00	3056.0
65.90	267.0
67.00	290.0
68.00	2656.0
69.00	166912.0
70.00	965.0
73.00	1727.0
74.00	17504.0
75.00	27952.0
76.00	9863.0
77.00	202432.0
78.10	13729.0
79.00	12750.0
80.00	9571.0
81.00	13928.0
82.00	3448.0
83.00	2992.0
83.90	547.0
85.00	2731.0
86.00	3865.0
87.00	1808.0
88.00	752.0
88.90	318.0
91.00	3088.0
92.00	3413.0
93.00	23120.0
94.00	1464.0
95.00	362.0
96.00	1151.0
97.00	415.0
98.00	18392.0
99.00	14746.0
100.00	1116.0
101.00	7909.0
102.00	445.0
103.00	2693.0
104.00	4910.0
105.00	4764.0
106.00	1421.0
107.00	61312.0
108.00	9211.0
109.00	1678.0
110.00	107856.0
111.00	16880.0
112.00	2022.0
112.90	690.0
115.10	194.0
116.00	3300.0
117.00	55456.0
118.00	3859.0
119.00	540.0
120.00	719.0
121.00	273.0
122.00	3972.0
123.00	6246.0
124.00	2776.0
125.00	2715.0
126.10	942.0
127.00	221632.0
128.00	16408.0
129.00	82376.0
130.00	7073.0
131.00	1314.0
131.90	648.0
132.90	395.0
134.00	2293.0
135.00	6660.0
136.00	2878.0
137.00	3926.0
138.00	768.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

138.90	403.0
139.90	984.0
141.00	11073.0
142.00	3341.0
142.90	2419.0
144.00	579.0
145.00	666.0
146.00	1901.0
147.00	5818.0
148.00	14008.0
149.00	2984.0
150.00	914.0
151.10	1464.0
151.90	810.0
153.00	3403.0
154.00	2523.0
155.00	6442.0
156.00	9518.0
157.00	1739.0
158.00	2006.0
159.00	1573.0
160.00	3519.0
161.00	5539.0
162.00	1600.0
162.90	390.0
164.00	604.0
165.00	3892.0
166.00	3491.0
167.00	24224.0
168.00	13296.0
169.00	2302.0
170.00	588.0
171.00	877.0
171.90	2014.0
173.00	2647.0
174.00	4478.0
175.00	8685.0
176.00	2477.0
177.00	3787.0
178.00	1387.0
179.00	16125.0
180.00	11807.0
181.00	5356.0
182.10	829.0
182.90	440.0
184.00	1510.0
185.00	7972.0
186.10	62344.0
187.00	17768.0
188.00	1934.0
189.00	3669.0
190.00	652.0
191.00	1946.0
192.00	5499.0
193.00	6204.0
194.00	1377.0
195.00	750.0
196.00	13129.0
196.80	2418.0
198.00	419776.0
199.00	28808.0
200.00	2221.0
201.50	2106.0
203.00	3058.0
204.10	16400.0
205.00	27744.0
206.10	113888.0
207.10	14553.0
208.00	4087.0
209.00	1460.0
210.20	1889.0
211.00	4794.0
213.00	411.0
215.00	1346.0
216.00	2460.0
217.00	30648.0
218.00	4102.0
219.00	343.0
221.00	16992.0
223.00	6693.0
224.00	67768.0
225.00	16592.0
226.10	1810.0
227.00	26576.0
228.00	3972.0
229.00	5431.0
230.00	730.0
231.00	2270.0
232.00	393.0
232.90	400.0
234.00	1770.0
234.90	1694.0
236.00	1296.0
237.10	2623.0
238.00	365.0
238.90	913.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

240.00	918.0
241.00	1600.0
242.10	3331.0
243.10	3702.0
244.10	47456.0
245.10	6532.0
246.00	10388.0
247.00	2163.0
248.00	522.0
249.00	1686.0
250.10	382.0
251.00	471.0
252.00	498.0
253.00	1064.0
254.10	1717.0
255.00	255104.0
256.00	38880.0
257.00	3030.0
258.00	16856.0
259.00	2304.0
260.00	483.0
261.00	469.0
262.90	187.0
264.10	511.0
265.00	6441.0
266.00	780.0
267.00	183.0
270.00	309.0
270.90	495.0
272.10	767.0
273.00	7708.0
274.10	21800.0
275.10	116112.0
276.10	15986.0
277.00	9943.0
278.00	2058.0
279.00	354.0
281.10	328.0
282.00	303.0
283.00	1354.0
284.10	739.0
285.00	1780.0
286.20	398.0
288.90	380.0
290.00	350.0
291.00	216.0
292.00	624.0
293.00	2373.0
294.00	518.0
295.10	688.0
296.00	32272.0
297.00	4828.0
298.00	368.0
301.00	471.0
302.10	491.0
303.00	4530.0
304.00	1105.0
308.00	491.0
309.00	365.0
310.00	520.0
313.00	289.0
314.00	1665.0
315.00	4177.0
316.00	2247.0
316.90	356.0
320.00	157.0
321.00	1090.0
322.10	704.0
323.00	11555.0
324.00	2159.0
325.00	197.0
326.00	213.0
326.90	2328.0
328.00	849.0
328.90	232.0
332.00	878.0
333.00	1117.0
334.00	7973.0
335.00	2041.0
336.00	218.0
339.10	196.0
340.00	222.0
341.00	1552.0
342.00	472.0
345.90	2794.0
347.00	592.0
351.00	214.0
352.10	3699.0
353.00	2765.0
354.10	3866.0
355.10	858.0
358.90	216.0
363.90	242.0
365.00	19928.0
366.00	2759.0
366.90	164.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292

370.00	377.0
371.10	948.0
372.00	7191.0
373.00	1566.0
374.00	175.0
377.00	153.0
383.00	2162.0
384.00	687.0
385.00	243.0
390.00	1081.0
391.00	796.0
392.00	364.0
401.00	373.0
402.00	2977.0
403.00	4461.0
404.00	1740.0
405.00	235.0
415.00	264.0
421.10	4264.0
422.10	3975.0
423.10	22200.0
424.00	7426.0
425.00	693.0
441.00	70312.0
442.10	407232.0
443.00	91344.0
444.10	6549.0
445.00	486.0

Instrument :
BNA_N
ClientSampleId :
DFTPP292