

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103124\
 Data File : BN034793.D
 Acq On : 01 Nov 2024 08:08
 Operator : RC/JU
 Sample : P4529-15
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F7H36

Quant Time: Nov 03 23:25:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 31 11:13:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.600	152	8035	0.400	ng/ul	0.00
4) Naphthalene-d8	10.366	136	25648	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.231	164	14563	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.979	188	29963	0.400	ng/ul	0.00
17) Chrysene-d12	21.171	240	21590	0.400	ng/ul	0.00
23) Perylene-d12	23.355	264	20702	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.170	96	25923	3.117	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.961	152	6630	0.206	ng/ul	0.00
18) Fluoranthene-d10	19.009	212	14734	0.210	ng/ul	0.00
Target Compounds						
					Qvalue	
20) Pyrene	19.376	202	2194	0.020	ng/ul#	1
26) Benzo(a)pyrene	23.214	252	3655	0.050	ng/ul#	42


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103124\
Data File : BN034793.D
Acq On : 01 Nov 2024 08:08
Operator : RC/JU
Sample : P4529-15
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F7H36

Quant Time: Nov 03 23:25:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 31 11:13:08 2024
Response via : Initial Calibration









#20
Pyrene
Concen: 0.020 ng/ul
RT: 19.376 min Scan# 3647
Delta R.T. -0.033 min
Lab File: BN034793.D
Acq: 01 Nov 2024 08:08

Instrument :
BNA_N
ClientSampleId :
F7H36

Tgt Ion:202 Resp: 2194
Ion Ratio Lower Upper
202 100
101 186.5 12.4 18.6#
100 1318.3 10.0 15.0#



#26
Benzo(a)pyrene
Concen: 0.050 ng/ul
RT: 23.214 min Scan# 4817
Delta R.T. -0.050 min
Lab File: BN034793.D
Acq: 01 Nov 2024 08:08

Tgt Ion:252 Resp: 3655
Ion Ratio Lower Upper
252 100
253 42.7 20.0 30.0#
125 65.3 19.8 29.6#

