

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103124\
 Data File : BN034801.D
 Acq On : 01 Nov 2024 12:57
 Operator : RC/JU
 Sample : PB164465BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK465

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/04/2024
 Supervised By :mohammad ahmed 11/05/2024

Quant Time: Nov 03 23:26:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 31 11:13:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.600	152	7818	0.400	ng/u1	0.00
4) Naphthalene-d8	10.366	136	24510	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.231	164	13392	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.975	188	26954m	0.400	ng/u1	0.00
17) Chrysene-d12	21.169	240	17860	0.400	ng/u1	0.00
23) Perylene-d12	23.352	264	14737m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.170	96	41938	5.182	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.961	152	9164	0.298	ng/u1	0.00
18) Fluoranthene-d10	19.009	212	17624	0.303	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103124\
Data File : BN034801.D
Acq On : 01 Nov 2024 12:57
Operator : RC/JU
Sample : PB164465BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK465

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/04/2024
Supervised By :mohammad ahmed 11/05/2024

Quant Time: Nov 03 23:26:36 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 31 11:13:08 2024
Response via : Initial Calibration

