

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103124\
 Data File : BN034830.D
 Acq On : 02 Nov 2024 07:02
 Operator : RC/JU
 Sample : PB164449BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS449

Quant Time: Nov 03 23:30:04 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 31 11:13:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.596	152	7021	0.400	ng/ul	0.00
4) Naphthalene-d8	10.361	136	20707	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.226	164	11324	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.975	188	23797	0.400	ng/ul	0.00
17) Chrysene-d12	21.166	240	15837	0.400	ng/ul	0.00
23) Perylene-d12	23.346	264	12352	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.170	96	5166	0.711	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.956	152	10521	0.405	ng/ul	-0.01
18) Fluoranthene-d10	19.004	212	20590	0.399	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.199	88	13371	1.676	ng/ul#	84
5) Naphthalene	10.410	128	20920	0.381	ng/ul	100
7) 2-Methylnaphthalene	12.033	142	13559	0.380	ng/ul	100
8) 1-Methylnaphthalene	12.253	142	13783	0.380	ng/ul	100
10) Acenaphthylene	13.944	152	19337	0.375	ng/ul	99
11) Acenaphthene	14.291	153	15131	0.384	ng/ul	100
12) Fluorene	15.281	166	16874	0.384	ng/ul	99
14) Pentachlorophenol	16.629	266	3261	0.694	ng/ul	99
15) Phenanthrene	17.013	178	27832	0.387	ng/ul	100
16) Anthracene	17.106	178	25045	0.376	ng/ul	100
19) Fluoranthene	19.037	202	29821	0.388	ng/ul	99
20) Pyrene	19.400	202	30896	0.386	ng/ul	99
21) Benzo(a)anthracene	21.148	228	24282	0.383	ng/ul	99
22) Chrysene	21.201	228	25335	0.399	ng/ul	100
24) Benzo(b)fluoranthene	22.694	252	21362	0.425	ng/ul	98
25) Benzo(k)fluoranthene	22.738	252	20987	0.414	ng/ul	99
26) Benzo(a)pyrene	23.249	252	17251	0.399	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	25.513	276	19499	0.423	ng/ul#	98
28) Dibenzo(a,h)anthracene	25.530	278	15004	0.431	ng/ul	100
29) Benzo(g,h,i)perylene	26.171	276	15619	0.431	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103124\
Data File : BN034830.D
Acq On : 02 Nov 2024 07:02
Operator : RC/JU
Sample : PB164449BS
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS449

Quant Time: Nov 03 23:30:04 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 31 11:13:08 2024
Response via : Initial Calibration

