

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103124\
 Data File : BN034753.D
 Acq On : 30 Oct 2024 19:06
 Operator : RC/JU
 Sample : SSTD0.828
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.8212

Quant Time: Oct 31 11:08:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 31 11:07:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.604	152	7372	0.400	ng/ul	0.00
4) Naphthalene-d8	10.372	136	22590	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.236	164	12667	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.983	188	26033	0.400	ng/ul	0.00
17) Chrysene-d12	21.178	240	17747	0.400	ng/ul	0.00
23) Perylene-d12	23.361	264	15252	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.174	96	6015	0.794	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.967	152	22553	0.796	ng/ul	0.00
18) Fluoranthene-d10	19.014	212	44679	0.771	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.208	88	6652	0.815	ng/ul#	92
5) Naphthalene	10.422	128	47313	0.792	ng/ul	99
7) 2-Methylnaphthalene	12.044	142	31100	0.798	ng/ul	99
8) 1-Methylnaphthalene	12.264	142	31489	0.796	ng/ul	100
10) Acenaphthylene	13.954	152	46020	0.797	ng/ul	99
11) Acenaphthene	14.301	153	35094	0.796	ng/ul	99
12) Fluorene	15.291	166	39517	0.803	ng/ul	100
14) Pentachlorophenol	16.633	266	3970	0.760	ng/ul	99
15) Phenanthrene	17.026	178	63000	0.801	ng/ul	100
16) Anthracene	17.114	178	59081	0.806	ng/ul	100
19) Fluoranthene	19.047	202	67395	0.781	ng/ul	99
20) Pyrene	19.404	202	70114	0.781	ng/ul	97
21) Benzo(a)anthracene	21.160	228	56708	0.799	ng/ul	100
22) Chrysene	21.213	228	56223	0.790	ng/ul	100
24) Benzo(b)fluoranthene	22.712	252	50358	0.811	ng/ul	96
25) Benzo(k)fluoranthene	22.753	252	50025	0.801	ng/ul	95
26) Benzo(a)pyrene	23.264	252	43055	0.806	ng/ul	93
27) Indeno(1,2,3-cd)pyrene	25.540	276	44516	0.782	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.554	278	33848	0.787	ng/ul	97
29) Benzo(g,h,i)perylene	26.191	276	34685	0.776	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103124\
Data File : BN034753.D
Acq On : 30 Oct 2024 19:06
Operator : RC/JU
Sample : SST0.828
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.8212

Quant Time: Oct 31 11:08:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 31 11:07:22 2024
Response via : Initial Calibration

