

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103124\
 Data File : BN034767.D
 Acq On : 31 Oct 2024 15:52
 Operator : RC/JU
 Sample : PB164418BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK418

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/04/2024
 Supervised By :mohammad ahmed 11/05/2024

Quant Time: Oct 31 18:09:44 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 31 11:13:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.604	152	7566	0.400	ng/u1	0.00
4) Naphthalene-d8	10.372	136	23269	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.235	164	12797	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.983	188	25752m	0.400	ng/u1	0.00
17) Chrysene-d12	21.174	240	16659	0.400	ng/u1	0.00
23) Perylene-d12	23.360	264	14333m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.170	96	47291	6.038	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.967	152	10215	0.350	ng/u1	0.00
18) Fluoranthene-d10	19.014	212	19272	0.355	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103124\
Data File : BN034767.D
Acq On : 31 Oct 2024 15:52
Operator : RC/JU
Sample : PB164418BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK418

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/04/2024
Supervised By :mohammad ahmed 11/05/2024

Quant Time: Oct 31 18:09:44 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 31 11:13:08 2024
Response via : Initial Calibration

