

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110121\
 Data File : BN017224.D
 Acq On : 01 Nov 2021 22:16
 Operator : CG/JU
 Sample : M4360-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-RW9-GW-58-60

Quant Time: Nov 02 04:08:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 01 17:00:24 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/02/2021
 Supervised By :mohammad ahmed 11/08/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.561	152	32121	0.400	ng	0.00
7) Naphthalene-d8	10.319	136	157051	0.400	ng	0.00
13) Acenaphthene-d10	14.194	164	79994	0.400	ng	0.00
19) Phenanthrene-d10	16.946	188	152455	0.400	ng	0.00
29) Chrysene-d12	21.144	240	145529	0.400	ng	#-0.01
35) Perylene-d12	23.352	264	145708	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.191	112	11593	0.149	ng	0.02
5) Phenol-d6	6.749	99	8090	0.095	ng	0.00
8) Nitrobenzene-d5	8.697	82	21758	0.226	ng	0.00
11) 2-Methylnaphthalene-d10	11.920	152	57461	0.259	ng	0.00
14) 2,4,6-Tribromophenol	15.691	330	8469	0.305	ng	0.00
15) 2-Fluorobiphenyl	12.811	172	76200	0.254	ng	0.00
27) Fluoranthene-d10	18.980	212	134786	0.348	ng	0.00
31) Terphenyl-d14	19.596	244	121553	0.382	ng	0.00
Target Compounds						
25) Phenanthrene	16.982	178	9064m	0.021	ng	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.091	149	56158	0.359	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110121\
Data File : BN017224.D
Acq On : 01 Nov 2021 22:16
Operator : CG/JU
Sample : M4360-02
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-RW9-GW-58-60

Quant Time: Nov 02 04:08:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 01 17:00:24 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021

