


4,6-Dinitro-2-methylphenol


Response Ratio



$R = 3.608e-003 A^2 + 4.214e-002 A - 1.810e-002$
 Coef of Det (r^2) = 0.999446 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN110121.M
 Calibration Table Last Updated: Mon Nov 01 17:00:24 2021

Pentachlorophenol

Response Ratio



$$\text{Response} = 1.160\text{e-}001 * \text{Amt} - 5.940\text{e-}002$$


Coef of Det (r^2) = 0.996408 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN110121.M

Calibration Table Last Updated: Mon Nov 01 17:00:24 2021

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 8.276\text{e-}001 * \text{Amt} - 3.570\text{e-}001$$

Coef of Det (r^2) = 0.996962 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN110121.M

Calibration Table Last Updated: Mon Nov 01 17:00:24 2021