

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110121\
 Data File : BN017232.D
 Acq On : 02 Nov 2021 04:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 02 05:27:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 01 17:00:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.285	0.286	-0.4	106	0.00
3	n-Nitrosodimethylamine	0.376	0.358	4.8	103	0.00
4 S	2-Fluorophenol	0.969	0.850	12.3	95	0.00
5 S	Phenol-d6	1.056	0.909	13.9	95	0.00
6	bis(2-Chloroethyl)ether	1.089	1.065	2.2	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.246	0.226	8.1	106	0.00
9	Naphthalene	1.077	1.022	5.1	99	0.00
10	Hexachlorobutadiene	0.196	0.194	1.0	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.565	0.552	2.3	103	0.00
12	2-Methylnaphthalene	0.663	0.658	0.8	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.139	0.119	14.4	107	0.00
15 S	2-Fluorobiphenyl	1.503	1.479	1.6	103	0.00
16	Acenaphthylene	1.643	1.574	4.2	105	0.00
17	Acenaphthene	1.100	1.096	0.4	105	0.00
18	Fluorene	1.303	1.292	0.8	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	-0.01
20	4,6-Dinitro-2-methylphenol	0.042	0.018	57.1#	91	0.00
21	4-Bromophenyl-phenylether	0.224	0.216	3.6	107	0.00
22	Hexachlorobenzene	0.241	0.240	0.4	108	0.00
23	Atrazine	0.152	0.139	8.6	106	0.00
24	Pentachlorophenol	0.076	0.049	35.5#	104	0.00
25	Phenanthrene	1.121	1.119	0.2	107	0.00
26	Anthracene	0.965	0.912	5.5	107	-0.01
27 SURR	Fluoranthene-d10	1.018	0.981	3.6	107	0.00
28	Fluoranthene	1.228	1.242	-1.1	108	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	109	-0.01
30	Pyrene	1.279	1.224	4.3	109	0.00
31 S	Terphenyl-d14	0.874	0.828	5.3	107	0.00
32	Benzo(a)anthracene	1.242	1.121	9.7	105	0.00
33	Chrysene	1.267	1.292	-2.0	114	0.00
34	Bis(2-ethylhexyl)phthalate	0.598	0.424	29.1#	109	0.00
35 I	Perylene-d12	1.000	1.000	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	1.410	1.354	4.0	110	0.00
37	Benzo(b)fluoranthene	1.287	1.247	3.1	107	0.00
38	Benzo(k)fluoranthene	1.277	1.284	-0.5	114	0.00
39 C	Benzo(a)pyrene	1.162	1.121	3.5	110	0.00
40	Dibenzo(a,h)anthracene	1.139	1.083	4.9	109	0.00
41	Benzo(g,h,i)perylene	1.238	1.156	6.6	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4