

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110121\
 Data File : BN017232.D
 Acq On : 02 Nov 2021 04:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 02 05:27:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 01 17:00:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	1,4-Dioxane	0.400	0.402	-0.5	106	0.00
3	n-Nitrosodimethylamine	0.400	0.381	4.8	103	0.00
4 S	2-Fluorophenol	0.400	0.351	12.3	95	0.00
5 S	Phenol-d6	0.400	0.345	13.8	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.391	2.3	103	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	103	0.00
8 S	Nitrobenzene-d5	0.400	0.368	8.0	106	0.00
9	Naphthalene	0.400	0.380	5.0	99	0.00
10	Hexachlorobutadiene	0.400	0.396	1.0	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	103	0.00
12	2-Methylnaphthalene	0.400	0.397	0.8	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.400	0.342	14.5	107	0.00
15 S	2-Fluorobiphenyl	0.400	0.394	1.5	103	0.00
16	Acenaphthylene	0.400	0.383	4.3	105	0.00
17	Acenaphthene	0.400	0.398	0.5	105	0.00
18	Fluorene	0.400	0.397	0.8	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	107	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.321	19.8	91	0.00
21	4-Bromophenyl-phenylether	0.400	0.386	3.5	107	0.00
22	Hexachlorobenzene	0.400	0.399	0.3	108	0.00
23	Atrazine	0.400	0.366	8.5	106	0.00
24	Pentachlorophenol	0.400	0.376	6.0	104	0.00
25	Phenanthrene	0.400	0.399	0.3	107	0.00
26	Anthracene	0.400	0.378	5.5	107	-0.01
27 SURR	Fluoranthene-d10	0.400	0.386	3.5	107	0.00
28	Fluoranthene	0.400	0.404	-1.0	108	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	109	-0.01
30	Pyrene	0.400	0.383	4.3	109	0.00
31 S	Terphenyl-d14	0.400	0.379	5.3	107	0.00
32	Benzo(a)anthracene	0.400	0.361	9.8	105	0.00
33	Chrysene	0.400	0.408	-2.0	114	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.377	5.8	109	0.00
35 I	Perylene-d12	0.400	0.400	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.384	4.0	110	0.00
37	Benzo(b)fluoranthene	0.400	0.388	3.0	107	0.00
38	Benzo(k)fluoranthene	0.400	0.402	-0.5	114	0.00
39 C	Benzo(a)pyrene	0.400	0.386	3.5	110	0.00
40	Dibenzo(a,h)anthracene	0.400	0.380	5.0	109	0.00
41	Benzo(g,h,i)perylene	0.400	0.373	6.8	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4