

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110121\
 Data File : BN017209.D
 Acq On : 01 Nov 2021 12:35
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Nov 01 15:16:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 01 15:13:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.552	152	32995	0.400	ng	0.00
7) Naphthalene-d8	10.319	136	148841	0.400	ng	0.00
13) Acenaphthene-d10	14.194	164	79951	0.400	ng	0.00
19) Phenanthrene-d10	16.946	188	144607	0.400	ng	0.00
29) Chrysene-d12	21.154	240	134516	0.400	ng	0.00
35) Perylene-d12	23.356	264	137654	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.173	112	8022	0.108	ng	0.00
5) Phenol-d6	6.740	99	8548	0.104	ng	0.00
8) Nitrobenzene-d5	8.697	82	8351	0.096	ng	0.00
11) 2-Methylnaphthalene-d10	11.920	152	21142	0.094	ng	0.00
14) 2,4,6-Tribromophenol	15.691	330	2190	0.076	ng	0.00
15) 2-Fluorobiphenyl	12.811	172	30293	0.101	ng	0.00
27) Fluoranthene-d10	18.985	212	34375	0.085	ng	0.00
31) Terphenyl-d14	19.595	244	28846	0.100	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.163	88	2597	0.148	ng	# 1
3) n-Nitrosodimethylamine	3.494	42	2908	0.101	ng	# 96
6) bis(2-Chloroethyl)ether	6.998	93	9349	0.108	ng	99
9) Naphthalene	10.370	128	45167	0.116	ng	100
10) Hexachlorobutadiene	10.662	225	7555	0.106	ng	# 99
12) 2-Methylnaphthalene	11.991	142	25869	0.106	ng	100
16) Acenaphthylene	13.902	152	28534	0.089	ng	100
17) Acenaphthene	14.257	154	21591	0.098	ng	100
18) Fluorene	15.247	166	24698	0.094	ng	99
20) 4,6-Dinitro-2-methylph...	15.349	198	671	0.042	ng	# 71
21) 4-Bromophenyl-phenylether	16.143	248	7540	0.095	ng	98
22) Hexachlorobenzene	16.252	284	8576	0.099	ng	99
23) Atrazine	16.423	200	4805	0.086	ng	99
24) Pentachlorophenol	16.605	266	1910	0.060	ng	99
25) Phenanthrene	16.982	178	40296	0.100	ng	99
26) Anthracene	17.080	178	31011	0.091	ng	99
28) Fluoranthene	19.015	202	42458	0.097	ng	100
30) Pyrene	19.377	202	42932	0.101	ng	100
32) Benzo(a)anthracene	21.133	228	38092	0.094	ng	99
33) Chrysene	21.186	228	43089	0.101	ng	99
34) Bis(2-ethylhexyl)phtha...	21.091	149	17122	0.100	ng	100
36) Indeno(1,2,3-cd)pyrene	25.574	276	45972	0.092	ng	98
37) Benzo(b)fluoranthene	22.695	252	40047	0.093	ng	97
38) Benzo(k)fluoranthene	22.740	252	43194	0.101	ng	# 91
39) Benzo(a)pyrene	23.256	252	35648	0.092	ng	96
40) Dibenzo(a,h)anthracene	25.591	278	35991	0.089	ng	96
41) Benzo(g,h,i)perylene	26.245	276	40977	0.094	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA N

ClientSampleId :
SSTDICC0.1


Quant Time: Nov 01 15:16:21 2021


Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110121.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Mon Nov 01 15:13:26 2021

Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.552 min Scan# 496
 Delta R.T. -0.000 min
 Lab File: BN017209.D
 Acq: 01 Nov 2021 12:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 32995
 Ion Ratio Lower Upper
 152 100
 150 157.3 127.4 191.0
 115 60.0 48.5 72.7



#2
 1,4-Dioxane
 Concen: 0.148 ng
 RT: 3.163 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: BN017209.D
 Acq: 01 Nov 2021 12:35

Tgt Ion: 88 Resp: 2597
 Ion Ratio Lower Upper
 88 100
 43 397.3 72.9 109.3#
 58 88.2 68.6 103.0









#3
n-Nitrosodimethylamine
Concen: 0.101 ng
RT: 3.494 min Scan# 51
Delta R.T. 0.009 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 42 Resp: 2908
Ion Ratio Lower Upper
42 100
74 130.4 100.6 150.8
44 8.2 4.6 6.8#



#4
2-Fluorophenol
Concen: 0.108 ng
RT: 5.173 min Scan# 238
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion: 112 Resp: 8022
Ion Ratio Lower Upper
112 100
64 58.6 47.1 70.7
63 29.7 23.9 35.9









#5
Phenol-d6
Concen: 0.104 ng
RT: 6.740 min Scan# 408
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Tgt Ion: 99 Resp: 8548




Ion	Ratio	Lower	Upper
99	100		
42	20.1	16.5	24.7
71	31.4	24.7	37.1




#6
bis(2-Chloroethyl)ether
Concen: 0.108 ng
RT: 6.998 min Scan# 436
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion: 93 Resp: 9349

Ion	Ratio	Lower	Upper
93	100		
63	79.8	64.5	96.7
95	33.1	26.5	39.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.319 min Scan# 737
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:136 Resp: 148841
Ion Ratio Lower Upper
136 100
137 10.9 8.7 13.1
54 7.5 6.1 9.1
68 5.5 4.6 6.8



#8
Nitrobenzene-d5
Concen: 0.096 ng
RT: 8.697 min Scan# 609
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion: 82 Resp: 8351
Ion Ratio Lower Upper
82 100
128 41.0 32.0 48.0
54 56.8 45.0 67.4









#9
Naphthalene
Concen: 0.116 ng
RT: 10.370 min Scan# 741
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	128	Resp:	45167
Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.8	13.2
127	13.0	10.3	15.5



#10
Hexachlorobutadiene
Concen: 0.106 ng
RT: 10.662 min Scan# 764
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:	225	Resp:	7555
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.2	50.6	76.0









#11
2-Methylnaphthalene-d10
Concen: 0.094 ng
RT: 11.920 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 21142
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.106 ng
RT: 11.991 min Scan# 939
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:142 Resp: 25869
Ion Ratio Lower Upper
142 100
141 89.7 71.6 107.4
115 33.6 27.0 40.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.194 min Scan# 1188
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 79951
Ion Ratio Lower Upper
164 100
162 99.0 79.8 119.8
160 43.5 35.4 53.2



#14
2,4,6-Tribromophenol
Concen: 0.076 ng
RT: 15.691 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:330 Resp: 2190
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 40.5 31.4 47.2









#15
2-Fluorobiphenyl
Concen: 0.101 ng
RT: 12.811 min Scan# 1079
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:172 Resp: 30293
Ion Ratio Lower Upper
172 100
171 34.6 27.9 41.9
170 23.0 18.4 27.6



#16
Acenaphthylene
Concen: 0.089 ng
RT: 13.902 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:152 Resp: 28534
Ion Ratio Lower Upper
152 100
151 19.3 15.5 23.3
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 0.098 ng
RT: 14.257 min Scan# 1193
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:154 Resp: 21591
Ion Ratio Lower Upper
154 100
153 113.5 90.5 135.7
152 55.3 44.1 66.1



#18
Fluorene
Concen: 0.094 ng
RT: 15.247 min Scan# 1271
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:166 Resp: 24698
Ion Ratio Lower Upper
166 100
165 98.7 78.5 117.7
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.946 min Scan# 1408
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:188 Resp: 144607
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.1 5.1 7.7



#20
4,6-Dinitro-2-methylphenol
Concen: 0.042 ng
RT: 15.349 min Scan# 1279
Delta R.T. 0.013 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:198 Resp: 671
Ion Ratio Lower Upper
198 100
182 28.3 12.6 19.0#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.095 ng
RT: 16.143 min Scan# 1342
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:248 Resp: 7540
Ion Ratio Lower Upper
248 100
250 94.2 74.7 112.1
141 68.8 51.9 77.9



#22
Hexachlorobenzene
Concen: 0.099 ng
RT: 16.252 min Scan# 1351
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:284 Resp: 8576
Ion Ratio Lower Upper
284 100
142 36.4 29.5 44.3
249 31.2 24.7 37.1









#23
Atrazine
Concen: 0.086 ng
RT: 16.423 min Scan# 1365
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:200 Resp: 4805
Ion Ratio Lower Upper
200 100
173 30.1 23.3 34.9
215 47.3 37.5 56.3



#24
Pentachlorophenol
Concen: 0.060 ng
RT: 16.605 min Scan# 1380
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:266 Resp: 1910
Ion Ratio Lower Upper
266 100
264 62.1 49.8 74.6
268 65.9 51.2 76.8









#25
Phenanthrene
Concen: 0.100 ng
RT: 16.982 min Scan# 1411
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:178 Resp: 40296
Ion Ratio Lower Upper
178 100
176 18.9 15.1 22.7
179 15.9 12.3 18.5



#26
Anthracene
Concen: 0.091 ng
RT: 17.080 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:178 Resp: 31011
Ion Ratio Lower Upper
178 100
176 18.1 14.6 22.0
179 15.1 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.085 ng
RT: 18.985 min Scan# 1691
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:212 Resp: 34375
Ion Ratio Lower Upper
212 100
106 13.0 10.6 15.8
104 7.3 5.8 8.6



#28
Fluoranthene
Concen: 0.097 ng
RT: 19.015 min Scan# 1697
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:202 Resp: 42458
Ion Ratio Lower Upper
202 100
101 11.2 8.9 13.3
203 17.0 13.8 20.6









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.154 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:240 Resp: 134516
Ion Ratio Lower Upper
240 100
120 5.8 4.2 6.4
236 25.0 20.2 30.2



#30
Pyrene
Concen: 0.101 ng
RT: 19.377 min Scan# 1770
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:202 Resp: 42932
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.4 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.100 ng
 RT: 19.595 min Scan# 11
 Delta R.T. -0.005 min
 Lab File: BN017209.D
 Acq: 01 Nov 2021 12:35




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244 Resp: 28846
 Ion Ratio Lower Upper
 244 100
 212 7.6 5.7 8.5
 122 13.2 8.3 12.5#



#32
 Benzo(a)anthracene
 Concen: 0.094 ng
 RT: 21.133 min Scan# 2003
 Delta R.T. 0.000 min
 Lab File: BN017209.D
 Acq: 01 Nov 2021 12:35

Tgt Ion:228 Resp: 38092
 Ion Ratio Lower Upper
 228 100
 226 27.2 21.6 32.4
 229 19.8 15.5 23.3









#33
Chrysene
Concen: 0.101 ng
RT: 21.186 min Scan# 2008
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:228 Resp: 43089
Ion Ratio Lower Upper
228 100
226 29.4 23.6 35.4
229 20.0 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.100 ng
RT: 21.091 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35


Tgt Ion:149 Resp: 17122
Ion Ratio Lower Upper
149 100
167 27.2 21.8 32.8
279 4.2 3.1 4.7








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.356 min Scan# 24
Delta R.T. 0.003 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1





Tgt Ion:264 Resp: 137654
Ion Ratio Lower Upper
264 100
260 24.5 19.5 29.3
265 23.3 18.8 28.2




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.092 ng
RT: 25.574 min Scan# 3048
Delta R.T. 0.007 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35



Tgt Ion:276 Resp: 45972
Ion Ratio Lower Upper
276 100
138 27.4 20.2 30.4
277 25.1 20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 0.093 ng
RT: 22.695 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 40047
Ion Ratio Lower Upper
252 100
253 23.6 17.6 26.4
125 11.1 8.2 12.4



#38
Benzo(k)fluoranthene
Concen: 0.101 ng
RT: 22.740 min Scan# 2220
Delta R.T. 0.003 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:252 Resp: 43194
Ion Ratio Lower Upper
252 100
253 24.6 17.7 26.5
125 17.1 8.5 12.7#









#39
Benzo(a)pyrene
Concen: 0.092 ng
RT: 23.256 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 35648
Ion Ratio Lower Upper
252 100
253 24.4 17.8 26.8
125 12.9 9.0 13.6



#40
Dibenzo(a,h)anthracene
Concen: 0.089 ng
RT: 25.591 min Scan# 3053
Delta R.T. 0.000 min
Lab File: BN017209.D
Acq: 01 Nov 2021 12:35

Tgt Ion:278 Resp: 35991
Ion Ratio Lower Upper
278 100
139 19.2 12.9 19.3
279 24.6 19.0 28.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.094 ng
 RT: 26.245 min Scan# 31
 Delta R.T. 0.003 min
 Lab File: BN017209.D
 Acq: 01 Nov 2021 12:35

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	276	Resp:	40977
Ion	Ratio	Lower	Upper
276	100		
277	24.3	19.0	28.4
138	22.1	17.2	25.8

