

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110220\
 Data File : BN012489.D
 Acq On : 02 Nov 2020 14:19
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 02 14:52:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN103020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 02 08:41:36 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.672	0.629	6.4	111	0.00
3	n-Nitrosodimethylamine	1.653	1.648	0.3	101	0.00
4 S	2-Fluorophenol	1.406	1.230	12.5	93	0.00
5 S	Phenol-d6	1.730	1.536	11.2	95	0.00
6	bis(2-Chloroethyl)ether	1.130	0.947	16.2	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	97	-0.01
8 S	Nitrobenzene-d5	0.461	0.410	11.1	97	-0.01
9	Naphthalene	1.129	1.026	9.1	97	-0.01
10	Hexachlorobutadiene	0.235	0.215	8.5	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.656	0.606	7.6	100	0.00
12	2-Methylnaphthalene	0.732	0.667	8.9	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	-0.01
14 S	2,4,6-Tribromophenol	0.270	0.245	9.3	101	0.00
15 S	2-Fluorobiphenyl	1.503	1.346	10.4	99	-0.01
16	Acenaphthylene	1.991	1.791	10.0	101	0.00
17	Acenaphthene	1.272	1.154	9.3	101	0.00
18	Fluorene	1.585	1.419	10.5	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.01
20	4,6-Dinitro-2-methylphenol	0.083	0.067	19.3	100	0.00
21	4-Bromophenyl-phenylether	0.243	0.221	9.1	101	0.00
22	Hexachlorobenzene	0.305	0.283	7.2	101	0.00
23	Atrazine	0.233	0.209	10.3	100	0.00
24	Pentachlorophenol	0.138	0.120	13.0	99	0.00
25	Phenanthrene	1.192	1.073	10.0	100	0.00
26	Anthracene	1.107	0.959	13.4	99	-0.01
27 SURR	Fluoranthene-d10	1.239	1.093	11.8	97	0.00
28	Fluoranthene	1.395	1.225	12.2	98	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
30	Pyrene	1.367	1.265	7.5	98	0.00
31 S	Terphenyl-d14	0.930	0.861	7.4	98	0.00
32	Benzo(a)anthracene	1.247	1.100	11.8	95	0.00
33	Chrysene	1.361	1.248	8.3	96	0.00
34	Bis(2-ethylhexyl)phthalate	0.815	0.727	10.8	93	0.00
35	Indeno(1,2,3-cd)pyrene	1.990	1.844	7.3	90	0.00
36 I	Perylene-d12	1.000	1.000	0.0	92	0.00
37	Benzo(b)fluoranthene	1.290	1.115	13.6	92	0.00
38	Benzo(k)fluoranthene	1.352	1.228	9.2	93	0.00
39 C	Benzo(a)pyrene	1.231	1.133	8.0	96	0.00
40	Dibenzo(a,h)anthracene	1.394	1.285	7.8	91	0.00
41	Benzo(g,h,i)perylene	1.455	1.345	7.6	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4