

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110220\
 Data File : BN012489.D
 Acq On : 02 Nov 2020 14:19
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 02 14:52:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN103020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 02 08:41:36 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.374	6.5	111	0.00
3	n-Nitrosodimethylamine	0.400	0.399	0.3	101	0.00
4 S	2-Fluorophenol	0.400	0.350	12.5	93	0.00
5 S	Phenol-d6	0.400	0.355	11.3	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.335	16.3	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	97	-0.01
8 S	Nitrobenzene-d5	0.400	0.356	11.0	97	-0.01
9	Naphthalene	0.400	0.363	9.3	97	-0.01
10	Hexachlorobutadiene	0.400	0.367	8.3	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	100	0.00
12	2-Methylnaphthalene	0.400	0.365	8.8	97	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.363	9.3	101	0.00
15 S	2-Fluorobiphenyl	0.400	0.358	10.5	99	-0.01
16	Acenaphthylene	0.400	0.360	10.0	101	0.00
17	Acenaphthene	0.400	0.363	9.3	101	0.00
18	Fluorene	0.400	0.358	10.5	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.348	13.0	100	0.00
21	4-Bromophenyl-phenylether	0.400	0.364	9.0	101	0.00
22	Hexachlorobenzene	0.400	0.371	7.3	101	0.00
23	Atrazine	0.400	0.359	10.3	100	0.00
24	Pentachlorophenol	0.400	0.347	13.3	99	0.00
25	Phenanthrene	0.400	0.360	10.0	100	0.00
26	Anthracene	0.400	0.346	13.5	99	-0.01
27 SURR	Fluoranthene-d10	0.400	0.353	11.8	97	0.00
28	Fluoranthene	0.400	0.351	12.3	98	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
30	Pyrene	0.400	0.370	7.5	98	0.00
31 S	Terphenyl-d14	0.400	0.370	7.5	98	0.00
32	Benzo(a)anthracene	0.400	0.353	11.8	95	0.00
33	Chrysene	0.400	0.367	8.3	96	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.357	10.8	93	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.371	7.3	90	0.00
36 I	Perylene-d12	0.400	0.400	0.0	92	0.00
37	Benzo(b)fluoranthene	0.400	0.346	13.5	92	0.00
38	Benzo(k)fluoranthene	0.400	0.363	9.3	93	0.00
39 C	Benzo(a)pyrene	0.400	0.368	8.0	96	0.00
40	Dibenzo(a,h)anthracene	0.400	0.369	7.8	91	0.00
41	Benzo(g,h,i)perylene	0.400	0.370	7.5	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4