

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110221\
 Data File : BN017244.D
 Acq On : 02 Nov 2021 17:23
 Operator : CG/JU
 Sample : PB140385BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS385

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/02/2021
 Supervised By :mohammad ahmed 11/08/2021

Quant Time: Nov 02 17:52:04 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN110221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 02 15:59:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.511	152	224946	20.000	ng/u1	0.00
20) Naphthalene-d8	10.287	136	1136865	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.169	164	749210	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.922	188	1590230	20.000	ng/u1	0.00
79) Chrysene-d12	21.139	240	1700360	20.000	ng/u1	0.00
88) Perylene-d12	23.339	264	1709059	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.034	96	28713	5.045	ng/uL	0.00
4) Pyridine-d5	3.428	84	403173	25.388	ng/u1	0.00
7) Phenol-d5	6.699	99	573286	27.610	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.864	67	333158	26.881	ng/u1	0.00
11) 2-Chlorophenol-d4	7.046	132	463429	28.087	ng/u1	0.00
15) 4-Methylphenol-d8	8.234	113	481081	28.102	ng/u1	0.00
21) Nitrobenzene-d5	8.664	128	233680	26.492	ng/u1	0.00
24) 2-Nitrophenol-d4	9.381	143	266268	27.064	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.916	165	475414	26.875	ng/u1	0.00
31) 4-Chloroaniline-d4	10.434	131	636734	24.018	ng/u1	0.00
46) Dimethylphthalate-d6	13.593	166	1470358	26.469	ng/u1	0.00
49) Acenaphthylene-d8	13.857	160	1832476	26.227	ng/u1	0.00
54) 4-Nitrophenol-d4	14.399	143	297942	27.326	ng/u1	0.00
60) Fluorene-d10	15.169	176	1270703	26.806	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.304	200	273990	26.935	ng/u1	0.00
73) Anthracene-d10	17.022	188	2012166	26.812	ng/u1	0.00
81) Pyrene-d10	19.328	212	2382498	25.894	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.204	264	2592775	28.081	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.070	88	62944	11.299	ng/uL	89
5) Pyridine	3.446	79	440980	27.258	ng/u1	100
6) Benzaldehyde	6.664	77	341926	32.448	ng/u1	98
8) Phenol	6.728	94	635159	30.190	ng/u1	98
10) Bis(2-Chloroethyl)ether	6.952	93	489334	29.351	ng/u1	99
12) 2-Chlorophenol	7.075	128	511203	30.307	ng/u1	99
13) 2-Methylphenol	7.964	108	489210	30.263	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.052	45	721506	29.225	ng/u1	100
16) Acetophenone	8.334	105	778569	30.235	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.328	70	391444	29.844	ng/u1	99
18) 4-Methylphenol	8.293	108	544216	30.401	ng/u1	99
19) Hexachloroethane	8.575	117	195832	29.722	ng/u1	95
22) Nitrobenzene	8.705	77	553664	27.922	ng/u1	99
23) Isophorone	9.228	82	1107804	28.106	ng/u1	99
25) 2-Nitrophenol	9.411	139	309726	29.205	ng/u1	96
26) 2,4-Dimethylphenol	9.487	107	592540	28.471	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.722	93	701852	27.891	ng/u1	98
29) 2,4-Dichlorophenol	9.940	162	513676	29.352	ng/u1	96
30) Naphthalene	10.334	128	1712930	28.038	ng/u1	100
32) 4-Chloroaniline	10.458	127	699362	26.274	ng/u1	98
33) Hexachlorobutadiene	10.622	225	294436	27.547	ng/u1	98
34) Caprolactam	11.210	113	181408m	28.154	ng/u1	
35) 4-Chloro-3-methylphenol	11.599	107	566821	28.881	ng/u1	98
36) 2-Methylnaphthalene	11.963	142	1181920	27.883	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110221\
 Data File : BN017244.D
 Acq On : 02 Nov 2021 17:23
 Operator : CG/JU
 Sample : PB140385BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SLCS385

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/02/2021
 Supervised By :mohammad ahmed 11/08/2021

Quant Time: Nov 02 17:52:04 2021

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN110221.M


Quant Title : SVOA CALIBRATION

QLast Update : Tue Nov 02 15:59:34 2021

Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.181	142	1188184	27.283	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.340	216	595196	27.739	ng/ul	98
40) Hexachlorocyclopentadiene	12.316	237	359495	26.294	ng/ul	96
41) 2,4,6-Trichlorophenol	12.587	196	412883	28.972	ng/ul	98
42) 2,4,5-Trichlorophenol	12.657	196	455778	29.164	ng/ul	98
43) 1,1'-Biphenyl	12.999	154	1575611	27.965	ng/ul	99
44) 2-Chloronaphthalene	13.034	162	1218920	28.051	ng/ul	98
45) 2-Nitroaniline	13.252	65	371355	30.709	ng/ul	96
47) Dimethylphthalate	13.640	163	1589387	28.731	ng/ul	100
48) 2,6-Dinitrotoluene	13.763	165	331384	30.530	ng/ul	99
50) Acenaphthylene	13.887	152	2058065	28.651	ng/ul	100
51) 3-Nitroaniline	14.087	138	336131	29.468	ng/ul	97
52) Acenaphthene	14.234	153	1315051	28.248	ng/ul	97
53) 2,4-Dinitrophenol	14.299	184	202897	28.692	ng/ul	93
55) 4-Nitrophenol	14.410	109	228530	30.177	ng/ul	94
56) Dibenzofuran	14.569	168	1878938	28.444	ng/ul	100
57) 2,4-Dinitrotoluene	14.551	165	493837	30.750	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.804	232	394877	29.752	ng/ul	96
59) Diethylphthalate	15.022	149	1603167	28.579	ng/ul	99
61) Fluorene	15.222	166	1499302	28.413	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.228	204	735827	28.031	ng/ul	99
63) 4-Nitroaniline	15.257	138	367531m	32.384	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.316	198	296932	29.404	ng/ul	98
67) N-Nitrosodiphenylamine	15.445	169	1345268	28.584	ng/ul	100
68) 4-Bromophenyl-phenylether	16.122	248	470486	28.605	ng/ul	95
69) Hexachlorobenzene	16.228	284	548595	28.689	ng/ul	98
70) Atrazine	16.410	200	496936	28.763	ng/ul	98
71) Pentachlorophenol	16.575	266	329201	27.994	ng/ul#	89
72) Phenanthrene	16.963	178	2461596	28.346	ng/ul	99
74) Anthracene	17.057	178	2537118	29.023	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.952	216	608604	27.319	ng/ul	99
76) Pentachlorobenzene	14.493	250	611512	26.411	ng/ul	97
77) Carbazole	17.334	167	2342584	29.904	ng/ul	97
78) Di-n-butylphthalate	17.922	149	2861201	30.321	ng/ul	99
80) Fluoranthene	18.992	202	3013660	27.815	ng/ul	96
82) Pyrene	19.363	202	3094288	27.884	ng/ul	97
83) Butylbenzylphthalate	20.292	149	1341084	29.055	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.069	252	1100130	28.112	ng/ul	98
85) Benzo(a)anthracene	21.122	228	3074860	27.578	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.081	149	2033766	28.911	ng/ul	98
87) Chrysene	21.175	228	3002469	27.725	ng/ul	97
89) Di-n-octyl phthalate	21.957	149	3553126	31.550	ng/ul	100
90) Benzo(b)fluoranthene	22.680	252	3438759	30.195	ng/ul	98
91) Benzo(k)fluoranthene	22.727	252	3264681	30.428	ng/ul	100
93) Benzo(a)pyrene	23.245	252	3400411	30.732	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	25.551	276	3974441	30.473	ng/ul	99
95) Dibenzo(a,h)anthracene	25.574	278	3392349	30.685	ng/ul	99
96) Benzo(g,h,i)perylene	26.227	276	3363806	30.480	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.299 ng/uL
RT: 3.070 min Scan# 14
Delta R.T. -0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385



Tgt Ion: 88 Resp: 62944
Ion Ratio Lower Upper
88 100
43 33.6 24.2 36.4
58 80.7 56.4 84.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#5
Pyridine
Concen: 27.258 ng/uL
RT: 3.446 min Scan# 78
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23


Tgt Ion: 79 Resp: 440980
Ion Ratio Lower Upper
79 100
52 81.7 65.3 97.9
51 38.2 30.8 46.2





#6
Benzaldehyde
Concen: 32.448 ng/ul
RT: 6.664 min Scan# 61
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385



Tgt Ion: 77 Resp: 341920
Ion Ratio Lower Upper
77 100
105 101.9 80.3 120.5
106 95.5 78.2 117.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#8
Phenol
Concen: 30.190 ng/ul
RT: 6.728 min Scan# 636
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23


Tgt Ion: 94 Resp: 635159
Ion Ratio Lower Upper
94 100
65 28.5 21.5 32.3
66 35.9 28.1 42.1





#10
Bis(2-Chloroethyl)ether
Concen: 29.351 ng/ul
RT: 6.952 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion: 93 Resp: 489334
Ion Ratio Lower Upper
93 100
63 80.2 63.8 95.8
95 32.7 26.6 39.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#12
2-Chlorophenol
Concen: 30.307 ng/ul
RT: 7.075 min Scan# 695
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23



Tgt Ion:128 Resp: 511203
Ion Ratio Lower Upper
128 100
64 45.5 37.0 55.6
130 32.3 26.3 39.5






#13
2-Methylphenol
Concen: 30.263 ng/ul
RT: 7.964 min Scan# 846
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :

BNA_N

ClientSampleId :

SLCS385






Tgt Ion:108 Resp: 489210
Ion Ratio Lower Upper
108 100
107 90.1 71.6 107.4


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#14
2,2'-oxybis(1-Chloropropane)
Concen: 29.225 ng/ul
RT: 8.052 min Scan# 861
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Tgt Ion: 45 Resp: 721506
Ion Ratio Lower Upper
45 100
77 13.0 10.3 15.5
79 10.3 8.2 12.4





#16
Acetophenone
Concen: 30.235 ng/ul
RT: 8.334 min Scan# 908
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:105 Resp: 778569
Ion Ratio Lower Upper
105 100
77 73.0 59.5 89.3
51 26.5 22.0 33.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#17
N-Nitroso-di-n-propylamine
Concen: 29.844 ng/ul
RT: 8.328 min Scan# 908
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Tgt Ion: 70 Resp: 391444
Ion Ratio Lower Upper
70 100
42 60.5 49.4 74.2
101 10.3 7.7 11.5
130 23.4 18.5 27.7





#18
4-Methylphenol
Concen: 30.401 ng/ul
RT: 8.293 min Scan# 902
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23



Instrument :
BNA_N
ClientSampleId :
SLCS385



Tgt Ion:108 Resp: 544210
Ion Ratio Lower Upper
108 100
107 113.8 89.8 134.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#19
Hexachloroethane
Concen: 29.722 ng/ul
RT: 8.575 min Scan# 950
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23


Tgt Ion:117 Resp: 195832
Ion Ratio Lower Upper
117 100
201 99.7 76.0 114.0
199 61.6 46.6 70.0





#22
Nitrobenzene
Concen: 27.922 ng/u1
RT: 8.705 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion: 77 Resp: 553664
Ion Ratio Lower Upper
77 100
123 47.9 37.8 56.6
65 13.5 10.6 15.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#23
Isophorone
Concen: 28.106 ng/u1
RT: 9.228 min Scan# 1061
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23



Tgt Ion: 82 Resp: 1107804
Ion Ratio Lower Upper
82 100
95 7.1 5.8 8.6
138 19.0 14.8 22.2





#25

2-Nitrophenol

Concen: 29.205 ng/ul

RT: 9.411 min Scan# 1092

Delta R.T. 0.000 min

Lab File: BN017244.D

Acq: 02 Nov 2021 17:23

Instrument :

BNA_N

ClientSampleId :

SLCS385





Tgt Ion:139 Resp: 309720

Ion	Ratio	Lower	Upper
139	100		
65	47.4	40.2	60.4
109	27.5	23.0	34.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#26

2,4-Dimethylphenol

Concen: 28.471 ng/ul

RT: 9.487 min Scan# 1105




Delta R.T. 0.000 min


Lab File: BN017244.D

Acq: 02 Nov 2021 17:23

Tgt Ion:107 Resp: 592540


Ion	Ratio	Lower	Upper
107	100		
121	53.6	40.6	60.8
122	90.3	71.4	107.2





#27
Bis(2-Chloroethoxy)methane
Concen: 27.891 ng/ul
RT: 9.722 min Scan# 1145
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385



Tgt Ion: 93 Resp: 701852
Ion Ratio Lower Upper
93 100
95 31.5 26.5 39.7
123 11.2 9.4 14.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#29
2,4-Dichlorophenol
Concen: 29.352 ng/ul
RT: 9.940 min Scan# 1182
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23


Tgt Ion:162 Resp: 513676
Ion Ratio Lower Upper
162 100
164 63.6 53.4 80.0
98 35.8 30.1 45.1





#30
Naphthalene
Concen: 28.038 ng/u1
RT: 10.334 min Scan# 1249
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385



Tgt Ion:128 Resp: 1712930
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.0 10.3 15.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#32
4-Chloroaniline
Concen: 26.274 ng/u1
RT: 10.458 min Scan# 1270
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23


Tgt Ion:127 Resp: 699362
Ion Ratio Lower Upper
127 100
129 31.9 26.6 40.0





#33
Hexachlorobutadiene
Concen: 27.547 ng/ul
RT: 10.622 min Scan# 1298
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:225 Resp: 294430
Ion Ratio Lower Upper
225 100
223 61.0 50.6 75.8
227 62.9 51.0 76.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#34
Caprolactam
Concen: 28.154 ng/ul m
RT: 11.210 min Scan# 1398
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23



Tgt Ion:113 Resp: 181408
Ion Ratio Lower Upper
113 100
55 171.7 137.8 206.8
56 128.9 99.0 148.4





#35
4-Chloro-3-methylphenol
Concen: 28.881 ng/ul
RT: 11.599 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :

BNA_N

ClientSampleId :





SLCS385

Tgt Ion:107 Resp: 56682
Ion Ratio Lower Upper
107 100
144 27.2 21.2 31.8
142 80.3 65.8 98.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#36
2-Methylnaphthalene
Concen: 27.883 ng/ul
RT: 11.963 min Scan# 1526
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23


Tgt Ion:142 Resp: 1181920
Ion Ratio Lower Upper
142 100
141 88.0 70.6 106.0





#37
1-Methylnaphthalene
Concen: 27.283 ng/ul
RT: 12.181 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:142 Resp: 1188184
Ion Ratio Lower Upper
142 100
141 92.1 75.4 113.0
116 3.8 3.0 4.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#39
1,2,4,5-Tetrachlorobenzene
Concen: 27.739 ng/ul
RT: 12.340 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Tgt Ion:216 Resp: 595196
Ion Ratio Lower Upper
216 100
214 78.7 60.8 91.2
179 21.1 17.0 25.6
108 17.0 13.5 20.3





#40
Hexachlorocyclopentadiene
Concen: 26.294 ng/ul
RT: 12.316 min Scan# 1586
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:237 Resp: 35949
Ion Ratio Lower Upper
237 100
235 64.1 48.6 72.8
272 14.0 10.0 15.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#41
2,4,6-Trichlorophenol
Concen: 28.972 ng/ul
RT: 12.587 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23



Tgt Ion:196 Resp: 412883
Ion Ratio Lower Upper
196 100
198 95.6 74.8 112.2
200 30.9 25.1 37.7





#42
2,4,5-Trichlorophenol
Concen: 29.164 ng/ul
RT: 12.657 min Scan# 10
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :

BNA_N

ClientSampleId :





SLCS385

Tgt Ion:196 Resp: 455778
Ion Ratio Lower Upper
196 100
198 88.8 72.4 108.6
200 30.4 24.8 37.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#43
1,1'-Biphenyl
Concen: 27.965 ng/ul
RT: 12.999 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23


Tgt Ion:154 Resp: 1575611
Ion Ratio Lower Upper
154 100
153 40.2 31.6 47.4
76 12.9 10.2 15.2





#44
2-Chloronaphthalene
Concen: 28.051 ng/u1
RT: 13.034 min Scan# 1708
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:162 Resp: 1218920
Ion Ratio Lower Upper
162 100
164 31.9 26.6 39.8
127 39.7 32.4 48.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#45
2-Nitroaniline
Concen: 30.709 ng/u1
RT: 13.252 min Scan# 1745
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Tgt Ion: 65 Resp: 371355
Ion Ratio Lower Upper
65 100
92 69.9 55.4 83.0
138 113.3 85.9 128.9





#47
Dimethylphthalate
Concen: 28.731 ng/ul
RT: 13.640 min Scan# 1811
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:163 Resp: 158938
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.2
164 10.3 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#48
2,6-Dinitrotoluene
Concen: 30.530 ng/ul
RT: 13.763 min Scan# 1832
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23



Tgt Ion:165 Resp: 331384
Ion Ratio Lower Upper
165 100
89 52.6 41.5 62.3
121 20.4 15.8 23.8






#50
Acenaphthylene
Concen: 28.651 ng/u1
RT: 13.887 min Scan# 1853
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :

BNA_N

ClientSampleId :

SLCS385






Tgt Ion:152 Resp: 205806
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 13.5 10.6 16.0


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#51
3-Nitroaniline
Concen: 29.468 ng/u1
RT: 14.087 min Scan# 1887
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Tgt Ion:138 Resp: 336131
Ion Ratio Lower Upper
138 100
108 10.3 8.5 12.7
92 110.5 91.0 136.4





#52
Acenaphthene
Concen: 28.248 ng/u1
RT: 14.234 min Scan# 1912
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:153 Resp: 131505
Ion Ratio Lower Upper
153 100
152 47.0 37.1 55.7
154 85.2 71.0 106.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#53
2,4-Dinitrophenol
Concen: 28.692 ng/u1
RT: 14.299 min Scan# 1923
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Tgt Ion:184 Resp: 202897
Ion Ratio Lower Upper
184 100
63 59.2 52.8 79.2
154 59.2 51.0 76.4





#55
4-Nitrophenol
Concen: 30.177 ng/ul
RT: 14.410 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385



Tgt Ion:109 Resp: 228530
Ion Ratio Lower Upper
109 100
139 145.6 124.8 187.2
65 137.7 114.3 171.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#56
Dibenzofuran
Concen: 28.444 ng/ul
RT: 14.569 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23


Tgt Ion:168 Resp: 1878938
Ion Ratio Lower Upper
168 100
139 36.4 29.4 44.0





#57
2,4-Dinitrotoluene
Concen: 30.750 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:165 Resp: 49383
Ion Ratio Lower Upper
165 100
63 42.8 34.0 51.0
182 2.8 2.2 3.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#58
2,3,4,6-Tetrachlorophenol
Concen: 29.752 ng/ul
RT: 14.804 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23



Tgt Ion:232 Resp: 394877
Ion Ratio Lower Upper
232 100
131 43.6 37.6 56.4
130 2.3 1.9 2.9
166 29.7 25.3 37.9






#59
Diethylphthalate
Concen: 28.579 ng/ul
RT: 15.022 min Scan# 2046
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :

BNA_N

ClientSampleId :

SLCS385






Tgt Ion:149 Resp: 160316
Ion Ratio Lower Upper
149 100
177 22.3 17.9 26.9
150 12.2 10.2 15.2


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#61
Fluorene
Concen: 28.413 ng/ul
RT: 15.222 min Scan# 2080
Delta R.T. -0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23



Tgt Ion:166 Resp: 1499302
Ion Ratio Lower Upper
166 100
165 95.0 78.7 118.1
167 13.6 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 28.031 ng/ul
RT: 15.228 min Scan# 2081
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :

BNA_N

ClientSampleId :





SLCS385

Tgt Ion:204 Resp: 73582
Ion Ratio Lower Upper
204 100
206 32.6 25.9 38.9
141 60.3 49.0 73.4

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#63
4-Nitroaniline
Concen: 32.384 ng/ul m
RT: 15.257 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Tgt Ion:138 Resp: 367531
Ion Ratio Lower Upper
138 100
92 51.0 40.3 60.5
108 71.4 55.1 82.7





#66
4,6-Dinitro-2-methylphenol
Concen: 29.404 ng/ul
RT: 15.316 min Scan# 2096
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :

BNA_N

ClientSampleId :





SLCS385

Tgt Ion:198 Resp: 296932
Ion Ratio Lower Upper
198 100
182 3.9 3.5 5.3
77 24.2 18.5 27.7

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#67
N-Nitrosodiphenylamine
Concen: 28.584 ng/ul
RT: 15.445 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Tgt Ion:169 Resp: 1345268
Ion Ratio Lower Upper
169 100
168 67.6 54.0 81.0
167 35.9 28.5 42.7





#68
4-Bromophenyl-phenylether
Concen: 28.605 ng/ul
RT: 16.122 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :

BNA_N

ClientSampleId :





SLCS385

Tgt Ion:248 Resp: 470480
Ion Ratio Lower Upper
248 100
250 94.7 79.1 118.7
141 77.8 58.5 87.7

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#69
Hexachlorobenzene
Concen: 28.689 ng/ul
RT: 16.228 min Scan# 2251
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23


Tgt Ion:284 Resp: 548595
Ion Ratio Lower Upper
284 100
142 34.9 29.0 43.6
249 30.9 24.0 36.0





#70
Atrazine
Concen: 28.763 ng/ul
RT: 16.410 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:200 Resp: 496930
Ion Ratio Lower Upper
200 100
173 27.8 21.4 32.2
215 48.1 39.4 59.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#71
Pentachlorophenol
Concen: 27.994 ng/ul
RT: 16.575 min Scan# 2310
Delta R.T. -0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23



Tgt Ion:266 Resp: 329201
Ion Ratio Lower Upper
266 100
264 68.7 44.4 66.6#
268 65.8 50.2 75.2









#72
Phenanthrene
Concen: 28.346 ng/u1
RT: 16.963 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :
BNA_N
ClientSampleId :
SLCS385

Tgt Ion:178 Resp: 2461590
Ion Ratio Lower Upper
178 100
179 15.0 12.2 18.4
176 18.9 14.6 22.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#74
Anthracene
Concen: 29.023 ng/u1
RT: 17.057 min Scan# 2392
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Tgt Ion:178 Resp: 2537118
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 19.4 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 27.319 ng/uL
RT: 12.952 min Scan# 1694
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :
BNA_N
ClientSampleId :
SLCS385

Tgt Ion:216 Resp: 608604
Ion Ratio Lower Upper
216 100
214 77.0 62.3 93.5
218 47.0 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#76
Pentachlorobenzene
Concen: 26.411 ng/uL
RT: 14.493 min Scan# 1956
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23


Tgt Ion:250 Resp: 611512
Ion Ratio Lower Upper
250 100
248 64.8 48.9 73.3
252 66.7 52.2 78.4





#77
 Carbazole
 Concen: 29.904 ng/ul
 RT: 17.334 min Scan# 2439
 Delta R.T. 0.000 min
 Lab File: BN017244.D
 Acq: 02 Nov 2021 17:23

Instrument :
 BNA_N
 ClientSampleId :
 SLCS385






Tgt Ion:167 Resp: 2342584


Ion	Ratio	Lower	Upper
167	100		
166	20.6	18.0	27.0
139	12.9	10.5	15.7

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/02/2021
 Supervised By :mohammad ahmed 11/08/2021






#78
 Di-n-butylphthalate
 Concen: 30.321 ng/ul
 RT: 17.922 min Scan# 2539
 Delta R.T. 0.000 min
 Lab File: BN017244.D
 Acq: 02 Nov 2021 17:23



Tgt Ion:149 Resp: 2861201


Ion	Ratio	Lower	Upper
149	100		
150	8.8	7.3	10.9
104	4.6	3.8	5.8





#80
Fluoranthene
Concen: 27.815 ng/u1
RT: 18.992 min Scan# 2721
Delta R.T. -0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:202 Resp: 3013660
Ion Ratio Lower Upper
202 100
101 12.7 8.8 13.2
100 9.7 6.7 10.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#82
Pyrene
Concen: 27.884 ng/u1
RT: 19.363 min Scan# 2784
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Tgt Ion:202 Resp: 3094288
Ion Ratio Lower Upper
202 100
101 13.6 11.9 17.9
100 11.0 9.6 14.4





#83
Butylbenzylphthalate
Concen: 29.055 ng/ul
RT: 20.292 min Scan# 2942
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:149 Resp: 1341084
Ion Ratio Lower Upper
149 100
91 68.1 54.7 82.1
206 20.5 16.1 24.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#84
3,3'-Dichlorobenzidine
Concen: 28.112 ng/ul
RT: 21.069 min Scan# 3074
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Tgt Ion:252 Resp: 1100130
Ion Ratio Lower Upper
252 100
254 61.4 50.5 75.7
126 11.8 9.0 13.6





#85
Benzo(a)anthracene
Concen: 27.578 ng/ul
RT: 21.122 min Scan# 3083
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385




Tgt Ion:228 Resp: 3074860
Ion Ratio Lower Upper
228 100
229 20.2 15.0 22.4
226 28.8 21.2 31.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#86
Bis(2-ethylhexyl)phthalate
Concen: 28.911 ng/ul
RT: 21.081 min Scan# 3076
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Tgt Ion:149 Resp: 2033766
Ion Ratio Lower Upper
149 100
167 27.9 21.6 32.4
279 4.2 3.2 4.8





#87
Chrysene
Concen: 27.725 ng/u1
RT: 21.175 min Scan# 3092
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23




Instrument :
BNA_N
ClientSampleId :
SLCS385



Tgt Ion:228 Resp: 3002469
Ion Ratio Lower Upper
228 100
226 29.8 25.3 37.9
229 19.4 16.8 25.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#89
Di-n-octyl phthalate
Concen: 31.550 ng/u1
RT: 21.957 min Scan# 3225
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Tgt Ion:149 Resp: 3553126






#90
Benzo(b)fluoranthene
Concen: 30.195 ng/u1
RT: 22.680 min Scan# 3348
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :

BN017244.D

ClientSampleId :

SLCS385






Tgt Ion:252 Resp: 3438759
Ion Ratio Lower Upper
252 100
253 21.7 16.5 24.7
125 11.1 8.7 13.1


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021




#91
Benzo(k)fluoranthene
Concen: 30.428 ng/u1
RT: 22.727 min Scan# 3356
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23



Tgt Ion:252 Resp: 3264681
Ion Ratio Lower Upper
252 100
253 21.2 17.0 25.6
125 10.6 8.8 13.2









#93
Benzo(a)pyrene
Concen: 30.732 ng/ul
RT: 23.245 min Scan# 3444
Delta R.T. 0.000 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :
BNA_N
ClientSampleId :
SLCS385

Tgt Ion:252 Resp: 340041
Ion Ratio Lower Upper
252 100
253 22.3 17.0 25.4
125 12.2 9.2 13.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#94
Indeno(1,2,3-cd)pyrene
Concen: 30.473 ng/ul
RT: 25.551 min Scan# 3836
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Tgt Ion:276 Resp: 3974441
Ion Ratio Lower Upper
276 100
138 24.3 19.4 29.0
277 22.8 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 30.685 ng/ul
RT: 25.574 min Scan# 3840
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Instrument :

BNA_N

ClientSampleId :





SLCS385

Tgt Ion:278 Resp: 3392349
Ion Ratio Lower Upper
278 100
139 17.6 14.1 21.1
279 23.8 19.7 29.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/02/2021
Supervised By :mohammad ahmed 11/08/2021



#96
Benzo(g,h,i)perylene
Concen: 30.480 ng/ul
RT: 26.227 min Scan# 3951
Delta R.T. 0.006 min
Lab File: BN017244.D
Acq: 02 Nov 2021 17:23

Tgt Ion:276 Resp: 3363806
Ion Ratio Lower Upper
276 100
138 23.6 18.6 27.8
277 23.8 18.4 27.6

