

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110221\
 Data File : BN017254.D
 Acq On : 03 Nov 2021 00:01
 Operator : CG/JU
 Sample : PB140383BS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS383

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2021
 Supervised By :mohammad ahmed 11/08/2021

Quant Time: Nov 03 01:23:50 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN110221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 02 15:59:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.511	152	210551	20.000	ng/u1	0.00
20) Naphthalene-d8	10.287	136	980723	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.169	164	641316	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.922	188	1328695	20.000	ng/u1	0.00
79) Chrysene-d12	21.139	240	1365495	20.000	ng/u1	0.00
88) Perylene-d12	23.333	264	1531683	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.034	96	30122	5.654	ng/uL	0.00
4) Pyridine-d5	3.428	84	417348	28.077	ng/u1	0.00
7) Phenol-d5	6.699	99	583607	30.028	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.863	67	342470	29.522	ng/u1	0.00
11) 2-Chlorophenol-d4	7.046	132	469329	30.389	ng/u1	0.00
15) 4-Methylphenol-d8	8.234	113	479659	29.935	ng/u1	0.00
21) Nitrobenzene-d5	8.663	128	235953	31.008	ng/u1	0.00
24) 2-Nitrophenol-d4	9.381	143	263060	30.995	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.916	165	468956	30.731	ng/u1	0.00
31) 4-Chloroaniline-d4	10.434	131	656236	28.694	ng/u1	0.00
46) Dimethylphthalate-d6	13.593	166	1462695	30.761	ng/u1	0.00
49) Acenaphthylene-d8	13.857	160	1847117	30.884	ng/u1	0.00
54) 4-Nitrophenol-d4	14.398	143	286414	30.688	ng/u1	0.00
60) Fluorene-d10	15.169	176	1264807	31.171	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.304	200	272995	32.120	ng/u1	0.00
73) Anthracene-d10	17.022	188	1983639	31.635	ng/u1	0.00
81) Pyrene-d10	19.328	212	2316924	31.356	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.198	264	2473144	29.887	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.070	88	64495	12.369	ng/uL	94
5) Pyridine	3.446	79	449309	29.672	ng/u1	99
6) Benzaldehyde	6.663	77	341984	34.672	ng/u1	99
8) Phenol	6.728	94	632676	32.128	ng/u1	100
10) Bis(2-Chloroethyl)ether	6.952	93	494263	31.674	ng/u1	98
12) 2-Chlorophenol	7.075	128	513996	32.556	ng/u1	99
13) 2-Methylphenol	7.963	108	487973	32.251	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.052	45	729553	31.571	ng/u1	99
16) Acetophenone	8.334	105	773911	32.109	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.328	70	386962	31.519	ng/u1	98
18) 4-Methylphenol	8.293	108	542032	32.349	ng/u1	100
19) Hexachloroethane	8.575	117	195016	31.622	ng/u1	94
22) Nitrobenzene	8.705	77	558825	32.669	ng/u1	100
23) Isophorone	9.228	82	1112691	32.724	ng/u1	99
25) 2-Nitrophenol	9.410	139	307662	33.629	ng/u1	97
26) 2,4-Dimethylphenol	9.487	107	592343	32.992	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.722	93	694087	31.974	ng/u1	99
29) 2,4-Dichlorophenol	9.940	162	504062	33.388	ng/u1	99
30) Naphthalene	10.334	128	1702465	32.303	ng/u1	99
32) 4-Chloroaniline	10.457	127	725414	31.592	ng/u1	99
33) Hexachlorobutadiene	10.622	225	303873	32.956	ng/u1	98
34) Caprolactam	11.210	113	175291m	31.536	ng/u1	
35) 4-Chloro-3-methylphenol	11.599	107	562430	33.219	ng/u1	100
36) 2-Methylnaphthalene	11.957	142	1182658	32.342	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110221\
 Data File : BN017254.D
 Acq On : 03 Nov 2021 00:01
 Operator : CG/JU
 Sample : PB140383BS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS383


Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2021
 Supervised By :mohammad ahmed 11/08/2021

Quant Time: Nov 03 01:23:50 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN110221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 02 15:59:34 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.181	142	1189288	31.656	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.340	216	610877	33.260	ng/ul	99
40) Hexachlorocyclopentadiene	12.316	237	354431	30.285	ng/ul	99
41) 2,4,6-Trichlorophenol	12.587	196	406745	33.343	ng/ul	99
42) 2,4,5-Trichlorophenol	12.657	196	439309	32.840	ng/ul	96
43) 1,1'-Biphenyl	12.998	154	1582447	32.812	ng/ul	99
44) 2-Chloronaphthalene	13.034	162	1214753	32.658	ng/ul	100
45) 2-Nitroaniline	13.251	65	355054	34.301	ng/ul	94
47) Dimethylphthalate	13.640	163	1552169	32.779	ng/ul	99
48) 2,6-Dinitrotoluene	13.763	165	319486	34.386	ng/ul	98
50) Acenaphthylene	13.887	152	2011449	32.713	ng/ul	100
51) 3-Nitroaniline	14.087	138	325732	33.361	ng/ul	96
52) Acenaphthene	14.234	153	1306514	32.786	ng/ul	99
53) 2,4-Dinitrophenol	14.298	184	192995	31.883	ng/ul	95
55) 4-Nitrophenol	14.410	109	219013	33.786	ng/ul	94
56) Dibenzofuran	14.569	168	1864249	32.969	ng/ul	98
57) 2,4-Dinitrotoluene	14.551	165	479224	34.861	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.804	232	376039	33.099	ng/ul	99
59) Diethylphthalate	15.022	149	1574382	32.787	ng/ul	99
61) Fluorene	15.222	166	1473058	32.612	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.228	204	721954	32.129	ng/ul	99
63) 4-Nitroaniline	15.257	138	353450m	36.382	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.316	198	287061	34.022	ng/ul	99
67) N-Nitrosodiphenylamine	15.445	169	1337788	34.020	ng/ul	97
68) 4-Bromophenyl-phenylether	16.122	248	462902	33.684	ng/ul	98
69) Hexachlorobenzene	16.228	284	537111	33.617	ng/ul	99
70) Atrazine	16.416	200	477021	33.045	ng/ul	99
71) Pentachlorophenol	16.575	266	325505	33.129	ng/ul	95
72) Phenanthrene	16.963	178	2430082	33.491	ng/ul	99
74) Anthracene	17.057	178	2492047	34.119	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	12.951	216	614231	32.998	ng/ul	98
76) Pentachlorobenzene	14.492	250	623691	32.239	ng/ul	94
77) Carbazole	17.334	167	2287515	34.949	ng/ul	98
78) Di-n-butylphthalate	17.916	149	2754971	34.941	ng/ul	100
80) Fluoranthene	18.992	202	2952312	33.932	ng/ul	96
82) Pyrene	19.357	202	2953790	33.145	ng/ul	99
83) Butylbenzylphthalate	20.292	149	1263707	34.092	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.069	252	1052275	33.483	ng/ul	99
85) Benzo(a)anthracene	21.122	228	2996170	33.463	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.080	149	1931295	34.187	ng/ul	97
87) Chrysene	21.174	228	2881482	33.133	ng/ul	97
89) Di-n-octyl phthalate	21.951	149	3341176	33.104	ng/ul	100
90) Benzo(b)fluoranthene	22.674	252	3232426	31.670	ng/ul	97
91) Benzo(k)fluoranthene	22.727	252	3142605	32.682	ng/ul	98
93) Benzo(a)pyrene	23.245	252	3229305	32.565	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.551	276	3754101	32.117	ng/ul	97
95) Dibenzo(a,h)anthracene	25.563	278	3197067	32.268	ng/ul	98
96) Benzo(g,h,i)perylene	26.227	276	3173294	32.083	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.369 ng/uL
RT: 3.070 min Scan# 14
Delta R.T. -0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383



Tgt Ion: 88 Resp: 6449
Ion Ratio Lower Upper
88 100
43 31.5 24.2 36.4
58 76.7 56.4 84.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#5
Pyridine
Concen: 29.672 ng/uL
RT: 3.446 min Scan# 78
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01


Tgt Ion: 79 Resp: 449309
Ion Ratio Lower Upper
79 100
52 81.0 65.3 97.9
51 37.6 30.8 46.2





#6
Benzaldehyde
Concen: 34.672 ng/ul
RT: 6.663 min Scan# 61
Delta R.T. 0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion: 77 Resp: 341984
Ion Ratio Lower Upper
77 100
105 101.7 80.3 120.5
106 97.7 78.2 117.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#8
Phenol
Concen: 32.128 ng/ul
RT: 6.728 min Scan# 636
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion: 94 Resp: 632676
Ion Ratio Lower Upper
94 100
65 27.0 21.5 32.3
66 35.3 28.1 42.1





#10
Bis(2-Chloroethyl)ether
Concen: 31.674 ng/ul
RT: 6.952 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion: 93 Resp: 49426
Ion Ratio Lower Upper
93 100
63 77.9 63.8 95.8
95 31.8 26.6 39.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#12
2-Chlorophenol
Concen: 32.556 ng/ul
RT: 7.075 min Scan# 695
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:128 Resp: 513996
Ion Ratio Lower Upper
128 100
64 47.5 37.0 55.6
130 33.0 26.3 39.5





#13
2-Methylphenol
Concen: 32.251 ng/ul
RT: 7.963 min Scan# 846
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383



Tgt Ion:108 Resp: 48797
Ion Ratio Lower Upper
108 100
107 89.2 71.6 107.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.571 ng/ul
RT: 8.052 min Scan# 861
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01


Tgt Ion: 45 Resp: 729553
Ion Ratio Lower Upper
45 100
77 13.1 10.3 15.5
79 10.5 8.2 12.4





#16
Acetophenone
Concen: 32.109 ng/ul
RT: 8.334 min Scan# 909
Delta R.T. 0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:105 Resp: 77391
Ion Ratio Lower Upper
105 100
77 73.9 59.5 89.3
51 26.6 22.0 33.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#17
N-Nitroso-di-n-propylamine
Concen: 31.519 ng/ul
RT: 8.328 min Scan# 908
Delta R.T. 0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01



Tgt Ion: 70 Resp: 386962
Ion Ratio Lower Upper
70 100
42 63.3 49.4 74.2
101 10.4 7.7 11.5
130 24.4 18.5 27.7






#18
4-Methylphenol
Concen: 32.349 ng/ul
RT: 8.293 min Scan# 902
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Instrument :

BNA_N

ClientSampleId :

SLCS383






Tgt Ion:108 Resp: 54203
Ion Ratio Lower Upper
108 100
107 112.6 89.8 134.8


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#19
Hexachloroethane
Concen: 31.622 ng/ul
RT: 8.575 min Scan# 950
Delta R.T. 0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:117 Resp: 195016
Ion Ratio Lower Upper
117 100
201 100.8 76.0 114.0
199 62.1 46.6 70.0





#22
Nitrobenzene
Concen: 32.669 ng/u1
RT: 8.705 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383



Tgt Ion: 77 Resp: 558825
Ion Ratio Lower Upper
77 100
123 47.2 37.8 56.6
65 13.4 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#23
Isophorone
Concen: 32.724 ng/u1
RT: 9.228 min Scan# 1061
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Tgt Ion: 82 Resp: 1112691
Ion Ratio Lower Upper
82 100
95 6.9 5.8 8.6
138 18.2 14.8 22.2









#25
2-Nitrophenol
Concen: 33.629 ng/ul
RT: 9.410 min Scan# 1092
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Instrument :
BNA_N
ClientSampleId :
SLCS383

Tgt Ion:139 Resp: 307662
Ion Ratio Lower Upper
139 100
65 47.9 40.2 60.4
109 28.0 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#26
2,4-Dimethylphenol
Concen: 32.992 ng/ul
RT: 9.487 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01


Tgt Ion:107 Resp: 592343
Ion Ratio Lower Upper
107 100
121 51.5 40.6 60.8
122 87.1 71.4 107.2





#27
Bis(2-Chloroethoxy)methane
Concen: 31.974 ng/ul
RT: 9.722 min Scan# 1145
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion: 93 Resp: 69408
Ion Ratio Lower Upper
93 100
95 32.8 26.5 39.7
123 11.5 9.4 14.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#29
2,4-Dichlorophenol
Concen: 33.388 ng/ul
RT: 9.940 min Scan# 1182
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:162 Resp: 504062
Ion Ratio Lower Upper
162 100
164 67.0 53.4 80.0
98 36.7 30.1 45.1





#30
Naphthalene
Concen: 32.303 ng/u1
RT: 10.334 min Scan# 1249
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:128 Resp: 1702465
Ion Ratio Lower Upper
128 100
129 11.2 8.8 13.2
127 13.3 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#32
4-Chloroaniline
Concen: 31.592 ng/u1
RT: 10.457 min Scan# 1270
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:127 Resp: 725414
Ion Ratio Lower Upper
127 100
129 32.6 26.6 40.0





#33
Hexachlorobutadiene
Concen: 32.956 ng/ul
RT: 10.622 min Scan# 1298
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:225 Resp: 30387
Ion Ratio Lower Upper
225 100
223 62.9 50.6 75.8
227 61.6 51.0 76.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#34
Caprolactam
Concen: 31.536 ng/ul m
RT: 11.210 min Scan# 1398
Delta R.T. 0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01



Tgt Ion:113 Resp: 175291
Ion Ratio Lower Upper
113 100
55 168.6 137.8 206.8
56 125.8 99.0 148.4





#35
4-Chloro-3-methylphenol
Concen: 33.219 ng/ul
RT: 11.599 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Instrument :

BNA_N

ClientSampleId :





SLCS383

Tgt Ion:107 Resp: 562430
Ion Ratio Lower Upper
107 100
144 27.2 21.2 31.8
142 82.4 65.8 98.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#36
2-Methylnaphthalene
Concen: 32.342 ng/ul
RT: 11.957 min Scan# 1525
Delta R.T. -0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Tgt Ion:142 Resp: 1182658
Ion Ratio Lower Upper
142 100
141 89.0 70.6 106.0






#37
1-Methylnaphthalene
Concen: 31.656 ng/ul
RT: 12.181 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Instrument :

BNA_N

ClientSampleId :

SLCS383






Tgt Ion:142 Resp: 118928
Ion Ratio Lower Upper
142 100
141 94.9 75.4 113.0
116 3.7 3.0 4.6


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#39
1,2,4,5-Tetrachlorobenzene
Concen: 33.260 ng/ul
RT: 12.340 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:216 Resp: 610877
Ion Ratio Lower Upper
216 100
214 76.3 60.8 91.2
179 20.0 17.0 25.6
108 16.2 13.5 20.3





#40
Hexachlorocyclopentadiene
Concen: 30.285 ng/ul
RT: 12.316 min Scan# 1586
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:237 Resp: 35443
Ion Ratio Lower Upper
237 100
235 61.4 48.6 72.8
272 13.6 10.0 15.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#41
2,4,6-Trichlorophenol
Concen: 33.343 ng/ul
RT: 12.587 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:196 Resp: 406745
Ion Ratio Lower Upper
196 100
198 93.9 74.8 112.2
200 30.9 25.1 37.7





#42
2,4,5-Trichlorophenol
Concen: 32.840 ng/u1
RT: 12.657 min Scan# 1644
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:196 Resp: 439309
Ion Ratio Lower Upper
196 100
198 95.3 72.4 108.6
200 31.7 24.8 37.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#43
1,1'-Biphenyl
Concen: 32.812 ng/u1
RT: 12.998 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:154 Resp: 1582447
Ion Ratio Lower Upper
154 100
153 40.4 31.6 47.4
76 12.7 10.2 15.2





#44
2-Chloronaphthalene
Concen: 32.658 ng/ul
RT: 13.034 min Scan# 1707
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:162 Resp: 121475
Ion Ratio Lower Upper
162 100
164 33.2 26.6 39.8
127 40.0 32.4 48.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#45
2-Nitroaniline
Concen: 34.301 ng/ul
RT: 13.251 min Scan# 1745
Delta R.T. 0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion: 65 Resp: 355054
Ion Ratio Lower Upper
65 100
92 71.6 55.4 83.0
138 116.0 85.9 128.9





#47
Dimethylphthalate
Concen: 32.779 ng/ul
RT: 13.640 min Scan# 1811
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:163 Resp: 1552169
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.2
164 10.4 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#48
2,6-Dinitrotoluene
Concen: 34.386 ng/ul
RT: 13.763 min Scan# 1832
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:165 Resp: 319486
Ion Ratio Lower Upper
165 100
89 52.7 41.5 62.3
121 21.0 15.8 23.8





#50
Acenaphthylene
Concen: 32.713 ng/ul
RT: 13.887 min Scan# 1853
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:152 Resp: 2011449
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.8
153 13.3 10.6 16.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#51
3-Nitroaniline
Concen: 33.361 ng/ul
RT: 14.087 min Scan# 1887
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:138 Resp: 325732
Ion Ratio Lower Upper
138 100
108 10.6 8.5 12.7
92 108.6 91.0 136.4





#52
Acenaphthene
Concen: 32.786 ng/ul
RT: 14.234 min Scan# 1912
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:153 Resp: 1306514
Ion Ratio Lower Upper
153 100
152 46.9 37.1 55.7
154 89.4 71.0 106.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#53
2,4-Dinitrophenol
Concen: 31.883 ng/ul
RT: 14.298 min Scan# 1923
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:184 Resp: 192995
Ion Ratio Lower Upper
184 100
63 62.4 52.8 79.2
154 58.8 51.0 76.4





#55
4-Nitrophenol
Concen: 33.786 ng/ul
RT: 14.410 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383



Tgt Ion:109 Resp: 21901
Ion Ratio Lower Upper
109 100
139 146.0 124.8 187.2
65 137.3 114.3 171.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#56
Dibenzofuran
Concen: 32.969 ng/ul
RT: 14.569 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01


Tgt Ion:168 Resp: 1864249
Ion Ratio Lower Upper
168 100
139 38.1 29.4 44.0





#57
2,4-Dinitrotoluene
Concen: 34.861 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:165 Resp: 479224
Ion Ratio Lower Upper
165 100
63 43.3 34.0 51.0
182 2.9 2.2 3.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#58
2,3,4,6-Tetrachlorophenol
Concen: 33.099 ng/ul
RT: 14.804 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01



Tgt Ion:232 Resp: 376039
Ion Ratio Lower Upper
232 100
131 46.3 37.6 56.4
130 2.2 1.9 2.9
166 30.6 25.3 37.9






#59
Diethylphthalate
Concen: 32.787 ng/ul
RT: 15.022 min Scan# 2046
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Instrument :

BNA_N

ClientSampleId :

SLCS383






Tgt Ion:149 Resp: 1574382
Ion Ratio Lower Upper
149 100
177 22.0 17.9 26.9
150 12.5 10.2 15.2


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#61
Fluorene
Concen: 32.612 ng/ul
RT: 15.222 min Scan# 2080
Delta R.T. -0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01



Tgt Ion:166 Resp: 1473058
Ion Ratio Lower Upper
166 100
165 96.7 78.7 118.1
167 13.1 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 32.129 ng/ul
RT: 15.228 min Scan# 2081
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Instrument :

BNA_N

ClientSampleId :





SLCS383

Tgt Ion:204 Resp: 721954
Ion Ratio Lower Upper
204 100
206 33.4 25.9 38.9
141 60.6 49.0 73.4

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#63
4-Nitroaniline
Concen: 36.382 ng/ul m
RT: 15.257 min Scan# 2086
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01


Tgt Ion:138 Resp: 353450
Ion Ratio Lower Upper
138 100
92 50.2 40.3 60.5
108 72.8 55.1 82.7





#66
4,6-Dinitro-2-methylphenol
Concen: 34.022 ng/ul
RT: 15.316 min Scan# 2096
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:198 Resp: 287061
Ion Ratio Lower Upper
198 100
182 4.0 3.5 5.3
77 23.7 18.5 27.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#67
N-Nitrosodiphenylamine
Concen: 34.020 ng/ul
RT: 15.445 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01



Tgt Ion:169 Resp: 1337788
Ion Ratio Lower Upper
169 100
168 64.0 54.0 81.0
167 35.6 28.5 42.7





#68
4-Bromophenyl-phenylether
Concen: 33.684 ng/ul
RT: 16.122 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Instrument :

BNA_N

ClientSampleId :





SLCS383

Tgt Ion:248 Resp: 462902
Ion Ratio Lower Upper
248 100
250 96.3 79.1 118.7
141 74.6 58.5 87.7

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#69
Hexachlorobenzene
Concen: 33.617 ng/ul
RT: 16.228 min Scan# 2251
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01


Tgt Ion:284 Resp: 537111
Ion Ratio Lower Upper
284 100
142 36.8 29.0 43.6
249 30.9 24.0 36.0





#70
Atrazine
Concen: 33.045 ng/ul
RT: 16.416 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383



Tgt Ion:200 Resp: 47702
Ion Ratio Lower Upper
200 100
173 27.2 21.4 32.2
215 48.4 39.4 59.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#71
Pentachlorophenol
Concen: 33.129 ng/ul
RT: 16.575 min Scan# 2310
Delta R.T. -0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01


Tgt Ion:266 Resp: 325505
Ion Ratio Lower Upper
266 100
264 61.7 44.4 66.6
268 61.2 50.2 75.2





#72
Phenanthrene
Concen: 33.491 ng/u1
RT: 16.963 min Scan# 2376
Delta R.T. -0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383



Tgt Ion:178 Resp: 243008
Ion Ratio Lower Upper
178 100
179 15.0 12.2 18.4
176 18.6 14.6 22.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#74
Anthracene
Concen: 34.119 ng/u1
RT: 17.057 min Scan# 2392
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01


Tgt Ion:178 Resp: 2492047
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.2
176 18.6 15.1 22.7





#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.998 ng/uL
RT: 12.951 min Scan# 1694
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:216 Resp: 61423
Ion Ratio Lower Upper
216 100
214 78.8 62.3 93.5
218 48.5 37.0 55.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#76
Pentachlorobenzene
Concen: 32.239 ng/uL
RT: 14.492 min Scan# 1956
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:250 Resp: 623691
Ion Ratio Lower Upper
250 100
248 62.5 48.9 73.3
252 58.1 52.2 78.4





#77
 Carbazole
 Concen: 34.949 ng/ul
 RT: 17.334 min Scan# 2439
 Delta R.T. 0.000 min
 Lab File: BN017254.D
 Acq: 03 Nov 2021 00:01




Instrument :
 BNA_N
 ClientSampleId :
 SLCS383




Tgt Ion:167 Resp: 2287511
 Ion Ratio Lower Upper
 167 100
 166 21.1 18.0 27.0
 139 13.2 10.5 15.7

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
 Supervised By :mohammad ahmed 11/08/2021




#78
 Di-n-butylphthalate
 Concen: 34.941 ng/ul
 RT: 17.916 min Scan# 2538
 Delta R.T. -0.006 min
 Lab File: BN017254.D
 Acq: 03 Nov 2021 00:01




Tgt Ion:149 Resp: 2754971
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.3 10.9
 104 4.9 3.8 5.8





#80
Fluoranthene
Concen: 33.932 ng/u1
RT: 18.992 min Scan# 2721
Delta R.T. -0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:202 Resp: 2952312
Ion Ratio Lower Upper
202 100
101 12.8 8.8 13.2
100 9.7 6.7 10.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#82
Pyrene
Concen: 33.145 ng/u1
RT: 19.357 min Scan# 2783
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01



Tgt Ion:202 Resp: 2953790
Ion Ratio Lower Upper
202 100
101 15.3 11.9 17.9
100 12.6 9.6 14.4





#83
Butylbenzylphthalate
Concen: 34.092 ng/ul
RT: 20.292 min Scan# 2942
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Instrument :

BNA_N

ClientSampleId :





SLCS383

Tgt Ion:149 Resp: 126370
Ion Ratio Lower Upper
149 100
91 67.6 54.7 82.1
206 21.3 16.1 24.1

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#84
3,3'-Dichlorobenzidine
Concen: 33.483 ng/ul
RT: 21.069 min Scan# 3074
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Tgt Ion:252 Resp: 1052275
Ion Ratio Lower Upper
252 100
254 64.2 50.5 75.7
126 10.8 9.0 13.6






#85
Benzo(a)anthracene
Concen: 33.463 ng/ul
RT: 21.122 min Scan# 3083
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Instrument :

BNA_N

ClientSampleId :

SLCS383






Tgt Ion:228 Resp: 2996170
Ion Ratio Lower Upper
228 100
229 19.7 15.0 22.4
226 27.1 21.2 31.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#86
Bis(2-ethylhexyl)phthalate
Concen: 34.187 ng/ul
RT: 21.080 min Scan# 3076
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01


Tgt Ion:149 Resp: 1931295
Ion Ratio Lower Upper
149 100
167 28.9 21.6 32.4
279 4.7 3.2 4.8





#87
Chrysene
Concen: 33.133 ng/ul
RT: 21.174 min Scan# 3092
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383



Tgt Ion:228 Resp: 288148
Ion Ratio Lower Upper
228 100
226 29.6 25.3 37.9
229 20.2 16.8 25.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#89
Di-n-octyl phthalate
Concen: 33.104 ng/ul
RT: 21.951 min Scan# 3224
Delta R.T. -0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01


Tgt Ion:149 Resp: 3341176





#90
Benzo(b)fluoranthene
Concen: 31.670 ng/u1
RT: 22.674 min Scan# 3347
Delta R.T. -0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:252 Resp: 3232420
Ion Ratio Lower Upper
252 100
253 22.3 16.5 24.7
125 11.5 8.7 13.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#91
Benzo(k)fluoranthene
Concen: 32.682 ng/u1
RT: 22.727 min Scan# 3356
Delta R.T. 0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01



Tgt Ion:252 Resp: 3142605
Ion Ratio Lower Upper
252 100
253 20.3 17.0 25.6
125 10.1 8.8 13.2






#93
Benzo(a)pyrene
Concen: 32.565 ng/ul
RT: 23.245 min Scan# 3444
Delta R.T. 0.000 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01

Instrument :

BNA_N

ClientSampleId :

SLCS383






Tgt Ion:252 Resp: 322930
Ion Ratio Lower Upper
252 100
253 21.7 17.0 25.4
125 11.4 9.2 13.8


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021




#94
Indeno(1,2,3-cd)pyrene
Concen: 32.117 ng/ul
RT: 25.551 min Scan# 3836
Delta R.T. 0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Tgt Ion:276 Resp: 3754101
Ion Ratio Lower Upper
276 100
138 26.7 19.4 29.0
277 24.1 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 32.268 ng/ul
RT: 25.563 min Scan# 319706
Delta R.T. -0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01




Instrument :
BNA_N
ClientSampleId :
SLCS383




Tgt Ion:278 Resp: 319706
Ion Ratio Lower Upper
278 100
139 18.5 14.1 21.1
279 24.0 19.7 29.5

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/03/2021
Supervised By :mohammad ahmed 11/08/2021



#96
Benzo(g,h,i)perylene
Concen: 32.083 ng/ul
RT: 26.227 min Scan# 3951
Delta R.T. 0.006 min
Lab File: BN017254.D
Acq: 03 Nov 2021 00:01



Tgt Ion:276 Resp: 3173294
Ion Ratio Lower Upper
276 100
138 22.8 18.6 27.8
277 23.7 18.4 27.6

