

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110321\
 Data File : BN017277.D
 Acq On : 03 Nov 2021 15:56
 Operator : CG/JU
 Sample : M4374-06
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BG337

Quant Time: Nov 03 16:28:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 03 10:42:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.514	152	5144	0.400	ng/ul	0.00
4) Naphthalene-d8	10.288	136	21325	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.169	164	13449	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.922	188	28934	0.400	ng/ul	0.00
17) Chrysene-d12	21.138	240	26349	0.400	ng/ul	0.00
23) Perylene-d12	23.332	264	28112	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.038	96	26576	4.647	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.888	152	10859	0.349	ng/ul	0.00
18) Fluoranthene-d10	18.965	212	31968	0.463	ng/ul	0.00
Target Compounds						
15) Phenanthrene	16.964	178	2122	0.023	ng/ul#	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110321\
Data File : BN017277.D
Acq On : 03 Nov 2021 15:56
Operator : CG/JU
Sample : M4374-06
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BG337

Quant Time: Nov 03 16:28:56 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 03 10:42:03 2021
Response via : Initial Calibration

