

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110321\
 Data File : BN017285.D
 Acq On : 03 Nov 2021 21:21
 Operator : CG/JU
 Sample : PB140308BS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS308

Quant Time: Nov 08 01:06:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 03 10:42:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.515	152	4000	0.400	ng/ul	0.00
4) Naphthalene-d8	10.289	136	15733	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.171	164	9664	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.924	188	21909	0.400	ng/ul	0.00
17) Chrysene-d12	21.139	240	21030	0.400	ng/ul	0.00
23) Perylene-d12	23.334	264	19150	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.042	96	3186	0.716	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.890	152	8685	0.378	ng/ul	0.00
18) Fluoranthene-d10	18.967	212	22929	0.416	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.072	88	8434	1.766	ng/ul#	78
5) Naphthalene	10.339	128	16073	0.369	ng/ul	100
7) 2-Methylnaphthalene	11.967	142	10867	0.360	ng/ul	99
8) 1-Methylnaphthalene	12.187	142	10897	0.353	ng/ul	100
10) Acenaphthylene	13.884	152	13891	0.363	ng/ul	100
11) Acenaphthene	14.231	153	12850	0.383	ng/ul	100
12) Fluorene	15.230	166	14769	0.377	ng/ul	99
14) Pentachlorophenol	16.586	266	2232	0.452	ng/ul	99
15) Phenanthrene	16.966	178	26114	0.376	ng/ul	99
16) Anthracene	17.059	178	20942	0.363	ng/ul	100
19) Fluoranthene	18.995	202	30226	0.402	ng/ul	100
20) Pyrene	19.357	202	30754	0.398	ng/ul	96
21) Benzo(a)anthracene	21.121	228	24010	0.349	ng/ul	99
22) Chrysene	21.174	228	30883	0.395	ng/ul	100
24) Benzo(b)fluoranthene	22.676	252	29039	0.407	ng/ul	99
25) Benzo(k)fluoranthene	22.720	252	33944	0.419	ng/ul	99
26) Benzo(a)pyrene	23.237	252	24415	0.381	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	25.543	276	32712	0.360	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.563	278	25798	0.359	ng/ul	98
29) Benzo(g,h,i)perylene	26.211	276	27994	0.364	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110321\
Data File : BN017285.D
Acq On : 03 Nov 2021 21:21
Operator : CG/JU
Sample : PB140308BS
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS308

Quant Time: Nov 08 01:06:47 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 03 10:42:03 2021
Response via : Initial Calibration

